Canon
CANON MEXICANA, S. DE R.L. DE C.V.

TITLE: **DEMONSTRATIONS**					
	VALIDITY		BEGIN	VERSION	
REF: HEALTHCARE-017-18	DAY	MONTH	YEAR		PAGE 1 OF 4
	31	05	2018	**1 .0**	

I. OBJECTIVE

Establish the necessary guidelines and technical procedures to carry out Healthcare device demonstrations.

II. SCOPE

This policy will be applicable to the Healthcare Solutions Sales and Support Area

III. RESPONSABILITIES

The Healthcare Area, in order to ensure the demonstration quality, effectiveness and efficiency, will use the following guidelines as a procedure

The Healthcare area is composed of two divisions, radiology and ophthalmology, and is responsible for carrying out the device demonstrations as well as ensuring that the applicable technical and safety conditions are complied.

IV. PROCEDURE

1) The Healthcare Manager or Specialist will submit demonstration requests to the Healthcare Service Department.

2) Demonstration setup. The following aspects should be covered.

    a. Configuration and performance tests. Request to the Healthcare Manager information regarding the customer's points of interest, type and physical facilities in order to properly configure the device.

PREPARED BY:	**REVIEWED BY:**	**APPROVED BY:**
DANIEL PEREZ HEALTHCARE SOLUTIONS SALES AND SUPPORT ENGINEER DATE: 31/05/2018	FEDERICO TAFOYA HEALTHCARE SOLUTIONS SALES AND SUPPORT MANAGER DATE: 31/05/2018	MORIHIRO ETO SALES ADMINISTRATION BISG DIRECTOR DATE: 31/05/2018

b. Each one of the device components must comply with the cleaning instructions in the operation and service manual, especially in devices with optical systems.

c. The devices must be packed preferably in the original shipping boxes, the specifically designed case as second option or in a package which ensures effective protection against hits and mistreatment. In any case, "fragile" labels will be added to all packages.

    i. Check the packages condition.
    ii. Check the state of the foams which hold the device
    iii. Assure the device packing inside the boxes or cases

d. Investigate relevant clinical applications related to the customer activity.

3) Once packaged, coordinate with the Healthcare Specialist shipment to the appropriate address.
    i. Print and deliver to the surveillance and showroom staff the device output formats.
    ii. Print and place shipping guidelines when using commercial parcel service (UPS)
    iii. Ask the Healthcare Specialist or Healthcare Manager for information regarding the entry and exit requirements of the device into the demonstration area.

4) Once confirmed that the device has arrived to the final destination, proceed with the demonstration. Because demonstrations are an essential part of the purchasing and decision process, it is crucial to carry out a professional labor with technical and clinical knowledge. In order to carry out effective and efficient demonstrations, it is mandatory that demonstrations are executed only by personnel belonging to or certified by the Healthcare Service Department.
    a. Only personnel authorized by the Healthcare Service Department are able to unpack the devices. In case of any missing or damaged component, report it immediately to the Healthcare Assistant, Service Supervisor or Healthcare Manager in order to take the appropriate actions.

b. Perform the installation and start up. In the case of the Radiology product line, it is mandatory to use radiological protection devices.
c. Focus the demonstration on the customer points of interest as well as on the device competitive advantages.
d. Bt the end of the demonstration, recapitulate the demonstrated points and verify that there are no questions or inconclusive comments.
e. In case the demonstration duration is longer than one day, provide basic training regarding operation and care in order that the participants are able to properly use the devices under their working conditions.
f. Provide contact information for support.
g. Verify the demonstration end date.
h. It is responsibility of the Healthcare Service Department staff to monitor the correct use of the device during the demonstration period, as well as to answer any technical questions or requests.

5) At the end of the demonstration, the Healthcare Manager or Service Supervisor will request the user support in order to perform the disinfection procedure according to the user and service manual, taking special care if the device was in contact with body fluids. It is the responsibility of the Healthcare Manager to assure that the Healthcare Service Staff is provided with latex gloves and disinfectant products as a protection measure for these cases.

6) The Healthcare Manager or Service Supervisor must assure that the devices are packed in the original boxes or cases in accordance with subsection 2.c, and returned to the Canon Mexicana facilities complying with the demonstration contract.

7) Ask the demonstration attendees to fill out the satisfaction survey

8) Once the devices are back at Canon facilities, the physical and functional status of the devices will be verified, reporting to the Service Supervisor or Healthcare Manager any anomaly in order to take the corresponding actions.

9) Each one of the devices components must comply with the cleaning and disinfection parameters according to the operation and service manual, especially in optical components and those that have been exposed to contact with the patient.