Canon
CANON MEXICANA, S. DE R.L. DE C.V.

TITLE: **MEDICAL DEVICE WARRANTY**					
	VALIDITY BEGIN			VERSION	
REF: HEALTHCARE-020/18	DAY	MONTH	YEAR		PAGE 1 OF 4
	31	05	2018	**1 .0**	

I. OBJECTIVE

Establish the guidelines in order to validate the Healthcare device warranty.

II. SCOPE

This policy is applicable to the Canon Mexicana employees and for dealers authorized to commercialize or provide service to Healthcare devices.

III. RESPONSABILITIES

The Healthcare Service Department has the qualification and training to adequately attend the customer warranty requests. In case that any other Canon Mexicana representative receive a Healthcare device warranty request, must refer it to the Healthcare Service Department in order to assure the service quality and customer satisfaction and carry out the following points as process.

It is responsibility of the Service Supervisor and the Healthcare Manager to grant authorization regarding services, spare parts, updates, freight and any other applicable subjects for the repair of the device under warranty, as well as the commercial and warranty conditions according to the current policies.

IV. PROCEDURE

1) As a standard, Canon Healthcare guarantees the products against manufacturing defects in parts, components, performance or labor, for a period of 1 year from the invoice date. Warranty policy does not include complete device replacement, only the damaged or defective parts will be considered.
2) It is responsibility of the Healthcare Manager to obtain authorization from the BISG Sales Management Director regarding any exceptions to the standard warranty terms.

PREPARED BY	**REVIEWED BY:**	**APPROVED BY:**
MARCIAL JIMENEZ SUPERVISOR HEALTHCARE SOLUTIONS SALES AND SUPPORT FECHA: 31/05/2018	FEDERICO TAFOYA GERENTE HEALTHCARE SOLUTIONS SALES AND SUPPORT FECHA: 31/05/2018	MORIHIRO ETO DIRECTOR ADMINISTRACIÓN DE VENTAS BISG FECHA: 31/05/2018

3) It is possible to offer extended warranty periods which will consider additional costs depending on the type of device. It is responsibility of the Healthcare Manager and Service Supervisor to confirm the corresponding terms and costs with Canon USA.
4) Receive a written request from the customer or dealer that must include name, phone number, email address, location, device model, serial number and invoice date.
5) Verify if the device was sold directly by Canon Mexicana or through a dealer, as well as the invoice date
6) Dealer case:
    i. In the event that the device was supplied directly by Canon Mexicana, it will be responsibility of Healthcare Service Department to execute the warranty. In this case any dealer will be able to perform the service. It will be responsibility of the Service Supervisor and the Healthcare Manager to evaluate and authorize any exceptions.
    ii. In the event that the device was supplied by a dealer certified and authorized to offer service to Healthcare devices, it will be their responsibility to inspect the device and decide if warranty applies, also to provide follow up and execute the warranty labors in accordance to the stated conditions in the distribution agreement, having technical support from Healthcare Service Department.
    iii. In the event that the device was supplied by a dealer non-certified or non-authorized to offer service to Healthcare devices, the Healthcare Service Department will be responsible to follow up and execute the warranty labor. In this case, the dealer must send the device to the Healthcare Service Department in order to inspect and correct the failure without charge to the dealer.
    iv. The dealer will diagnose the reason of the device malfunction and confirm the necessary spare parts and corrective actions.
    v. The dealer will request in writing to the Healthcare Service Department the spare parts needed to correct the device failure. The parts will be requested to the Service Administration Department using the valid format. Prior to shipment to the dealer, the Healthcare Service Department must receive the damaged parts for review, evaluation and final disposal by the Service Management Department.
    vi. The spare parts will be sent to the dealer's address without charge and the dealer will be responsible for the replacement procedure and to run the appropriate good performance tests prior to returning the device to the customer.

vii. The dealer must deliver the repaired device at the customer facilities, free of charge for both, customer and Canon Mexicana.
viii. The dealer will be responsible to follow up the device good performance.

B. Customer Case:
   i. If it the device was directly supplied by a dealer, they will be responsible to execute the warranty according to points 6.ii and 6.iii, offering the customer a quality service and attention at all time.
   ii. In case the device was supplied directly by Canon Mexicana, the situation must be evaluated in order to know the device failure and determine the possible causes.
   iii. If the failure is determined to be caused by fraud or device misuse, warranty will not be valid and proceed according to the policy of technical support HEALTHCARE-015.
   iv. If the device presents a failure that could be considered under warranty coverage, the Healthcare Service Department must perform a detailed evaluation of the device. In all cases the evaluation must be carried out at the Healthcare Service Department facilities. The customer will be responsible for proper packaging, preferably in the original boxes, shipping and all related costs. Under no circumstances Canon will reimburse the customer for shipping costs or accept shipments under account or charges upon delivery.
   v. The service format authorized by the Healthcare Service Supervisor will be sent to the customer, which will include an identification number for warranty process control and follow-up, and must be attached in the device package.
   vi. The customer must include all the accessories and cables originally included in the device, unless Healthcare Service Department states otherwise.
   vii. Once the device is at Healthcare Service Department facilities, a thorough technical review will be carried out in order to determine the failure origin.
   viii. Once a diagnosis obtained, and if the warranty conditions are not fulfilled, a written notification must be sent to the customer and perform a technical support procedure according to the HEALTHCARE-015 policy.

ix. If the warranty conditions are fulfilled, the appropriate parts will be requested to the Service Administration Department using the valid format. Notify the customer regarding the estimated time of repair.
x. Once the parts are available, proceed with the replacement, adjustment and any corrective action needed to repair the failure according to the instructions in the service manual and in the training by Canon USA. It is the responsibility of the person who executes the service to procure all the necessary tools (hardware, software, manuals, support material) for the correct execution of the service.
xi. Run all good performance tests according to the operation and service manuals.
xii. Fulfill the service order in the authorized format by the Healthcare Supervisor, sending a copy to the customer for signature and acceptance.
xiii. Properly pack and send the device back to the customer, confirming the adequate reception of it. In case of using parcel service, the shipment must have insurance.

7) Register the service in the Lotus Notes database "BISG HC Support", attaching the service order.
8) Follow up on the device performance within 30 days after the end of the service.
9) Send the satisfaction survey to the customer or dealer.