

Form B03

## Scheduled Maintenance Work Order

Quantum Medical Solutions sdn bhd  
transcending boundaries, transforming life

Format Ref. - QMS/TSD-003 Rev. 00

Work Order No	PO0322036	Scheduled Month	FEBRUARY 2018
Work Order Date	11/2/18	Completed Date	13/2/2018
Clinic Name	KO BUKIT KAMAM	Clinic Code	WPL011
BE No	WPL000200	District	WPL Labuan
BE Category	PETAL <del>HEART</del> DETECTORS	WO Assigned to	MOH D. ASHMANI
Ownership	<input type="checkbox"/> Existing Equipment	<input type="checkbox"/> New & Purchased Equipment	
BE Condition	<input type="checkbox"/> Under Warranty	<input type="checkbox"/> BER Proposed	
Work Order Type	<input type="checkbox"/> Preventive Maintenance (PM) <input type="checkbox"/> Routine Inspection (RI)	<input type="checkbox"/> Third Party Calibration (TPC) <input type="checkbox"/> Statutory Certification (SC)	

## BE Third Party Calibration / Statutory Certification Details

Company Name		Cal / Cert Date	
Contact Number		Cal / Cert Expiry Date	

## Action Taken

→ Do PPM as per Checklist.

## Schedule Maintenance Execution Details

SI No	QMS Engineer / Technician Name	Date	Start Time	End Time
	MOHD. ASHMANI & MOHD. NISHAM BIOMEDICAL TECHNICIAN, QMS 019-2634231	13/2/2018	10.00am	10.15am


## Customer Remarks

--

Engineer / Technician Signature Name Date MOHD. ASHMANI & MOHD. NISHAM BIOMEDICAL TECHNICIAN, QMS 019-2634231 13/2/2018	Customer Signature Name Designation Date Stamp JURAINI AHMAD KETUA JURUPAHAY KESIHATAN URUP LAM: 24420 13.2.2018
---	--

For Internal Use only

First Verification  
QMS Circle Incharge
 JULIUS LIANSUN  
 BIOMEDICAL ENGINEER, QMS  
 019-3620179
Final Verification  
QMS State Incharge
 DICKY LEE  
 SABAH STATE MANAGER  
 QUANTUM MEDICAL SOLUTIONS SDN BHD



### BEMS Planned Preventive Maintenance Checklist

Fetal Heart Detectos.Ultrasonic

BE CODE : 11-696

CHECKLIST NO: CL-070  
REV.000

## PART 6 QUANTITATIVE TASKS

Tick (✓) where appropriate

[illegible]

## PART 7 ELECTRICAL SAFETY TEST

ELECTRICAL SAFETY TEST, (attach report)

(In accordance to IEC 60601)

☐ PASS☐ FAIL☒ NA

## PART 8 NOTES

↳ EST not perform cause unit using battery.

JURAINI AHMAD  
KETUA JURUDAMAT KESIMPATAN USKAP  
1.04.24426

☐ CORRECTIVE MAINTENANCE REQUIRED

 FUNCTIONING☐ NOT FUNCTIONING

WORK ORDER NO. NA

NEXT PPM DATE ▶ 12/2/2019


PPM has been performed in accordance to the checklist and the equipment is functioning to the intended purpose.

COMPLETED BY:

MOND, ASHMAVAT MOND USHAM  
BIOLOGICAL TECHNICIAN, QMS  
010-2584291

DATE : \_\_\_\_\_

13/12/2018



**BEMS Planned Preventive Maintenance Checklist**  
**Fetal Heart Detectos,Ultrasonic**  
BE CODE : 11-696

CHECKLIST NO: CL-070  
REV.000

## PART 6 QUANTITATIVE TASKS

Tick (✓) where appropriate

[illegible]

## PART 7 ELECTRICAL SAFETY TEST

ELECTRICAL SAFETY TEST, (attach report)

(In accordance to IEC 60601)

☐ PASS☐ FAIL☒ NA

## PART 8. NOTES

↳ EST not perform cause unit using Battery.

JURAINI AHMAD  
KETUA JURUSAN KESAMPAHAN USF  
19111940

☐ CORRECTIVE MAINTENANCE REQUIRED

☒ FUNCTIONING

☐ NOT FUNCTIONING

WORK ORDER NO ▶ NA

NEXT PPM DATE ▶ 12/2/2019

PPM has been performed in accordance to the checklist and the equipment is functioning to the intended purpose.

COMPLETED BY:

MOND. APRIL 2 1968  
BIOCHEMICAL PHYS. LAB, QMS  
613-2544391

DATE:

13/12/2018