



# KEMENTERIAN KESIHATAN MALAYSIA

MEET Planned Preventive Maintenance Checklist  
Aspirators  
BE CODE : 10-208

CHECKLIST NO: CL-021-000 REV.000

WORK ORDER NO ▶

## PART 6 QUANTITATIVE TASKS

Tick (✓) where appropriate

No	Description	Units / UOM	Set Values	Measured Values	Limit/Tolerance	PASS	FAIL	NA
	Vacuum level (set at maximum vacuum)	mmHg				( )	( )	( )

## PART 7 ELECTRICAL SAFETY TEST

ELECTRICAL SAFETY TEST, (attach report)

Tick (✓) where appropriate

Standard use :

☐ IEC 60601

☐ IEC 61010

☐ IEC 62353

Result :

☒ PASS

☐ FAIL

☐ NA

## PART 8 NOTES

- ppm done.

☐ CORRECTIVE MAINTENANCE REQUIRED

☒ FUNCTIONING

☐ NOT FUNCTIONING

WORK ORDER NO ▶

NEXT PPM DATE ▶ 11/2019

PPM has been performed in accordance to the checklist and the equipment is functioning to the intended purpose.

COMPLETED BY:

FOSTER ANAK ENKASAN  
QUANTUM BIOMEDICAL ENGINEERING  
KUCHING SARAWAK.

DATE :

15/11/18