



# KEMENTERIAN KESIHATAN MALAYSIA

MEET Planned Preventive Maintenance Checklist  
Sterilization Process Indicators, Biological  
BE CODE : 13-732

CHECKLIST NO: CL-146-000  
REV.000

WORK ORDER NO ▶

## PART 6 QUANTITATIVE TASKS

Tick (✓) where appropriate

No	Description	Units / UOM	Set Values	Measured Values	Limit/Tolerance	PASS	FAIL	NA
	Temperature reading	°C				( )	( )	( )

## PART 7 ELECTRICAL SAFETY TEST

ELECTRICAL SAFETY TEST, (attach report)

Tick (✓) where appropriate

Standard use :

☐ IEC 60601

☐ IEC 61010

☐ IEC 62353

Result :

☒ PASS

☐ FAIL

☐ NA

## PART 8 NOTES

- ppm done.

☐ CORRECTIVE MAINTENANCE REQUIRED

☒ FUNCTIONING

☐ NOT FUNCTIONING

WORK ORDER NO ▶

NEXT PPM DATE ▶ 11/2019

PPM has been performed in accordance to the checklist and the equipment is functioning to the intended purpose.

COMPLETED BY:

FOSTER ANAK ENKASAN  
QUANTUM BIOMEDICAL ENGINEERING  
KUCHING SARAWAK.

DATE: 27/11/19