

**KEMENTERIAN KESIHATAN MALAYSIA**

MEET Planned Preventive Maintenance Checklist  
Sterilizing units, Steam  
BE CODE : 13-746

CHECKLIST NO: CL-148-000 REV.001

WORK Order No

**PART 6 SERVICE CHECKLIST**

No	Description	Checked	Service/ Clean	Replaced	Not Applicable	No	Description	Checked	Service/ Clean	Replaced	Not Applicable
1	Fuses	( / )	( )	( )	( )	16	Discharge Valve	( / )	( )	( )	( )
2	Main Control Board	( / )	( )	( )	( )	17	Infill Valve	( / )	( )	( )	( )
3	Control Board	( / )	( )	( )	( )	18	Heating Element	( / )	( )	( )	( )
4	Fan Running Rear & Side	( / )	( )	( )	( )	19	Heat Exchanger/Cooling system (if applicable)	( / )	( )	( )	( )
5	Water Reservoir Float Switch/Water Level Sensor	( / )	( )	( )	( )	20	Vacuum Pump	( / )	( )	( )	( )
6	Water Reservoir	( / )	( )	( )	( )	21	Door Gasket	( / )	( )	( )	( )
7	Electric Door Lock	( / )	( )	( )	( )	22	Chamber Guard Seal (if applicable)	( / )	( )	( )	( )
8	Pressure Door Lock	( / )	( )	( )	( )	23	Solid State Relay	( / )	( )	( )	( )
9	Discharge Door Lock	( / )	( )	( )	( )	24	Thermostat Assembly	( / )	( )	( )	( )
10	Pressure Transducer	( / )	( )	( )	( )	25	Temperature Sensor	( / )	( )	( )	( )
11	Pressure Gauge	( / )	( )	( )	( )	26	Thermal Printer (if applicable)	( / )	( )	( )	( )
12	Safety Valve	( / )	( )	( )	( )	27	Vacuum Pump (if applicable)	( / )	( )	( )	( )
13	Air Valve/Steam Solenoid Valve	( / )	( )	( )	( )	28	Filters / strainer	( / )	( )	( )	( )
14	Air Valve/Steam Solenoid Valve	( / )	( )	( )	( )						
15	Air Valve/Steam Solenoid Valve	( / )	( )	( )	( )						

**PART 7 Technician Remarks**

- Functioning good, ppm done