



**Pratik Tamang**

**2214130**

**Infrared Sugar Level Tester**

BSc (Hons) Computer Science and Software Engineering

Contextual Report University of Bedfordshire

Undergraduate Final Thesis Report

Supervisor: Pawan Kc

AY 2025



## ASSIGNMENT TOP SHEET

Faculty of Creative Arts, Technologies & Science

Department of Computer Science & Technology

Student Ref. No 2214130

|                                                                        |                              |                                                               |
|------------------------------------------------------------------------|------------------------------|---------------------------------------------------------------|
| Unit Name: Undergraduate Project                                       |                              | Unit Code: CIS017-3                                           |
|                                                                        |                              | Deadline for Submission(s)<br>12 <sup>th</sup> December, 2025 |
| Student's Surname<br>Tamang                                            | Student's Forename<br>Pratik |                                                               |
| Unit Leader's Name: Ajaya Kumar Sharma                                 |                              | Supervisor: Pawan KC                                          |
| Assignment Details:<br>Assessment 2: Undergraduate Final Thesis Report |                              |                                                               |

### Instructions to Student:

Please note: Work presented in an assessment must be the student's own. Plagiarism is where a student copies work from another source, published or unpublished (including the work of a fellow student) and fails to acknowledge the influence of another's work or to attribute quotes to the author. Plagiarism is an academic offence.

Work presented in an assessment must be your own. Plagiarism is where a student copies work from another source, published or unpublished (including the work of another student) and fails to acknowledge the influence of another's work or to attribute quotes to the author. Plagiarism is an academic offence and the penalty can be serious. The University's policies relating to Plagiarism can be found in the regulations at <http://www.beds.ac.uk/aboutus/quality/regulations>. To detect possible plagiarism we may submit your work to the national plagiarism detection facility. This searches the Internet and an extensive database of reference material including other students' work to identify. Once your work has been submitted to the detection service it will be stored electronically in a database and compared against work submitted from this and other universities. It will therefore be necessary to take electronic copies of your materials for transmission, storage and comparison purposes and for the operational back-up process. This material will be stored in this manner indefinitely.

I have read the above information and I confirm that this work is my own and that it may be processed and stored in the manner described.

Signature (Print Name): PRATIK TAMANG

Date: 12<sup>th</sup> december, 2025

Extension deadline

CAAS agrees that the assignment may be submitted \_\_\_\_ days after the deadline and should be marked without penalty.

CAAS confirmation.....

Please leave sufficient time to meet this deadline and do not leave the handing-in of assignments to the last minute. You need to allow time for any system problems or other issue.

## Abstract

The study paper focuses on the use of IoT And AI to develop a non-invasive infrared sugar level tester that uses near-infrared spectroscopy and Partial Least Squares Regression(PLSR) to measure blood sugar levels in real time. Diabetes control requires frequent glucose monitoring, but traditional methods are invasive, painful, and inconvenient. This research explores the possibility of using NIR light absorption to predict glucose concentration without requiring blood sampling. Traditional glucose measurement techniques include finger-prick tests and continuous glucose monitoring involves invasive procedures leading to infection, high cost, and discomfort. This System uses NIR LED wavelength ranging from 500nm to 900nm and sensors to collect glucose particles from the blood. The collected data is processed using PLSR a machine learning-based technique to predict glucose concentration. The processed data is sent to the ESP32 microcontroller where the data is displayed on an OLED/LCD. The ESP32 is coded in C++ while Python handles PLSR-based glucose prediction. This project aims to develop a portable and non-invasive glucose monitoring device that minimizes dependence on painful invasive methods while improving accessibility and conscience for diabetic patients.

Keywords: ESP32, NIR-SPECTROSCOPY, NON-Invasive Glucose Level Tester, IoT, AI, PLSR

## Acknowledgment

I would like to whole heartedly thank all the individuals who ensured the timely and successful completion of the research.

Above all, I would like to extend my sincere appreciation to my supervisor, Mr Pawan Kc, for his time, dedication, and direction, and for the boosting comments as well. His dedication and helpful comments have been crucial for the project to take the shape it has now taken.

I would also like to express my sincere gratitude to our unit coordinator Mr Ajaya Sharma and our lecturer Mr Pawan Kc for the generous guidance and support given to me for my studies. They helped get all my academic reports ready, including this one.

I would like to utilize this opportunity to thank from the bottom of my heart to the University of Bedfordshire and my college Patan College For Professional Studies for the opportunity to pursue my academic goals, the available resources, and the serene academic atmosphere necessary to conduct this research. Heartfelt thanks to all the faculty members for their constant and timely support and for the trailblazing path of innovation and research.

I would like to express my gratitude to all the respondents who answered my survey for the infrared sugar level tester with their responses and opinions. Response provided valuable data for establishing the acceptability and feasibility of the non-invasive system for blood glucose monitoring.

Most importantly, I must acknowledge my seniors, colleagues, and fellow students who were with me throughout for relentless inspiration, healthy arguments, and unstinting encouragement in times of need.

Pratik Tamang

2214130

BSC(Hons)CSSE

## Table of Contents

|                                                                                                                                                      |    |
|------------------------------------------------------------------------------------------------------------------------------------------------------|----|
| Abstract.....                                                                                                                                        | I  |
| Acknowledgment.....                                                                                                                                  | II |
| Chapter 1: Introduction.....                                                                                                                         | 1  |
| 1.1 Background.....                                                                                                                                  | 1  |
| 1.2 Problem Statement.....                                                                                                                           | 2  |
| 1.3 Proposed Solution.....                                                                                                                           | 2  |
| 1.4 Aims and Objectives.....                                                                                                                         | 3  |
| Aims 3                                                                                                                                               |    |
| Objectives.....                                                                                                                                      | 3  |
| 1.5 Feasibility Study.....                                                                                                                           | 4  |
| Technical Feasibility.....                                                                                                                           | 4  |
| Financial Feasibility.....                                                                                                                           | 4  |
| Operational Feasibility.....                                                                                                                         | 4  |
| Environmental Feasibility.....                                                                                                                       | 5  |
| 1.6 Intellectual Challenge.....                                                                                                                      | 5  |
| Chapter 2: Project Plan.....                                                                                                                         | 6  |
| 2.1 Gantt Chart.....                                                                                                                                 | 6  |
| 2.2 Risk Analysis.....                                                                                                                               | 7  |
| 2.3 Work Flow Diagram.....                                                                                                                           | 8  |
| Chapter: 3 Literature Review.....                                                                                                                    | 9  |
| 3.1 Overview/Background.....                                                                                                                         | 9  |
| 3.2 Review of existing project.....                                                                                                                  | 9  |
| 3.2.1 Non-Invasive Blood Glucose Monitoring Using NIR Spectroscopy (Smith et al., 2021).....                                                         | 9  |
| 3.2.2 PLSR-Based Predictive Model for Glucose Estimation (Chen & Li, 2020).....                                                                      | 10 |
| 3.3 Literature Review.....                                                                                                                           | 10 |
| Key Terms.....                                                                                                                                       | 10 |
| 3.3.1 Advances in NIR Spectroscopy for Blood Sugar Measurement (Gupta et al., 2019).....                                                             | 11 |
| 3.3.2 Comparative Study of Chemometric Models for NIR Glucose Detection (Rahman et al., 2022).....                                                   | 12 |
| 3.3.3 Real-Time Blood Glucose Monitoring Using IoT-Enabled NIR Sensors (Kumar & Patel, 2023).....                                                    | 13 |
| 3.3.4 Design of Non-Invasive Glucose Meter Using Near-Infrared Technique.....                                                                        | 13 |
| 3.3.5 Evaluation of a Novel Noninvasive Blood Glucose Monitor Based on Mid-Infrared Quantum Cascade Laser Technology and Photothermal Detection..... | 14 |
| 3.3.6 Development of a Noninvasive Blood Glucose Monitoring System Prototype: Pilot Study.....                                                       | 14 |

|                                                                                                                                                           |    |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------|----|
| 3.3.7 An Infrared Non-Invasive System for Measuring Blood Glucose: A Primary Study Using Serum Samples.....                                               | 15 |
| 3.3.8 Noninvasive Blood Glucose Monitoring Systems Using Near-Infrared Technology—A Review .....                                                          | 15 |
| 3.3.9 Non-Invasive Glucose Monitoring Using NIR Spectroscopy.....                                                                                         | 16 |
| 3.3.10 Development of Spectroscopic Sensor System for an IoT Application of Adulteration Identification on Milk Using Machine Learning.....               | 16 |
| 3.3.11 Non-Invasive Blood Glucose Monitoring Using PPG Signals with Machine Learning .....                                                                | 17 |
| 3.3.12 Application and Improvement of Partial-Least-Squares in Fourier Transform Infrared Spectroscopy. ....                                              | 17 |
| 3.3.13 Optimization of Partial Least Squares Regression Calibration Models in Near-Infrared Spectroscopy: A Novel Algorithm for Wavelength Selection..... | 18 |
| 3.4 Summary.....                                                                                                                                          | 18 |
| 3.5 Justification .....                                                                                                                                   | 19 |
| Chapter 4: Primary Research.....                                                                                                                          | 21 |
| 4.1 Objective of primary research .....                                                                                                                   | 21 |
| 4.1.1 Market Research .....                                                                                                                               | 21 |
| 4.2 Survey Questionnaire.....                                                                                                                             | 22 |
| 4.2.1 Survey Data Analysis.....                                                                                                                           | 24 |
| 4.3 Data Collection and Analysis .....                                                                                                                    | 30 |
| 4.3.1 Data Collections.....                                                                                                                               | 31 |
| Chapter 5: Artefact Planning .....                                                                                                                        | 33 |
| 5.1 Requirement Analysis .....                                                                                                                            | 33 |
| 5.1.1 Functional Requirements .....                                                                                                                       | 33 |
| 5.1.2 Non-Functional Requirements .....                                                                                                                   | 33 |
| 5.1.3 Usability Requirements.....                                                                                                                         | 34 |
| 5.2 Software Requirement.....                                                                                                                             | 35 |
| 5.3 Hardware requirement.....                                                                                                                             | 35 |
| 5.3.1 Development Requirements.....                                                                                                                       | 35 |
| 5.3.2 Deployment Requirement .....                                                                                                                        | 35 |
| 5.4.2 Activity Diagram .....                                                                                                                              | 37 |
| 5.4.3 DB Schema .....                                                                                                                                     | 37 |
| Chapter 6: Testing and Evaluation Strategy.....                                                                                                           | 38 |
| 6.1 Testing Strategy/Plan.....                                                                                                                            | 38 |
| 6.1.1 Test Case .....                                                                                                                                     | 38 |
| 6.2 Evaluation Strategy.....                                                                                                                              | 40 |
| Chapter 7: Development .....                                                                                                                              | 42 |
| 7.1 Methodology .....                                                                                                                                     | 42 |

|                                                                     |    |
|---------------------------------------------------------------------|----|
| 7.2 Tools and technologies used.....                                | 43 |
| 7.2.1 Hardware Implementation .....                                 | 43 |
| 7.2.2 Tech Stack and Development Tools .....                        | 43 |
| 7.3 Risk Encountered.....                                           | 43 |
| 7.4 Implementation Process.....                                     | 45 |
| 7.5 Data Collection and Pre-Processing .....                        | 46 |
| 7.6 Ai Model Training and Optimization .....                        | 47 |
| 7.7 Backend and Api Development.....                                | 49 |
| 7.8 Front End Development and Real Time Monitoring.....             | 50 |
| 7.9 Challenges and solutions .....                                  | 51 |
| Chapter 8: Testing.....                                             | 53 |
| 8.1 Component Tests (Hardware & Integration) .....                  | 54 |
| 8.2 Data Preprocessing Tests (Software Logic) .....                 | 55 |
| 8.3 Partial Least Squares Regression (PLSR) Model Test .....        | 56 |
| 8.4 System-Level Validation and Reliability Tests .....             | 56 |
| Chapter 9: Evaluation .....                                         | 61 |
| 9.1 User Evaluation .....                                           | 61 |
| 9.1.1 Appreciations (Strengths and User Experience).....            | 62 |
| 9.1.2 Criticisms (Improvement Areas and Adoption Barriers).....     | 62 |
| Chapter 10: Critical Analysis .....                                 | 64 |
| Chapter 11: Limitation and Future Scope .....                       | 66 |
| 11.1 Limitation: .....                                              | 66 |
| 11.1.1 Hardware Constraints (Standard vs. Specialized Optics) ..... | 66 |
| 11.1.2 Data and Accuracy Issues.....                                | 66 |
| 11.2 Future Scope:.....                                             | 67 |
| 11.2.1 Hardware Upgrades (High-Fidelity Sensing) .....              | 67 |
| 11.2.2 Dataset Creation and Algorithms .....                        | 67 |
| 11.2.3 Mechanical Design .....                                      | 68 |
| Chapter 12- Conclusion.....                                         | 69 |
| References .....                                                    | 71 |
| Appendix .....                                                      | 78 |
| Project Proposal .....                                              | 78 |
| Ethics Form.....                                                    | 84 |
| Weekly Reports from UG.....                                         | 86 |
| Testing Results .....                                               | 93 |
| Model Training.....                                                 | 94 |
| Testing Code Snippets.....                                          | 97 |

|                                   |     |
|-----------------------------------|-----|
| Testing Results and Formula ..... | 101 |
| Project Poster:.....              | 103 |

## Table of Figures:

|                                                             |    |
|-------------------------------------------------------------|----|
| Figure 1: Sem-1 RMET Gantt Chart.....                       | 6  |
| Figure 2: Sem-1 UG Gantt Chart.....                         | 6  |
| Figure 3: Sem-2 UPdated Gantt Chart.....                    | 7  |
| Figure 4: Work Flow Diagram.....                            | 8  |
| Figure 5: Age group.....                                    | 24 |
| Figure 6: Diabetes Status.....                              | 25 |
| Figure 7: Sugar Consumption.....                            | 25 |
| Figure 8: Blood Monitor.....                                | 26 |
| Figure 9: Medication use.....                               | 26 |
| Figure 10: Sugar Check.....                                 | 27 |
| Figure 11: Intrest in non invasive method.....              | 28 |
| Figure 12: Wereable preferred.....                          | 28 |
| Figure 13: Concern.....                                     | 29 |
| Figure 14: Accuracy Expected.....                           | 29 |
| Figure 15: Beta Testing.....                                | 30 |
| Figure 16: Additional Features.....                         | 30 |
| Figure 17: Use case.....                                    | 36 |
| Figure 18: Activity Diagram.....                            | 37 |
| Figure 19: Database.....                                    | 37 |
| Figure 20: Final Implemented Device.....                    | 45 |
| Figure 21: Pre Processing.....                              | 48 |
| Figure 22: Mean Regression.....                             | 48 |
| Figure 23: Algorithm.....                                   | 49 |
| Figure 24: Cofficient Extraction.....                       | 49 |
| Figure 25: Extracted Cofficient.....                        | 50 |
| Figure 26: LCD Display.....                                 | 51 |
| Figure 27: Sensor Testing Code.....                         | 53 |
| Figure 28: Sensor Test Result.....                          | 54 |
| Figure 29: No finger in Sensor<br>Placed In the sensor..... | 58 |
| Figure 30: Finger<br>Placed In the sensor.....              | 58 |
| Figure 31: No button Placed.....                            | 59 |
| Figure 32: Button Pressed.....                              | 59 |
| Figure 33: Finger Placed.....                               | 59 |
| Figure 34: glucode Measure<br>.....                         | 59 |
| Figure 35: Finger Removed while measuring.....              | 59 |
| Figure 36: Test Case Result.....                            | 60 |

|                                         |    |
|-----------------------------------------|----|
| Table 1: Risk Table                     | 8  |
| Table 2: Survey Question.....           | 24 |
| Table 3: Functional requirement.....    | 33 |
| Table 4:Non Functional Requirement..... | 34 |
| Table 5:Usability Requirement .....     | 35 |
| Table 6: Test Case.....                 | 40 |
| Table 7: Hardware Test.....             | 55 |
| Table 8: Data Preprocessing Test .....  | 55 |
| Table 9: Plsr Model Test.....           | 56 |
| Table 10: System Test.....              | 58 |
| Table 11: User Evaluation .....         | 62 |

## Abbreviations

IoT: Internet of Things

AI: Artificial Intelligence

ML: Machine Learning

PLSR: Partial Least Square Regression

IR: Infrared

NIR: Near-Infrared

RMSECV - Root Mean Square Error of Cross-Validation

RMSEP - Root Mean Square Error of Prediction

OGTT - Oral Glucose Tolerance Test

SVM - Support Vector Machine

ANN - Artificial Neural Network

SAE - Stacked Auto-Encoder

$R^2$  - Coefficient of Determination (used for regression accuracy)

R - Correlation Coefficient

## Chapter 1: Introduction

### 1.1 Background

Diabetes mellitus has become a significant global health problem over the past several decades. The occurrence of diabetes-related complications, such as cardiovascular disease, renal failure, and neuropathy, can be significantly reduced by frequent monitoring and maintaining blood glucose levels within the normal range. Most commercially available glucose measurement systems, however, are invasive or minimally invasive. Invasive methods, i.e., finger-pricking, are not only painful but also inconvenient for frequent testing, while minimally invasive devices are usually limited by accuracy and stability. This necessitates an inexpensive, small-sized, painless, and easy-to-use non-invasive glucose monitoring system that can promote frequent testing, which will ultimately benefit people in managing and controlling their blood glucose levels(Edward, 2022).

The project intends to leverage the newest breakthroughs in IoT and AI technologies to create an infrared-based sugar tester to bridge this gap. Infrared spectroscopy is employed by this accurate, real- time, non-invasive glucose monitoring system for determining the level of blood sugar without the drawing of blood or without a finger prick. Machine learning algorithms sort the data that the device detects, thereby enhancing the accuracy of the system. Accurate measurement and more tailored diabetes care are provided through the adjustment of the AI model for individual differences in the level of blood glucose. The project intends to make diabetes easier and more comfortable for users by being less expensive and invasive than prevailing methods of glucose testing.

However, such technologies are costly and are not easily accessible to individuals, especially in resource-constrained environments. Therefore, the project is interested in identifying low-cost, scalable, and easily accessible technologies for non-invasive glucose monitoring. The motivation behind the research is developing a practical and easily adoptable IoT and AI-based system for real-time non- invasive glucose monitoring. Within the affordability and accessibility constraints, the project aims to leverage low-cost, simple, yet effective components to create actionable insights for users. Through the ongoing update of the AI model against data collection, the system enhances the accuracy and reliability of the glucose measurement, making it a worthy

alternative for users who wish to measure blood glucose levels without using invasive tests.(Wilson and Baietto, 2011)

## 1.2 Problem Statement

Current methods for blood glucose monitoring, particularly invasive methods like finger-pricking, are a cause of significant discomfort to users. For diabetic patients or those at risk of diabetes, frequent blood testing is typically necessary in order to monitor their condition. Repeated testing is both a physical and emotional burden, imposing stress and discomfort that dissuades frequent monitoring. As a result, the majority of patients do not achieve desired glucose levels, and the threat of complications like heart disease, kidney damage, and nerve damage is increased.(Aminah Hina, 2022)

In non-invasive glucose monitoring, newer technologies like the Internet of Things (IoT) and Near- Infrared (NIR) spectroscopy are also developing fast and offering promising solutions. IoT allows real- time data collection and monitoring of sensors, and the data—that is, continuous glucose level—can be remotely accessed or through connected devices.(Al-Fuqaha et al., 2015). Because physicians or caregivers can see a patient's trend of glucose and react accordingly without multiple hospital visits, this feature allows more personalization of care. In addition, there is a new method that is referred to as NIR spectroscopy measures blood glucose without blood extraction through infrared light. By measuring the light absorption properties of the skin to compute blood glucose levels, NIR technology offers a painless alternative to others.(Yadav et al., 2015).

## 1.3 Proposed Solution

The suggested solution would investigate the viability of an affordable and painless blood glucose monitoring system through infrared spectroscopy and AI and IoT technologies. The system will involve the use of Near-Infrared (NIR) spectroscopy, an infrared radiation methodology that measures glucose from under the skin and without need to pierce the finger or extract blood. The reading from the NIR sensor will be analyzed by an abbreviated AI model that adds more data exposure to glucose predictions. The system would include IoT capability to supplement real-time monitoring and transmission and provide the subject with instant glucose level feedback.

This is corroborated by current studies to address the feasibility and possibility of

utilizing low-cost and non-invasive glucose monitoring devices. The efficacy and effectiveness in applying NIR spectroscopy to quantify glucose as an alternative way to achieve this in a non-invasive way with the support of advanced signal processing techniques and machine learning algorithms were illustrated by Almeida et al. (2018), and Jain and Banerjee (2020). The addition of predictive artificial intelligence algorithms will further enhance the performance with learning from real-time input and adjusting to inter-personal variance in glucose level.

Through an emphasis on cheap and scalable devices, the system avoids the use of costly and invasive equipment and is cheaper among those who live in resource-constrained environments. The use of simple and efficient AI models and IoT syncing allows glucose monitoring under real-time and non-invasive conditions with better quality of life among diabetics. The system provides an easy, efficient, and cheap process of blood glucose regulation with efficient treatment for diabetes and prevention against its complications.

## 1.4 Aims and Objectives

### Aims

The main aim of the research is to create a non-invasive real-time blood glucose monitoring system using Near-Infrared (NIR) Spectroscopy and Partial Least Squares Regression (PLSR).

### Objectives

- Develop an affordable method of non-invasive glucose monitoring based on NIR spectroscopy and machine learning algorithms.
- Design and implement a full glucose monitoring system utilizing NIR sensors and a microcontroller (Arduino) to acquire and process real-time data.
- Utilize the PLSR algorithm to predict NIR absorption values and the concentration of blood glucose precisely.
- Develop and design a real-time display module to display glucose concentration levels based on sensor inputs
- Transcend the barriers of noise, environmental variation, and inter-person differences to provide stable and reliable glucose readings.
- Create a simple user interface to display glucose levels and provide ease of use

in running the device.

## 1.5 Feasibility Study

Financial, technical, and operational feasibilities must be taken into consideration for the effective development and execution of the Infrared Sugar Level Tester. In this project, as the available means were constrained to Near-Infrared (NIR) spectroscopy and Partial Least Squares Regression (PLSR) only, consideration of the feasibility of meeting the intended objectives of accurate, non-destructive sugar level determination is necessary.

### Technical Feasibility

The system proposed utilizes traditional NIR spectroscopy methodologies combined with PLSR for predictive modeling, allowing for high accuracy and reliability of sugar content measurement. NIR- based sensing has been shown by research works to be an effective and popular technique of ascertaining sugar content in food and agricultural produce (Bittner et al., 2020). Additionally, coupling with machine learning algorithms for the analysis of data improves its predictive nature, and it can be utilized for continuous monitoring of sugar levels in real-time. Utilization of inexpensive spectrometers and open-source computational software such as MATLAB renders it cost-effective without compromising accuracy.

### Financial Feasibility

One of the main obstacles in spectroscopy-based testing is the cost of high-precision sensors and spectrometers. However, this project aims to bypass this obstacle by utilizing low-cost NIR sensors and open-source data analysis software, which significantly reduces the cost. It is evident from the available literature that the application of such low-cost yet effective spectroscopy systems can bring high returns by preventing expensive laboratory-based sugar testing (Li et al., 2021). Large-scale production and advancements in optical sensor design can also reduce the cost further, increasing the commercial viability of the system.

### Operational Feasibility

System design focuses on user-friendly implementation to enable easy use by end users like food manufacturers, farmers, and researchers. The non-destructive nature of NIR spectroscopy allows rapid and repetitive testing without affecting the composition

of samples. Model training and initial calibration do require technical expertise, though, and that can prove to be a drawback. Automated calibration methods and pre-trained models can be added for that to reduce user input.(Pasquini, 2018). Portable device configurations also allow easy deployment in diverse environments, and the system can be scalable to different industry needs.(Li et al., 2020).

### Environmental Feasibility

The proposed system is in line with sustainable food processing because it reduces the dependency on chemical-based sugar testing, which has a tendency to generate waste and uses toxic reagents. NIR- based analysis is a green technique that minimizes sample destruction and the use of resources. Furthermore, accurate monitoring of sugar content helps in the optimization of food production processes, reduces waste, and improves efficiency in general in the food and agricultural industries (FAO, 2019). This aids in realizing a more sustainable and resource-efficient industry without sacrificing high-quality food standards.

### 1.6 Intellectual Challenge

The design of an infrared sugar level tester entails some intellectual hurdles that need to be overcome to enable successful development. One of the greatest hurdles is the accurate calibration of NIR sensors for proper measurement of sugar levels in different types and conditions of samples. Environmental fluctuations in temperature, humidity, and sample texture can affect spectral measurements, and therefore, complex compensation techniques (Gowen et al., 2015). The application of Partial Least Squares Regression (PLSR) also entails the acquisition and processing of large datasets to train models, which can be computationally intensive and prone to overfitting if not well managed (Zhang et al., 2020). Another major challenge is the design of a system that can be adapted to suit different food and agricultural products, as different sugar compositions may require different spectral preprocessing techniques . Integration of the NIR system with an interface that is user- friendly and can be operated by non-experts is another usability challenge, where user-friendly software design and automatic analysis functions are necessary. Lastly, ensuring regulatory compliance with food quality and safety standards is another challenge, which necessitates strenuous validation and testing (FAO, 2019). Overcoming these intellectual challenges is key to the development of a reliable and widely acceptable infrared sugar level tester.

## Chapter 2: Project Plan

### 2.1 Gantt Chart




Figure 1: Sem-I RMET Gantt Chart




Figure 2: Sem-I UG Gantt Chart




Figure 3: Sem-2 UPdated Gantt Chart

## 2.2 Risk Analysis

Risk analysis involves the identification and analysis of the possible risk factors or issues that can affect the realization of the Infrared Sugar Level Tester. The process looks into the likelihood and impact of the risks and considers the necessary mitigation measures to allow the accuracy, efficiency, and reliability of the system. The following are the risks and mitigation measures identified:

| Risk                           | Impact | Likelihood | Mitigation Strategies                                                                            |
|--------------------------------|--------|------------|--------------------------------------------------------------------------------------------------|
| Sensor Accuray Issues          | High   | High       | Use high-quality IR sensors.                                                                     |
| Environmental Interference     | Medium | Medium     | Shield sensors from external factors (e.g., temperature/humidity); use noise-filtering software. |
| Arduino Processing Limitations | High   | High       | Consider using higher-capacity microcontrollers like ESP32.                                      |

|                                               |         |        |        |                                                                                                                                                               |
|-----------------------------------------------|---------|--------|--------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Incorrect Calibration                         | Initial | High   | High   | Calibrate the device using a large dataset of real blood glucose samples across diverse populations (e.g., different skin types, ages, and hydration levels). |
| Continuous NIR LED Operation Draining Battery |         | Medium | Medium | Optimize LED usage with pulse modulation rather than continuous operation.                                                                                    |

Table 1: Risk Table

## 2.3 Work Flow Diagram




Figure 4: Work Flow Diagram

## Chapter: 3 Literature Review

### 3.1 Overview/Background

Recent innovations in non-invasive medical technology have opened up new possibilities for monitoring physiological parameters like blood glucose levels without the use of traditional invasive techniques like finger pricking. Of the technologies, the use of infrared (IR) spectroscopy has shown significant promise in developing an infrared sugar level tester. The technology is founded on near- infrared (NIR) spectroscopy, which measures the absorption of light by glucose molecules in the skin and provides real-time, non-invasive glucose monitoring. With the use of this technology, a more user- friendly and convenient system for monitoring blood sugar can be developed, providing individuals, particularly diabetics, with an easier way to control their condition.

The combination of NIR spectroscopy and artificial intelligence (AI) models, i.e., Partial Least Squares Regression (PLSR), provides precise and robust prediction of blood glucose levels from spectral information. Such systems are capable of analyzing data from low-cost infrared sensors and hence are a cheaper alternative to conventional methods of determination of glucose. The developed infrared sugar level tester is aimed at precise, non-invasive determination of glucose with the prospect of usage on a large scale in clinical as well as domestic settings.

### 3.2 Review of existing project

#### 3.2.1 Non-Invasive Blood Glucose Monitoring Using NIR Spectroscopy

(Smith et al., 2021)

The research investigated the use of Near-Infrared (NIR) spectroscopy for non-invasive measurement of blood glucose. The researchers applied NIR wavelengths (900-2500 nm) in measuring the glucose level in humans. The research concluded that the most precise algorithm was PLSR (Partial Least Squares Regression) with  $R^2 = 0.92$ , which was very reliable for non-invasive glucose measurement. The research concluded interference from skin pigmentation and water content, which may affect accuracy. This study validates my method by demonstrating PLSR to be effective in NIR-based glucose sensing ( $R^2 = 0.92$ ). However, it also introduces complications like interference from skin pigmentation and hydration that I must take into consideration while calibrating

my instrument for sensing sugar levels.

Limitations:

Although with promising findings, some of those studies were also noted to have some constraints of studies that may hinder their ready application under real-world conditions. Some of those notable ones were interferences caused by skin pigmentation and water content variability that lead to errors in glucose measurements. There was hardly any regard to interference by ambient conditions like temperature and humidity with tendencies to interfere with NIR absorption. Subsequent studies can incorporate adaptive calibration techniques and population trials to enhance precision.

### 3.2.2 PLSR-Based Predictive Model for Glucose Estimation (Chen & Li, 2020)

This work focused on the calibration performances for glucose prediction using NIR spectra with PLSR models. The comparison of the performances using PLSR and classical regression models, as reported, indicated that PLSR performed better with a Mean Absolute Error of 0.15 mmol/L. The work strongly emphasized the application of data preprocessing techniques including the baselines correction and the normalisation, which enhanced the model performances dramatically. I endorse my decision to employ PLSR in MATLAB, with the best fit (MAE = 0.15 mmol/L). This also demonstrates the need for normalisation and baseline correction, which I must implement to achieve better spectral preprocessing.

Limitations:

A limitation here, though, is that it used laboratory conditions that do not reflect real-world variability. The study neglected real-world glucose variability, subject-to-subject physiological variability, and environmental conditions such as skin moisture and body temperature. The study neglected hybrid models that could further improve prediction accuracy. Future work will require real-world testing with heterogeneous subjects and addressing integration with other pre-processing techniques.

## 3.3 Literature Review

### Key Terms

#### 1. Near-Infrared(NIR)Spectroscopy

NIR spectroscopy uses 780–2500 nm light for non-invasive examination of materials. It measures light absorption in proportion to glucose concentration. Quicker and non-harmful compared to most other techniques, it is omnipresent

in body-worn glucose monitors. It may be affected by skin color, water content, and tissue variation, requiring effective calibration and data processing to reach a proper outcome.

## 2. Partial Least Squares Regression (PLSR)

PLSR is a suitable method of regression for highly correlated variables in a dataset. It is applied in NIR glucose analysis to identify the latent factors describing glucose concentration. It offers better prediction accuracy as well as the ability to cope with noise, particularly in conjunction with data preprocessing. Extensions such as MWPLS as well as RPLS enhance model performance.

## 3. IoT (Internet of Things)

IoT links devices to the internet for real-time data exchange. In glucose monitoring, NIR sensors with IoT send measurements to apps or cloud servers, facilitating remote monitoring as well as alertness. This adds to user convenience as well as physician accessibility. Challenges involve maintaining data privacy, sensor reliability, as well as seamless communication among devices in different environments.

## 4. Calibration and Preprocessing Techniques

IoT interconnects devices over the internet for real-time information exchange. In glucose tracking, NIR sensors coupled with IoT send readings to apps or cloud services, promoting distant tracking and alerts. This is convenient for users as well as for physician accessibility. Challenges are ensuring information privacy, sensor reliability, as well as seamless communication of devices in varied settings.

## 5. Machine Learning & Hybrid Models

Machine learning refines glucose prediction from NIR data by detecting complicated patterns. Popular methods include artificial neural networks and SVMs, often in hybrid models with PLSR or deep networks. These models deliver better accuracy at the cost of possibly suffering from overfitting as well as high computational requirements, particularly in real-time wearable devices.

### 3.3.1 Advances in NIR Spectroscopy for Blood Sugar Measurement (Gupta et al., 2019)

The research covered the latest developments in NIR-based glucose monitoring technologies with an emphasis on the development of sensors. The authors provided an

explanation on the miniaturized NIR sensors embedded in Internet-of-Things devices that allowed the monitoring of real-time glucose levels. The research, according to the authors, also incorporated that the coupling with AI-based selection and PLSR enhanced the accuracy of the model but noted that calibration against different skin types remains an issue. This work introduces miniaturized NIR sensors and AI-based feature selection, which align with my goal to improve the accuracy of the prediction. This work also addresses the issue of calibration with the sensors, a most significant issue that I must address in my work.

#### Limitations:

The central failing of this research is that there is no experimental evidence of the suggested sensor innovations. Although it suggests theoretical innovations, there is no rigorous real-world testing to support the fact that the miniaturized sensors work in varied conditions. Sensor calibration problems with varying skin types and physiological conditions are also not addressed. More research needs to be done to develop more precise sensors, enhance calibration processes, and analyze the reliability of IoT- based glucometers.

### 3.3.2 Comparative Study of Chemometric Models for NIR Glucose Detection (Rahman et al., 2022)

In this research, the authors compared PLSR, SVM, and ANN for glucose detection based on NIR. PLSR worked better with low-noise conditions, and ANN worked better with nonlinear variation. The authors suggested the use of a hybrid model that integrates PLSR with deep learning so that the system becomes more efficient with different datasets. This research indicates that although PLSR performs optimally under low-noise conditions, that ANN is less forgiving to non-linearity. This leads me to consider hybrid models to achieve higher accuracy in my sugar level reader.

#### Limitations:

Even though it contained some valuable information, the research was not perfect. The authors were more concerned with model performance without regard to computation efficiency, a very basic issue in real-world applications. The ANN models, as effective as they were at their function, have high computational requirements and overfit if trained using small datasets. The research did not include mention of what effect environmental conditions such as skin moisture and ambient interference would have on model precision. Future research will have to make it their agenda to find a balance

between hybrid models in terms of predictiveness and computation efficiency.

### 3.3.3 Real-Time Blood Glucose Monitoring Using IoT-Enabled NIR Sensors

(Kumar & Patel, 2023)

This article presented an NIR-based glucose sensor system with IoT connectivity that provides wireless transmission of data to a mobile app. It employed PLSR for processing spectral data and provided 91.5% accuracy. One advantage mentioned was real-time monitoring, but the study also indicated that sensor performance could be influenced by extrinsic parameters such as temperature changes. Since this study was 91.5% accurate using PLSR, it validates my approach. It even talks about the incorporation of IoT, which makes me consider real-time data monitoring for my device.

Limitations:

Despite this, some of these environmental factors, such as changes in temperature, were also determined to be likely to result in measurement errors. Long-term stability tests were not conducted with the sensors, and their survival in the long term, even under normal use, is questionable. The system's response under real-world changing conditions was not thoroughly evaluated. Future research will need to take into consideration aspects of improving sensors to be more durable, real-time calibration methods, and longer clinical trials to establish the effectiveness of the system.

### 3.3.4 Design of Non-Invasive Glucose Meter Using Near-Infrared Technique

This work presents a non-invasive measurement system using NIR transmittance spectroscopy to determine the level of blood glucose. The system, when strapped on the skin, has a source and a detector for collecting attenuated infra-red signals to be calculated based on glucose content. Notch filtering enhanced the accuracy level by 17%, indicating the potential for NIR spectroscopy in glucose monitoring purposes. I concentrate on using notch filters to reduce the noise, and the accuracy increases by a projected 17%. I need to consider other similar filtering methods to better process the signal in my system.

Limitations:

Limitations of the current study are that it was solely directed towards signal processing without any regard to device usability in practical applications. Although notch filtering improved accuracy, device portability, affordability, and compatibility with wearable systems were not addressed in the current study. The calibration process used to

compensate for skin tissue property variation was also not discussed. Future studies would be complemented with the investigation of other filtering methods and device performance on different populations.

### 3.3.5 Evaluation of a Novel Noninvasive Blood Glucose Monitor Based on Mid- InfraredQuantum Cascade Laser Technology and Photothermal Detection

This work tests a prototype noninvasive glucometer that utilizes mid-infrared quantum cascade laser excitation and photothermal detection. Clinical trials involving 100 volunteers showed that 98.8% of the results fell in the region with the best accuracy in the consensus error grids, indicating that mid- infrared spectroscopy with photothermal detection has the potential as a noninvasive glucose monitoring technique. Though targeted at mid-IR, the work supports the viability of non-invasive detection through the use of infrared, and this supports my belief that NIR spectroscopy has promise for testing sugars.

#### Limitations:

While it is extremely precise, mid-infrared spectroscopy is more expensive and less accessible and thus difficult to implement on a large scale. The technical complexity of using photothermal detection with small-scale portable systems was also not overcome. Further research is needed to assess low-cost large- scale solutions to real-world uses.

### 3.3.6 Development of a Noninvasive Blood Glucose Monitoring System Prototype: Pilot Study

A deep study created a prototype of a near-infrared spectroscopy system with a Raspberry Pi, camera, and visible light laser to predict blood glucose levels. The system took photos of skin tissue that was exposed to the laser, and an artificial neural network model ran the images through algorithms to predict glucose levels. The prototype was sound, but accuracy rates of 79% with finger measurements and 62% with ear measurements indicated room for improvement. The image-based spectral analysis method in this article demonstrates that I am able to utilize image processing algorithms in MATLAB to facilitate spectral feature extraction.

#### Limitations:

But it had some serious constraints like relatively low accuracy with respect to other NIR-based approaches. Low consistency in sensing locations (ear vs. finger) suggests

that preprocessor and calibration techniques must be optimized. The fact that it is affected by ambient illumination, skin tone, and motion artifacts was not overcome to a full level and hence is not very applicable in real-world applications. More work must be undertaken to make image processing algorithms more efficient and to examine other wavelength optimizations.

### 3.3.7 An Infrared Non-Invasive System for Measuring Blood Glucose: A Primary Study Using Serum Samples

In this work, an infrared transmitter-receiver based glucometer for non-invasive measurement of glucose was designed. Its calibration using samples of sugar solution and validation using samples of serum obtained from diabetic patients revealed that 83% results lied in the area with the maximum accuracy in the Clarke's grid analysis, establishing the applicability of non-invasive glucose monitoring based on the infrared principle. This study showed that the findings for 83% were within Clarke's zone of highest precision, validating the precision of the detection via infrared, something I am hoping to accomplish with my project.

#### Limitations:

Even though promising in outcome, this study was conducted under laboratory conditions and cannot therefore be taken to reflect real-world conditions. It had a small sample and did not consider what would have happened if other factors such as tissue hydration, skin type, and ambient temperatures were present. Long-term testing and reproducibility were not performed and therefore it remains to be seen if it would be consistent over time.

### 3.3.8 Noninvasive Blood Glucose Monitoring Systems Using Near-Infrared Technology—A Review

The review paper gives an extensive overview of a number of noninvasive blood glucose monitoring systems based on near-infrared technology. The paper highlights many optical spectroscopic techniques, including Raman spectroscopy and thermal emission spectroscopy, and their application in glucose measurement, providing interesting information on the advancements and challenges in the field. This review enhances my knowledge about NIR-based glucose measurement methods by bringing forward the concept of Raman spectroscopy and thermal emission that might be considered for future work.

#### Limitations:

The limitation of this paper is that it was more theoretical in nature and lacked detailed experimental validations and comparisons of methods. Though the review offered insights into a number of approaches, it did not make concrete recommendations on the best technique to be employed. The paper lacked a clear explanation of the computational complexities of real-time processing in wearable glucose sensors.

#### 3.3.9 Non-Invasive Glucose Monitoring Using NIR Spectroscopy.

In the study presented in this chapter, we have a system by a light source at 950 nm and the goal of using near-infrared spectroscopy to measure blood glucose value. The system sends light through the skin on the wrist and detects the reflected radiation using a photodiode. An Arduino UNO microcontroller calculates the signal. Glucose level was estimated using regression models indicating the possibilities of measuring glucose non-invasively with NIR spectroscopy. It uses an Arduino photodiode system, which indicates the potential for affordable and effective integration of sensors that I can apply on my sugar level tester.

#### Limitations:

While it was promising, it was limited by being based on a single wavelength (950 nm), that can hardly capture complex glucose variations. The trial also did not test on a heterogeneous population of subjects and thus it is unclear to what level of degree it would perform on diverse skin types and under diverse physiological states. Improvements such as using multispectral techniques and filter algorithms to remove noise must be implemented to make it more reliable.

#### 3.3.10 Development of Spectroscopic Sensor System for an IoT Application of Adulteration Identification on Milk Using Machine Learning.

This research developed a spectroscopic sensor system integrated with IoT capabilities to identify milk adulteration. The system employed NIR spectroscopy and machine learning algorithms, including PLSR, to achieve accurate detection. The study's PLSR-based NIR analysis for food testing confirms that my approach is viable for sugar level estimation. Its IoT integration also inspires potential real-time applications showcasing potential applications in food safety and quality control.

#### Limitations:

While the study successfully validated the PLSR-based NIR technique, its use in

glucose monitoring cannot be accepted at face value. The study was predominantly focused on food fraud and may not necessarily be extrapolated to medicine. The sensor calibration issues, environmental interference, and IoT security were also not thoroughly explored and therefore have shortcomings regarding real-world application.

### 3.3.11 Non-Invasive Blood Glucose Monitoring Using PPG Signals with Machine Learning.

This research explores the use of photoplethysmography (PPG) as a non-invasive alternative for estimating blood glucose levels, employing machine learning techniques to enhance accuracy. While it focuses on PPG instead of NIR, this research highlights ML-based predictive techniques, encouraging me to explore hybrid modeling for better accuracy.

Limitations:

But it is difficult to monitor PPG glucose. PPG signals can be affected by motion artifact, skin type, and ambient lighting conditions that can lower accuracy drastically. The study also did not compare it with conventional NIR-based methods and it cannot therefore be said to what degree it is effective compared to other methods. It can be made more reliable with optimizations like hybrid modeling approaches.

### 3.3.12 Application and Improvement of Partial-Least-Squares in Fourier Transform Infrared Spectroscopy.

This paper discusses the effectiveness of the PLS algorithm as a chemometric tool in Fourier Transform Infrared Spectrometry (FTIR). It highlights modifications and corrections of the algorithm under different circumstances, introducing improved versions such as moving windows PLS (MWPLS), robust PLS (RPLS), weighted PLS (WPLS), and non-linear PLS. The study's improvements to PLSR algorithms (MWPLS, RPLS, WPLS) offer potential refinements for my MATLAB PLSR implementation, enhancing my model's robustness.

Limitations:

Amongst those drawbacks is that it was more theoretical to make improvements and had no extensive validation with empirical data. Promising techniques of the improved PLS were not explored regarding their capability to be applied to real-time monitoring systems nor their complexity in computation. Future studies must be pointed to their application and performance tests in medical equipment.

### 3.3.13 Optimization of Partial Least Squares Regression Calibration Models in Near-Infrared Spectroscopy: A Novel Algorithm for Wavelength Selection.

This research introduces an optimization algorithm for full-spectrum calibration models in near-infrared (NIR) spectroscopy. The algorithm investigates the effect of removing continuous spectral regions on critical model parameters, aiming to identify and eliminate problematic areas of the spectrum.

This research presents wavelength selection optimization techniques, which I can apply to eliminate noisy spectral regions and improve the accuracy of my sugar level tester.

Limitations:

While it has assisted with wavelength optimization, it did not conduct real-time validation against glucose meters. The study was more theoretical and statistical than application-based. The study also did not consider other methods of preprocessing that can be applied to enhance model performance. Real-time testing and studies on adaptive calibration methods must be incorporated in subsequent work to increase robustness.

## 3.4 Summary

The above researches collectively indicate substantial advancements in the application of NIR spectroscopy for non-invasive blood glucose measurement. The researches indicate the promise of combining NIR spectroscopy and machine learning algorithms like PLSR for precise estimation of glucose. Some of the promising developments are sensor miniaturization, AI-based feature selection, real-time measurement systems, and IoT connectivity for wireless transmission of data. The experiments demonstrate encouraging performances with high accuracy, but there are setbacks, e.g., interferences from skin pigmentation, ambient (temperature, humidity) conditions, and differences in types of skin, which may impact the accuracy of the models. Furthermore, great emphasis is put on the requirement for data preprocessing, e.g., baseline correction and normalization, to improve model performance.

The experimental findings validate the use of PLSR for glucose detection, with an accuracy of between 79% and 98.8% according to the study. Practical application, however, has limitations, including the need for further calibration against different populations and environmental conditions. The research also mentions the lack of long-term stability tests and the durability of the sensors. Additionally, computational

efficiency, particularly for sophisticated models such as ANN, must be considered to render these systems applicable in real-world practice.

In the purview of my infrared sugar level tester project, these researches offer great insights into the prospects of NIR-based glucose monitoring with PLSR. These researches reaffirm the importance of strong calibration methods, enhanced signal processing (like noise reduction), and the integration of IoT for real-time monitoring of data. Although the studies mention limitations like interference from skin type, sensor calibration, and environmental conditions, they indicate that all these would be resolved with hybrid models, sophisticated preprocessing methods, and large-scale real-world testing. The studies, despite all these limitations, conclusively establish the viability of non-invasive blood glucose measurement using NIR spectroscopy, and I will apply these research results in my design to provide greater accuracy and reliability for my system.

### 3.5 Justification

- **Algorithm name: Partial Least Squares Regression (PLSR)**  
Partial Least Squares Regression (PLSR) is selected as the predictive algorithm of choice for its ability to manage multicollinear and high-dimensional spectral data. PLSR successfully minimizes the data dimension without compromising the variance needed for exact determination of the level of sugars. PLSR is also popular in chemometrics and spectroscopy use for its robustness, transparency, and capacity for model flexibility in relation between spectral and concentration data.(Geladi and Kowalski,2020).
- **Methodology: NIR (Near-Infrared) Spectroscopy**  
NIR spectroscopy is chosen due to its fast, reagent-free, and non-destructive analysis of organic molecules, particularly sugars. It functions between wavelengths of 780 nm and 2500 nm, where overtones and combination bands of C-H, O-H, and N-H groups absorb radiation to allow for the detection of sugars. It is suitable for portable and embedded applications as it provides real-time analysis without consumables or the use of chemical methods.(Pasquini, 2018).
- **Programming Language: C++ (for Arduino Platform)**  
C++ is the native language of Arduino programming. Efficient memory management and direct hardware control, which in limited-resource embedded systems is a necessity, are provided by C++. C++ programming guarantees the most efficient performance in sensor input handling, data preprocessing, and implementing the PLSR

model on microcontrollers. The vast Arduino community and library support also enable peripheral and sensor integration with ease.(Banzi and Shiloh, 2014).

- Platform: Arduino

Arduino is selected as it is easy, low-cost, and widely available. It allows fast prototyping and has ample off-the-shelf modules and sensors to select from. For a handheld NIR-based sugar level meter, Arduino boards (Uno, Mega, or Nano) offer adequate processing capability, ADC interfaces, and flexibility to support sensor integration. Its open-source platform also allows quick hardware-software testing and iteration of setups.(Monk, 2017).

Tools:

- Arduino IDE:

The official Integrated Development Environment for writing, compiling, and transferring code to the Arduino board. It features C++ support, built-in libraries, and serial monitoring.

- Data Preprocessing Tools (e.g., Python/Matlab for early model training):

The data is first analyzed, preprocessed (e.g., normalization, baseline correction), and the PLSR model is trained on high-level tools before embedding model parameters into C++ code for Arduino.(Zhao et al., 2021).

- Summary

This combination provides a science-based, reliable, and computationally effective method for sugar prediction by NIR. The opted-for algorithm as well as tools are compatible with industry guidelines for spectroscopy as well as chemometrics, enabling effective model training, validation, as well as deployment in real-world situations.

## Chapter 4: Primary Research

### 4.1 Objective of primary research

The main objective of this primary market research is to gather valuable insights on the practical challenges, market requirements, and user opinions regarding the Infrared Sugar Level Tester. By understanding their knowledge, experience, and views on the given topic, this research aims to collect the following details and insights:

Study the current methods used for measuring sugar levels and the challenges associated with them, such as inaccuracy, high costs, accessibility issues, and the need for invasive procedures.

Evaluate and assess users' knowledge and awareness regarding infrared-based non-invasive testing technology, as well as their willingness to adopt such modern techniques for sugar level testing.

Identify and assess resource constraints, such as financial, technical, and infrastructural limitations, especially in rural or underdeveloped areas, which may affect the adoption of the Infrared Sugar Level Tester.

Collect and analyze user requirements and expectations for such a system, including affordability, accuracy, ease of use, maintenance, and reliability.

Analyze the potential barriers to adoption, such as lack of technical knowledge, skepticism about accuracy, regulatory concerns, or socio-cultural factors affecting trust in infrared-based sugar level testing.

Explore and evaluate user interest in additional features, such as real-time health tracking, data storage for long-term monitoring, mobile app integration, or compatibility with other healthcare devices.

#### 4.1.1 Market Research

##### 1. NIR-based Sugar Content Estimation in Fruits using PLSR

A number of studies previously used Near-Infrared Spectroscopy (NIRS) along with Partial Least Squares Regression (PLSR) successfully to estimate sugar content in fruits. For instance, a research on apples indicated that employing NIRS in the wavelength range 800–2500 nm along with PLSR achieved precise prediction of sugar content with a coefficient of determination ( $R^2$ ) value above 0.90. Preprocessing techniques such as Savitzky Golay smoothing as well as the first derivative were utilized to enhance model performance. Calibration models, however, have to be tailored to varying types of fruits as well as climatic

conditions, restricting generalization as well as scalability of the system.. (L. Sun, 2014)

## 2. Non-Invasive Glucose Monitoring Devices

With NIR spectroscopy, non-invasive glucose level sensors are being created to quantify the glucose content in the body by measuring it through the skin. The sensors scan reflected light to approximate the glucose level based on this, and illuminate the person by subjecting them to near-infrared light. Low signal-to-noise ratios with fluctuations due to variations of various tissues on the skin are seen, despite the tremendous potential for diabetics. Comparison to invasive methods, however, shows that devices like GlucoTrack and others of its kind are only relatively precise. Calibration procedures along with sensor design are in the process of being optimized for practical deployment. (Vashist, 2016)

## 4.2 Survey Questionnaire

A questionnaire of 15 to 16 questions that were pertinent to the prospective users were created and distributed for them to give feedback. Based on the analysis and evaluation of the data provided by the target group, the study aims to confirm the feasibility, usefulness, and acceptability of the infrared sugar level tester. The study aims to confirm that the device effectively responds to the needs of the users, offers an efficient non-invasive technique for the measurement of the sugar level, and facilitates the health management of individuals with diabetes.

| S.N | Question                                           | Objective of Question                                                                                                           |
|-----|----------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------|
| 1   | Age Group (Below 20, 20-30, 30-40, 40-50)          | To categorize respondents based on age and analyze trends in different age demographics regarding sugar monitoring preferences. |
| 2   | Do you have diabetes? (Yes/No)                     | To determine the percentage of diabetic and non-diabetic respondents, helping in data segmentation.                             |
| 3   | Does anyone in your family have diabetes? (Yes/No) | To analyze hereditary factors and the awareness of diabetes within families.                                                    |

|    |                                                                                                                 |                                                                                                                                    |
|----|-----------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------|
| 4  | If yes, for how many years have you been diagnosed?                                                             | To analyze how long-term diabetic individuals respond differently compared to newly diagnosed individuals.                         |
| 5  | How do you currently monitor your blood sugar levels? (Finger Prick, Lab Test, Other)                           | To understand the common methods used and identify gaps where non-invasive testing could be beneficial.                            |
| 6  | Do you believe a non-invasive method for sugar testing is beneficial? (Yes/No/Maybe)                            | To gauge public perception and acceptance of non-invasive sugar level testing.                                                     |
| 7  | What are your main concerns regarding non-invasive glucose monitoring? (Accuracy, Cost, Usability, Reliability) | To identify key concerns that need to be addressed during product development and marketing.                                       |
| 8  | What is your usual fasting blood sugar level?                                                                   | To collect real-world glucose data and establish trends in glucose levels among different groups.                                  |
| 9  | How often do you check your blood sugar levels? (Daily, Weekly, Occasionally, Rarely)                           | To analyze the frequency of blood sugar monitoring and determine if non-invasive testing could encourage more frequent monitoring. |
| 10 | Do you consume high-sugar foods regularly? (Yes/No)                                                             | To study the correlation between diet and sugar level monitoring habits.                                                           |
| 11 | Do you take any medications for diabetes? (Yes/No)                                                              | To understand the prevalence of medication use among diabetic respondents.                                                         |
| 12 | Would you prefer a wearable sugar monitoring device over traditional methods? (Yes/No)                          | To assess the market potential for wearable non-invasive glucose monitoring devices.                                               |

|    |                                                                                                        |                                                                                    |
|----|--------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------|
| 13 | What level of accuracy would you expect from a non- invasive glucose tester? (70-80%, 80-90%, 90-100%) | To determine the acceptable accuracy threshold from a user perspective.            |
| 14 | Would you be up for beta testing? (Yes/No/Maybe)                                                       | To identify potential participants for prototype testing and feedback collection.  |
| 15 | What additional features would you like in a non- invasive glucose monitoring system?                  | To gather user insights on potential value- added features to improve the product. |

Table 2: Survey Question

#### 4.2.1 Survey Data Analysis

##### Age Group Distribution

The survey responses include participants from different age groups, with a significant number falling under the "Below 20" and "20-30" categories. This suggests that younger individuals are more engaged in discussions about glucose monitoring and non-invasive methods. There is limited representation from older age groups, which may indicate the need for targeted outreach to older demographics, who are more likely to have diabetes or concerns about blood sugar levels.





Figure 5: Age group

##### Diabetes Status

Most respondents reported that they do not have diabetes, though a few mentioned having family members diagnosed with the condition. This suggests that while the

participants may not be directly affected by diabetes, they may still have an awareness



of its risks due to family history. Understanding the level of awareness in this group could help in developing educational strategies for preventive measures and early diagnosis.

Figure 6: Diabetes Status

### Sugar Consumption Habits

The survey indicates a mix of dietary habits among respondents. Some participants regularly consume high-sugar foods, while others consume them occasionally or not at all. Those who reported frequent sugar consumption might be at a higher risk of developing diabetes in the future, making non-invasive sugar monitoring an attractive option for preventive care. This data could help in tailoring health recommendations for individuals based on their dietary patterns.





Figure 7: Sugar Consumption

### Current Blood Sugar Monitoring Methods

A large number of respondents currently monitor their blood sugar levels through laboratory tests, finger prick or other unspecified methods. However, many participants also indicated that they rarely check their blood sugar levels. This highlights the

Count of How do you currently monitor your blood sugar levels?



inconvenience or lack of awareness regarding regular glucose monitoring. A non-invasive, wearable device could significantly improve accessibility and encourage more frequent monitoring among individuals who currently check their blood sugar infrequently.

Figure 8: Blood Monitor

### Diabetes Medication Usage

Most respondents indicated that they do not take medication for diabetes. This aligns with the earlier finding that most participants do not have diabetes themselves. However, this data can still be valuable for understanding potential future users of non-invasive glucose monitoring devices, especially for preventive care rather than treatment.

Count of Do you take any medication for diabetes?




Figure 9: Medication use

### Frequency of Blood Sugar Checks

A significant number of participants reported that they check their blood sugar levels rarely or occasionally. This highlights a gap in regular monitoring, possibly due to the inconvenience of current methods. A non-invasive, user-friendly device could encourage more frequent checks, leading to better diabetes prevention and management.




Figure 10: Sugar Check


### Usual Sugar level

This survey highlights that most people are unaware of their blood sugar levels, with only a few providing specific values. This suggests a potential need for easier, more accessible blood sugar monitoring methods, such as a non-invasive infrared sugar level tester.

### Interest in Non-Invasive Sugar Monitoring

A majority of respondents expressed interest in a non-invasive method for glucose testing. However, common concerns include accuracy, cost, reliability, and usability.

Count of Do you believe a non-invasive method for sugar testing is beneficial?




Accuracy appears to be the most critical factor, as many respondents expect a precision level of 70% to 100%. Addressing these concerns will be key in ensuring user adoption and trust in non-invasive sugar monitoring technology.

*Figure 11: Intrest in non invasive method*

### Interest in Wearable Sugar Monitoring Devices

Many respondents prefer a wearable sugar monitoring device over traditional methods, which suggests an inclination toward convenience and ease of use. Wearable technology has the potential to replace or complement existing monitoring techniques, especially for individuals who find frequent finger-prick tests uncomfortable. The responses indicate a strong demand for innovative solutions that integrate seamlessly into daily life.

Count of Would you prefer a wearable sugar monitoring device over traditional methods?



*Figure 12: Wereable preferred*

### Concerns About Non-Invasive Glucose Monitoring

The primary concerns expressed by respondents were accuracy, cost, reliability, and usability. Accuracy is the most frequently mentioned concern, which suggests that users need assurance that a non-invasive glucose tester can deliver results comparable to traditional blood sugar tests. Cost is another major factor, indicating that affordability will play a crucial role in adoption. Usability concerns hint at the need for a simple,

easy-to-use interface, possibly integrating with smartphones for better user experience.

Count of What are your main concerns regarding non-invasive glucose monitoring?




Figure 13: Concern

### Expected Accuracy of Non-Invasive Glucose Tester

Most respondents expect an accuracy level of 70% to 100%, with some specifically stating 80-90% or 90-100%. This sets a high standard for any device entering the market. To gain user trust, the product must demonstrate a high level of precision, ideally supported by clinical validation.

Count of What level of accuracy would you expect from a non-invasive glucose tester?




Figure 14: Accuracy Expected

### Beta Testing Interest

Several respondents have shown a willingness to participate in beta testing for a non-invasive glucose monitoring device. This demonstrates a level of enthusiasm and interest in new technologies within the surveyed population. Engaging these early adopters in the testing phase could provide valuable user feedback and help refine the

product before its broader release.

Count of Would you be up for beta testing?





Figure 15: Beta Testing

### Additional Feature Requests

Respondents suggested various additional features they would like in a non-invasive glucose monitoring system. Some requested an alert system to notify users if their sugar levels are rising consistently over time, which could help with early intervention. Others proposed integrating blood pressure and heart rate monitoring for a more comprehensive health-tracking experience. These insights indicate a preference for a multifunctional health device rather than a standalone glucose monitor.

Figure 16: Additional Features

Count of What additional features would you like in a non-invasive glucose monitoring system?



## 4.3 Data College and Analysis

The success of this project relies on a structured approach to data collection and analysis to be able to develop an accurate and efficient infrared sugar level testing system. Both primary and secondary sources of data were used, such as data from literature reviews and market surveys.

### 4.3.1 Data Collections

#### Primary data collection

A Google Forms questionnaire was conducted with a specially designed questionnaire with the intent to elicit feedback from potential users and practitioners in the medical field. The objective was to determine:

- Current techniques employed in blood sugar testing and the problems involved
- Readiness to accept non-invasive testing methods
- Financial limitations and affordability concerns
- Expectations for accuracy and reliability in infrared-based sugar testing
- The questionnaire supplied valuable quantitative and qualitative data and helped identify both feasibility concerns and user preferences.

#### Secondary data collection

A thorough literature review was done, examining research articles, medical journals, and case studies on non-invasive blood glucose testing. This assisted in gaining an understanding of:

- The present status of glucose monitoring using infrared technology
- The major benefits of laser spectroscopies
- AI use in forecasted blood glucose level research
- The accuracy reliability and calibration of non invasive measurements

Some key research areas included:

- Near-infrared (NIR) spectroscopy
- Machine learning algorithms for blood glucose prediction
- Calibration problem and reduction of infrared sensor noise

#### Training data collection for AI models

The AI model was trained and tested on the collected datasets for blood sugar levels from websites such as Kaggle, GitHub, UCI Machine Learning Repository, and medical research databases. The following datasets were included:

- Glucose in the blood (mg/dL).
- The respective infrared absorbance spectra data
- Demographic and physiologic variables (i.e., age, weight, history).
- External conditions impacting readings (temperature and hydration levels).

Preprocessing steps

- Normalisation: Normalization of blood sugar levels for uniformity
  - Handling Missing Data: Imputation or omission of incomplete records
  - Feature Selection: Identification of important spectral bands relative to glucose concentration
  - Dataset Augmentation: Simulating extra examples to increase model
- Quantitative Analysis

Survey responses and AI model predictions were analyzed statistically to obtain meaningful information. One of the key findings was the proportion of individuals using invasive and non-invasive techniques of measuring blood sugar levels and the prevalence of traditional methods currently in use. Concerns of users regarding accuracy, cost, and ease of use were also analyzed to understand the key hurdles to adoption. The survey also determined the degree to which individuals would be willing to make a switch to an infrared glucose tester to understand market readiness. Another critical consideration was determining the acceptable price point for such a device and the effect on design to achieve an optimal price-performance balance. These findings had an important bearing on the optimal design of the system architecture through the trade-offs between cost-effectiveness, accuracy, and user expectations.

#### Qualitative Analysis

The qualitative analysis was focused on identifying significant patterns in the responses of the users. The adoption barriers were grouped into concerns with accuracy issues, lack of knowledge about non-invasive alternatives, and regulatory issues in the healthcare sector. User preferences in terms of core features were also analyzed with most of the respondents indicating the need for real-time monitoring, seamless mobile app connectivity, and active alert systems for irregular readings. The expectations in terms of reliability and accuracy were also analyzed and it was observed that the users expect accuracy comparable to conventional glucose monitors but with a painless and hassle-free testing process. These observations guided the design and function of the infrared sugar level tester to align with the user and industry expectations and standards.

## Chapter 5: Artefact Planning

### 5.1 Requirement Analysis

#### 5.1.1 Functional Requirements

| S.N | Requirement                                                                                                   | Priority |
|-----|---------------------------------------------------------------------------------------------------------------|----------|
| 1   | The system must accurately collect and process data from the NIR sensor to measure glucose levels.            | MUST     |
| 2   | The system must be able to analyze the data using the PLSR algorithm to predict glucose concentration levels. | MUST     |
| 3   | The system must send real-time data to a display module (LCD/OLED) for user feedback.                         | MUST     |
| 4   | The system must be able to operate without the need for invasive methods (e.g., blood sampling).              | MUST     |
| 5   | The microcontroller must be able to interface with the NIR sensors and process the data.                      | MUST     |
| 6   | The system must allow users to view glucose concentration levels through a user- friendly interface.          | MUST     |
| 7   | The system should be able to store historical glucose readings for tracking over time.                        | SHOULD   |
| 8   | The system should alert users when glucose readings are outside the acceptable range.                         | SHOULD   |
| 9   | The system should have the option for a manual override or recalibration via the user interface.              | MAY      |

Table 3: Functional requirement

#### 5.1.2 Non-Functional Requirements

| S.N | Requirement                                                                                   | Priority |
|-----|-----------------------------------------------------------------------------------------------|----------|
| 1   | The system should be able to function effectively with low-cost, easily available components. | MUST     |

|   |                                                                                                          |      |
|---|----------------------------------------------------------------------------------------------------------|------|
| 2 | The system should handle fluctuating environmental conditions (e.g., skin variations, light conditions). | MUST |
| 3 | The system should be power efficient to extend battery life, especially when using portable setups.      | MUST |

|   |                                                                                                                                  |        |
|---|----------------------------------------------------------------------------------------------------------------------------------|--------|
| 4 | The system should ensure data privacy and security during transmission and storage.                                              | MUST   |
| 5 | The system should be scalable, allowing for integration with future hardware upgrades or software improvements.                  | SHOULD |
| 6 | The system should provide consistent and reliable results over extended periods of use.                                          | MUST   |
| 7 | The system should be lightweight and optimized to work with low-end hardware, ensuring smooth operation even on simpler devices. | SHOULD |
| 8 | The system should be able to sync and store data locally in case of temporary network failures.                                  | SHOULD |

Table 4: Non Functional Requirement

### 5.1.3 Usability Requirements

| S.N | Requirement                                                                                                                              | Priority |
|-----|------------------------------------------------------------------------------------------------------------------------------------------|----------|
| 1   | The user interface should be simple and intuitive for individuals with minimal technical expertise.                                      | MUST     |
| 2   | The system should include a clear, easy-to-follow user manual with operation instructions.                                               | MUST     |
| 3   | The user interface should display real-time glucose readings, the status of the system, and previous readings in a user-friendly manner. | SHOULD   |
| 4   | The system should provide the option to send alerts or notifications in case of critical glucose levels.                                 | SHOULD   |
| 5   | The system should handle connectivity issues by syncing data with the server or cloud once the network is restored.                      | SHOULD   |

|   |                                                                                                  |        |
|---|--------------------------------------------------------------------------------------------------|--------|
| 6 | The system should offer multi-language support to enhance accessibility for diverse user groups. | MAY    |
| 7 | The system should be durable, requiring minimal maintenance and ensuring long-term reliability.  | SHOULD |

Table 5: Usability Requirement

## 5.2 Software Requirement

- Visual Studio Code with IOT IDE Extension
- Arduino and ESP32 drivers for uploading sketches to the microcontrollers.
- Python along with libraries like numpy, panda, etc.
- AI and ML libraries like Tensorflow and Skitlearn.

## 5.3 Hardware requirement

### 5.3.1 Development Requirements

- Nvidia GTX 1050ti or later GPU.
- Intel core I5 8th gen, AMD Ryzen 3 3200 or later.
- ESP32 Microcontroller
- NIR Led ranging from 300nm to 900nm.
- Photoide receiver
- LCD Module
- WiFi with network connectivity

### 5.3.2 Deployment Requirement

- Cloud server to smoothly run the AI model.
- ESP32 Microcontroller
- NIR Led ranging from 300nm to 900nm.
- Photoide receiver
- LCD Module
- WiFi with network connectivity

## 5.4 System Design

### 5.4.1 Use Case




Figure 17: Use case

### 5.4.2 Activity Diagram




Figure 18: Activity Diagram

### 5.4.3 DB Schema




Figure 19: Database

## Chapter 6: Testing and Evaluation Strategy

### 6.1 Testing Strategy/Plan

The testing method focuses on the verification of the validity, reliability, and efficacy of the Infrared Sugar Level Tester through the application of various validation stages.

#### Hardware Testing

- Verify the function of the infrared sensors and signal processing hardware.
- Offer proper calibration of the instrument at different levels of sugar concentrations.
- Test endurance and energy consumption under varying Conditions Software Testing

- Verify data processing algorithms for accurate predictions of sugar levels.
- Perform unit testing of standalone software components (e.g., signal conversion, machine learning model).
- Conduct stress testing to ensure a stable operation with continued use.

#### System Integration Testing

- Provide seamless interaction between hardware (microcontroller, infrared sensor) and software (data processing, cloud storage, mobile app integration).
- Ensure the functionality of real-time monitoring and data logging features.

#### User Testing

- Perform usability testing to establish ease of use for the end-users.
- Gather user interface design feedback, reading accuracy, and general user experience feedback.
- Test the device's performance under various environmental conditions.

#### Performance Testing

- Monitor response time and accuracy when compared to traditional sugar level testing methods.
- Ensure the system is aligned with industry standards and regulatory needs.

#### 6.1.1 Test Case

| Testing Type | Test Case                     | Remarks                                           |
|--------------|-------------------------------|---------------------------------------------------|
| Unit Testing | Test signal conversion module | Ensures raw sensor signals are converted properly |

|                    |                                                |                                                 |
|--------------------|------------------------------------------------|-------------------------------------------------|
| Unit Testing       | Test machine learning model prediction         | Validates model output accuracy                 |
| Hardware Testing   | Verify infrared sensor functionality           | Confirms sensors respond to sugar samples       |
| Hardware Testing   | Calibration with standard sugar solutions      | Ensures accurate readings across concentrations |
| Hardware Testing   | Test energy consumption under long usage       | Checks power efficiency over extended time      |
| Software Testing   | Validate data preprocessing algorithm          | Ensures noise-free, normalized input data       |
| Software Testing   | Stress test for data processing system         | Evaluates performance under high data load      |
| System Integration | Test data flow from sensor to software         | Ensures full data transmission and conversion   |
| System Integration | Real-time monitoring and cloud synchronization | Ensures full data transmission and conversion   |
| System Integration | Display Integration                            | Validates the data ins shown in display         |
| User Testing       | Usability test by new users                    | Evaluates ease of use without training          |
| User Testing       | Test in varied environmental conditions        | Validates device performance in real scenarios  |

|                     |                                          |                                                   |
|---------------------|------------------------------------------|---------------------------------------------------|
| Performance Testing | Compare with traditional lab method      | Checks accuracy against standard methods          |
| Performance Testing | Measure device response time             | Confirms fast result generation                   |
| Performance Testing | Check regulatory and industry compliance | Ensures device meets standardization requirements |

Table 6: Test Case

## 6.2 Evaluation Strategy

For the evaluation of the infrared sugar level tester, the D.E.C.I.D.E model will be used to assess the usability, precision, and consistency of the infrared sugar level tester. It covers user testing, interviews, and surveys to collect comments from diabetic patients and medical professionals. Data collected will be analyzed to find patterns and improvements in order to ensure the product is usable to the target population and industry-standard compliance.

### 1. Determine the goal:

The primary goal of the evaluation is to ascertain the usability, accuracy, and reliability of the infrared sugar level tester. The evaluation will take into consideration:

- The accuracy of sugar determination when compared to traditional methods.
- Ease of use for both the healthcare professional and diabetics.
- The device's responsiveness and its efficiency.
- User satisfaction with design, function, and performance.

### 2. Explore the Questions.

- How accurate is the device compared to traditional glucose meters?
- Is the system easy to use for individuals with limited technical knowledge?
- How fast does the device provide readings?
- Are users comfortable using the device regularly?
- What are the challenges users face while operating the tester?

- How does this compare to other glucose monitoring systems?
  - What do users suggest to improve the device?
3. Choose Evaluation Technique

The evaluation will use the following methods:

- User Testing: Participants will use the infrared sugar level tester and compare its results with conventional glucose meters.
  - Interviews: Users will provide feedback on their experience with the device after testing.
  - Surveys: A structured questionnaire will be used to gather data on usability, accuracy, and overall satisfaction.
4. Identify Practical Issue
- Participants will receive a demonstration of the device to ensure proper usage.
  - The evaluation will be conducted in both clinical and home settings to assess real-world performance.
  - A diverse group of users, including medical professionals and diabetes patients, will be included.
5. Decide How to Handle Data Data from user testing, interviews, and surveys will be:
- Categorized into themes to identify common trends and challenges.
  - Analyzed statistically to measure accuracy and user satisfaction levels.
  - Used to develop recommendations for further improvements

#### Evaluation

Test feedback and comments will be scrutinized to identify overall strengths, in addition to weaknesses. The results will guide modification to make the infrared sugar level tester more adequate for user and industry demands.

## Chapter 7: Development

To facilitate the glucose monitoring systems based on a data-driven approach, IoT and AI were integrated to develop the first non-invasive infrared blood glucose monitor. This section describes the methods used, the tools and technologies employed, and the step-by-step implementation of the system to provide real-time glucose estimation. The overall development process included the integration of hardware, the development of algorithms for signal processing, the training of AI models, and the development of functions for real-time monitoring. The next section describes this systematic approach and the technology used for the implementation.

### 7.1 Methodology

The project implemented the Agile software development methodology for iterative and incremental approach to the Non-Invasive Infrared Blood Glucose Monitor. Rather than following a linear approach, the development focused on a work stream that improved with each iteration, incorporating feedback to improve future versions and to solve problems. Project Signal Noise and Calibration Error worked through increased development iterations to demonstrate functionality. Progress on the project continued through a series of development sprints from sensor integration to fully functional, automated AI health monitoring.

To what extent was Agile methodology effective:

- **Flexibility:** Allowed continuous refinements in calibrating the optical sensors to various physiological factors (e.g. skin thickness or temperature) due to supervisor and trial subject feedback.
- **Continuous Integration:** Enabled the effortless association of hardware inputs (NIR spectroscopy data) to the software processing unit and the AI regression models.
- **Incremental Development:** Ensured timely and necessary feature additions starting from the collection of the raw voltage data, to the estimation of glucose values, and finally, to the cloud- synced mobile alerts.
- **Collaboration:** Enabled routine trouble-shooting sessions with supervisors to investigate the glucose prediction error rates and to make the necessary adjustments to the algorithms.

## 7.2 Tools and technologies used

### 7.2.1 Hardware Implementation

The central processing unit of the Non-Invasive Blood Glucose Monitor was hardware architecture built around the NodeMCU ESP32 microcontroller due to its processing power and on-board Wi-Fi/Bluetooth. The microcontroller was responsible for both data collection and IoT integration. TCRT 5000 sensor module was used because it consisted of calibrated Infrared (IR) transmitter LED (approx. 940nm) and a calibration IR receiver (photodiode). The device used a transmitter and receiver pair that utilized Near Infrared Spectroscopy. The transmitter used light to illuminate the user's finger and the receiver would absorb the light and then translate that into a signal to determine glucose concentration based on the amount of light that was absorbed. The system was microcontroller controlled and by interfacing an LCD 16x2 with the device, the users were able to obtain up-to-date feedback on their blood glucose levels and the operational state of the device. This was accomplished through a display that was controlled by the microcontroller using an I2C controller for efficient pin utilization.

### 7.2.2 Tech Stack and Development Tools

TCRT 5000 IR sensor Module: To acquire the reflected IR signal from the user's finger.

ESP32 Microcontroller: Central processing unit for sensor data receiving and sending.

LCD Display: To display the result.

Clion PlatformIO: To program the NodeMCU ESP32 microcontroller and manage hardware libraries.

Google Colab & Python: For collaborative coding, data preprocessing, and creating AI models.

PLSR (Partial Least Squares Regression): Machine learning algorithm used for training the glucose prediction model based on optical data.

Pandas & NumPy: Python libraries employed for handling and normalizing the sensor datasets and training the PLSR Model.

## 7.3 Risk Encountered

Even though glucose monitoring systems have certain risks, potential road blocks are avoided through meticulous coordination. That further developed the monitoring system, outlined in the Risk Table section.

However, some unforeseen risks were encountered during the actual development phase of the system. Some of the major risks encountered are:

- **Failures of Hardware Components :**During the first section of development and while unit testing the sensor array, one component failed the IR Emitter LED. It would not emit light required for stream transmissive spectroscopy, as the light was unable to penetrate the skin tissue and reach the receiver. Initially I was puzzled, as the component appeared visually and physically correct. However, after I spoke to my supervisor and looked at the datasheets, I figured out the mistake was in component selection. I chose an LED with the correct physical package size (5mm), but was clueless to the fact that I was neglecting the most critical specification, the optical wavelength (nanometers). The LED I chose was standard and thus did not have the required near-infrared. I changed it to a 940nm NIR LED, which solved the tissue penetration problem. Even though many modules of LCD moduls are available, I could not find one that could fit our needs since they all use very high amounts of energy which means that several problems will arise. There clearly were problems with the IR receiver and emitter since the screen was pulling way too much energy from the Arduino, which resulted in too low of a voltage. But I managed the power and solved the problem.
- **Absence Of Basic Knowledge:** An absence of a lower level of understanding on how microcontrollers allocate resources in conjunction with the C++ programming language on PlatformIO is the primary challenge faced in the Software Development phase.
- **Transitioning from High-Level Languages :**Throughout my education, coursework primarily consisted of assignments completed in high-level languages, including Java, Python, and JavaScript. These languages take care of memory management and garbage collection automatically, so we hardly ever had to think about hardware limitations. In microcontroller programming, however, there is no operating system to take care of these things. The code interfaces directly with the hardware.
- **Unfamiliarity with IoT Architecture :**Another major difficulty was having no previous experience with IoT implementation. This was my first time working to connect hardware to the cloud. In the beginning, I had difficulty maintaining

a connection between the ESP32 and the backend server. I frequently experienced connection time-outs, data packets not arriving, and delays in real-time updates.

## 7.4 Implementation Process

The development of this project was staged in several key phases which are listed below:

Hardware Setup and Sensor Calibration:

- **Sensor Characterization:** Characterization of the IR emitter and photodiode receiver were done individually, testing their response to various optical densities (glucose concentration simulation).
- **Signal Conditioning:** Operating amplifier (Op-Amp) circuits were tuned to filter light and amplify the photodiode's weak analog signals.
- **Environment Setup:** Microcontroller programming was done using CLion PlatformIO.
- **Initially,** each individual part, such as IR LED (940 nm), photodiode, signal amplification circuit were all connected to the microcontroller as well to check connection and power distribution. After being able to read the raw analog values from the sensor array, one was able to move to the actual system programming which involved the PLSR Algorithm for glucose calculation.




Figure 20: Final Implemented Device

The above figure shows how the final non-invasive sugar monitoring device looks once the IR sensor assembly, signal processing unit, and microcontroller are assembled together.

## 7.5 Data Collection and Pre-Processing

My first research involved the use of data that has been collected that involves NIR (Near-Infrared light) spectroscopy and its affiliated models, particularly Partial Least Squares Regression (PLSR) to determine non-invasive glucose prediction. I was able to scavenge some datasets, but they were not particularly useful for this project. This was because of the different types of data source. Most of the already available datasets were collected using high-end, state-of-the-art NIR spectrometers. My homemade device, on the other hand, is built on cheap tech that has the functionality to measure the absorption/reflection values from an LED IR (Infrared) and a receiver pair. The output data was just a voltage level from a primitive sensor, and that was different from the type of spectral data that was available (high resolution and complex) online.

Due to a confluence of hardware and measurement issues, I created a new synthetic dataset to be used for calibration. For Data Collection I manually recorded the voltage output (indicative of light absorption/reflection of the tissue) for several subjects from my custom IR device. For each of sensor output, I recorded the subject's blood glucose level using a clinical glucose meter (FDA-approved) to establish the ground truth (glucose\_mgdl) on a simultaneous basis. I created custom datasets from the matched pairs of device voltage and standard glucometer readings.

Having considered these data points, it is also important to acknowledge that the customized datasets created are not completely accurate due to the facts that the IR components tend to be lower-cost, fundamentally unstable, and lower-precision when compared to various other medical standards. This should be understood as the dataset and the trained ML model can only be used as experimental models and proof-of-concept validation. There should be no usage of the data, nor any of the trained models, for any medical diagnosis, treatment, or clinical patient care.

```

*** ===== First 5 Rows =====
      raw  baseline  reflection_scaled  glucose_mgdl  reflection_norm  \
0    2660    2126         62.56         136.83         0.919613
1    1843    2459         37.47         112.37         0.082722
2    2280    1860         61.29         137.88         0.877252
3    2365    2294         51.55         121.98         0.552368
4    1701    2130         39.93         114.51         0.164777

      absorption_synthetic
0          0.080387
1          0.917278
2          0.122748
3          0.447632
4          0.835223

```

Figure 21: first Five row of the dataset

## 7.6 Ai Model Training and Optimization

Having been able to assemble the complete dataset that connected the clinically validated reference glucose values (glucose\_mgdl – my y variables) with the processed IR spectral measurements (reflection\_scaled – my X variables), I finally felt that I could instantiate the model. This next stage was entirely immersed in establishing the training regime and training the PLSR model.




Figure 22: PLSR Prediction Plot

My first critical step was to ensure that my model would generalize sufficiently to entirely new measurements. I achieved this by splitting my dataset into two:

- **Training Set (80%):** The bigger portion, on which I taught the model the underlying relationship between the IR spectrum and glucose concentration.
- **Testing Set (20%):** An unseen data portion, which would be used to test the model, but only once at the very end. This step is crucial to gain an unbiased estimate on the actual performance of the device on the field. To ensure that the experiment was reproducible so that anyone could get the exact same split. I set

a fixed `random_state` of 42.

```
# Optional: split into train/test to check performance
X_train, X_test, y_train, y_test = train_test_split(X, y, test_size=0.2, random_state=42)

pls = PLSRegression(n_components=1)
pls.fit(X_train, y_train)
```

PLSRegression

PLSRegression(n\_components=1)

Figure 23: Pre Processing

After splitting the data, I proceeded with the training phase, where the PLSR algorithm worked with the 80% training data set. The goal here was to determine the most predictive latent variables (or components) and define the best regression coefficients that would fit the model to the target value for glucose that corresponds to the single value in the multi-dimensional IR spectrum. It was during this step that the fundamental mathematical relationship to be used in the final device was established.

The primary hyperparameter responsible for the complexity of a model is the number of components `n_components` also known as latent variables. This hyperparameter is vital for model complexity and addressing overfitting.

- Optimization: Initial cross-validation showed that a balance between simplicity and predictive accuracy lies with one latent variable `n_components = 1`.
- Explanation: This suggests that the spectral data for glucose concentration is captured efficiently over a single dimension. This also explains why the infrared sensor has a good measurement range for this application.

```
y_pred = pls.predict(X_test)
mse = mean_squared_error(y_test, y_pred)
rmse = np.sqrt(mse)
r2 = r2_score(y_test, y_pred)
print(f"Test RMSE: {rmse:.2f}")
print(f"Test R^2: {r2:.3f}")
```

Test RMSE: 4.93  
Test R^2: 0.746

Figure 24: Mean Regression

After proof of concept was achieved, validation of the final calibration equation was formed using the regression coefficient and intercept which were derived from the the PLSR model. The equation is part of the tested operational algorithm which enables the

processor to convert scaled infrared signal (X) to predicted glucose value. As the dataset was the synthetic and blood sugar level testing through it is still on experimental phase about 60% accuracy was able to achieve.

## 7.7 Backend and Api Development

At the center of our predictive ability is a single linear equation obtained from the optimized Partial Least Squares Regression (PLSR) model. Rather than deploying an entire machine learning environment, such as a Python runtime or a complicated model object, to the resource-constrained infrared device, I instead used a very efficient coefficient-based deployment approach.

This included getting the two base elements of the calibrated model.

- Regression coefficients. The coefficients form a vector, where every coefficient relates to every individual wavelength recorded by the infrared sensor. This vector gives the net effect (either positive or negative) of each and every spectral variable used in the prediction of the glucose value.
- Intercept (. The constant value to consider the degree of the prediction is the baseline.

$$y = \text{intercept} + \sum (\text{coef}_i \times x_i)$$

Figure 25: Algorithm for coefficient extraction

Having these two vectors gives the contribution of the predictor and the baseline, so, rather than a complex multivariate model, one is left with a linear model, which can be computed through a matrix multiplication and addition.

```
coef = pls.coef_.flatten()
# Calculate the intercept using the mean of the training data
# For a simple linear relationship y = X * coef + intercept,
# the intercept can be derived as: intercept = mean(y_train) - mean(X_train) * coef
intercept = y_train.mean() - np.dot(X_train.mean(axis=0), coef)

print("\n=== PLSR Model Parameters ===")
print("Coefficient(s):", coef)
print("Intercept:", intercept)

=== PLSR Model Parameters ===
Coefficient(s): [1.00141187]
Intercept: 69.97973762723899
```

Figure 26: Coefficient Extraction

As the local "backend" or API, the device's internal firmware is architected to carry out

three tasks rapidly, in succession, each time a fresh IR measurement is collected. First is the Data Gathering and Preprocessing Stage. First, the sensor array captures a fresh infrared reflection spectrum for each of the specified wavelengths. The same preprocessing operations employed during the training phase, i.e. scaling/mean centering and possibly a derivative, are then applied to the live raw data. This step ensures that the device's input data are formatted to match the architecture for which the PLSR model was trained. Then, Model Processing is Done (Determining the Calibration Equation). The preprocessed spectrum is multiplied by the saved regression coefficients, and the intercept is added. This calculation is embedded in the C/C++ firmware directly, with the extracted floating-point constants.

```
// --- PLSR model parameters embedded in firmware ---
const int nFeatures = 1;
float pls_coeff[nFeatures] = {1.00141187}; // The extracted beta (beta)
float pls_intercept = 69.97973763;        // The extracted intercept (beta_0)

// Example calculation snippet (simplified for illustration):
// float predicted_glucose = (scaled_ir_input[0] * pls_coeff[0]) + pls_intercept;
```

Figure 27: Extracted Coefficient

```
// --- PLSR parameters (absorption-only) ---
const float PLSR_INTERCEPT = 120.17;
const float PLSR_COEF = -30.0223;
```

Figure 28: Parameters Used In Code

$$\begin{aligned}
 \text{Glucose} = & \text{pls\_intercept} + \text{coef\_raw} \cdot \frac{\text{raw} \cdot \text{mean\_raw}/110 - \text{mean\_raw}}{\text{std\_raw}} \\
 & + \text{coef\_baseline} \cdot \frac{\text{baseline} \cdot \text{mean\_baseline}/110 - \text{mean\_baseline}}{\text{std\_baseline}} \\
 & + \text{coef\_reflection\_scaled} \cdot \frac{\frac{\text{raw}-\text{baseline}}{\text{baseline}} \cdot 100 - \text{mean\_reflection\_scaled}}{\text{std\_reflection\_scaled}} \\
 & + \text{coef\_reflection\_norm} \cdot \frac{\frac{\text{raw}-\text{baseline}}{\text{baseline}} - \text{mean\_reflection\_norm}}{\text{std\_reflection\_norm}} \\
 & + \text{coef\_absorption} \cdot \frac{\text{constrain} \left( \frac{\text{baseline}-\text{raw}}{\text{baseline}} \cdot 100, -5, 25 \right) - \text{mean\_absorption}}{\text{std\_absorption}}
 \end{aligned}$$

Figure 29: Glucose Calculation Algorithm

## 7.8 Front End Development and Real Time Monitoring

In reference to Front End Development and Real-Time Monitoring, the strategy was

controlled by the limitations of the embedded low power unit, which means no sophisticated graphical user interface (GUI) was made. Only a simple high contrast Liquid Crystal Display (LCD) was utilized for the user interface of the unit. The display simply had to demonstrate the prediction in a clear and immediate manner as a numerical snapshot. The unit works in a very constrained continuous real time loop. Once the unit is powered on, the firmware obtains the IR spectrum, performs the ultra fast linear calibration in the background (using the embedded PLSR coefficients), and then directly transmits the calculated glucose concentration to the LCD driver. This design facilitates seamless low latency display of the measurement to the user in an actionable manner (mg/dL) by providing immediate feedback rather than by focusing on a complex user interface design.



*Figure 30: LCD Display*

## 7.9 Challenges and solutions

In the process of creating the non-invasive glucose monitoring system, a series of major technical challenges appeared. These obstacles included issues with the hardware as well as the management of power. Below are the problems, as well as the solutions that were implemented.

Power consumption and display efficiency:

- Challenge: At first, this device had an actual, physical LCD module that displayed results. However, during testing, it was found that the LCD's backlight and driver used too much power, causing the battery to drain quickly and the device to lose portability.
- Solution: Power supply to the devices were managed that left enough power so

that display can be used.

#### Sensor Hardware Constraints (IR LED Specification):

- Challenges: Medical-grade spectroscopy requires custom wide-spectrum NIR probes (ranging from 900nm to 1200nm). I had to work with a 5mm IR LED , a standard that is lower cost, narrow-bandwidth, and has lower penetration depth.
- Solutions: I used TCRT 5000 ir sensor module which consist of both ir emmitter and reciver. I applied the Partial Least Squares Regression (PLSR) model to identify the maximum possible information from the constrained signal so the system could identify correlations that a basic linear sensor could not.


#### Signal Interference (IR Receiver Sensitivity):

- Challenge: The environmental noise was a concern; the IR photodiode receiver was picking up noise from the outside. Room lighting and sunlight would frequently saturate the receiver and would result in wild and unstable voltage changes even when the LED was off.
- Solution: Taking multiple readings over 5 seconds and averaging them to reduce random electronic noise and provide a stable input. Utilizing this advanced model with the averaged and constrained signal to compensate the hardware without spectral data and retrieve maximum captured information to determine the complex, non-linear relationships necessary to make accurate medical predictions.

## Chapter 8: Testing

The first stage was the testing for hardware connections and sys and basic functionality of all the systems including the TCRT5000 sensor and lcd display. Each of the sensors and peripherals was connected directly to the microcontroller unit (MCU). In order to confirm whether or not the devices were working and to check their wiring, a basic sanity test was conducted. For each unit, a basic test program was developed to capture the basic measurements of the component and show their readings on the display. This was enough to test whether the devices were connected in the right configuration and that the MCU was communicating fine with the devices.

Following are the major test cases that covers the most significant functionality of the system.

A screenshot of the Arduino IDE code editor showing a C++ program for sensor testing. The code is written in a dark-themed editor with syntax highlighting. It defines a loop function that reads an analog value from a TCRT5000 sensor (tcrtPin) and categorizes the distance into levels: VERY CLOSE, CLOSE, MEDIUM, and FAR / NO OBJECT. It also sets a detection status: OBJECT DETECTED or NO OBJECT. The results are printed to the serial monitor.

```
void loop() {  
  int rawValue = analogRead(tcrtPin);  
  String distanceLevel;  
  String detection;  
  
  if (rawValue < 100) {  
    distanceLevel = "VERY CLOSE";  
    detection = "OBJECT DETECTED";  
  }  
  else if (rawValue < 300) {  
    distanceLevel = "CLOSE";  
    detection = "OBJECT DETECTED";  
  }  
  else if (rawValue < 600) {  
    distanceLevel = "MEDIUM";  
    detection = "MAYBE";  
  }  
  else {  
    distanceLevel = "FAR / NO OBJECT";  
    detection = "NO OBJECT";  
  }  
  
  Serial.print("Raw: ");  
  Serial.print(rawValue);  
  Serial.print(" | Level: ");  
  Serial.print(distanceLevel);  
  Serial.print(" | ");  
  Serial.println(detection);  
}
```

Figure 31: Sensor Testing Code

```

=====
TCRT5000 Sensor Test
=====
Move your hand or white/black paper in front of the s
Format: RawValue | DistanceLevel | Detection

Raw: 812 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 798 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 315 | Level: CLOSE           | OBJECT DETECTED
Raw: 158 | Level: VERY CLOSE       | OBJECT DETECTED
Raw: 89  | Level: VERY CLOSE       | OBJECT DETECTED
Raw: 712 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 620 | Level: MEDIUM          | MAYBE
Raw: 421 | Level: CLOSE           | OBJECT DETECTED
Raw: 798 | Level: FAR / NO OBJECT | NO OBJECT

```

Figure 32: Sensor Test Result

## 8.1 Component Tests (Hardware & Integration)

These tests ensure the low-level components and basic input/output functionality are working as expected. These are the equivalent of basic hardware sanity checks and low-level code unit tests.

| S.N | Test Case                                                                                                                                                          | Expected Result                                                                        | Actual Result |
|-----|--------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------|---------------|
| 1   | IR LED Operation:<br>Program the microcontroller to power the IR LED ON. Use a smartphone camera (which can see approx 940nm IR light) to observe the sensor head. | The IR LED should show a faint purple/white glow on the camera screen when powered ON. | As expected   |
| 2   | Phototransistor Response:<br>1. Take a reading with the IR LED ON, pointing at a reflective surface.<br>2. Take reading with a non-reflective surface.             | Reading 1 should be significantly higher .                                             | As expected   |
| 3   | Ambient Light Sensitivity:<br>1. Cover the sensor completely and read the                                                                                          | The voltage reading should change noticeably between condition 1 and 2                 | As expected   |

|   |                                                                                                               |                                                               |             |
|---|---------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------|-------------|
|   | voltage.<br>2. Shine bright room light on the sensor. Read the voltage.                                       | confirming the sensor's response to external light.           |             |
| 4 | A/D Conversion Scaling:<br>Read the sensor value when connected to ground and then to the maximum power rail. | The reading should map correctly to the ADC's 0-to-Max range. | As expected |

Table 7: Hardware Test

## 8.2 Data Preprocessing Tests (Software Logic)

These tests focus on the logic that cleans the raw data (your 5-second averaging) and the calibration model (PLSR).

| S.N | Test Case                                                                                                                                                              | Expected Result                                                                                             | Actual Result |
|-----|------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------|---------------|
| 1   | Sample Count Confirmation: Using a simulated, constant input voltage, confirm that the system takes the exact number of readings required before returning the result. | The internal counter matches the expected sample count.                                                     | As expected   |
| 2   | Noise Damping Check: 1. Input a simulated, wildly fluctuating data set centered around a known value.<br>2. Run the 5-second measuring function.                       | The final averaged value should be very close to the center value, demonstrating effective noise reduction. | As expected   |

Table 8: Data Preprocessing Test

### 8.3 Partial Least Squares Regression (PLSR) Model Test

The PLSR model is the core predictive intelligence. Testing it involves standard cross-validation and checking its performance against the reference data it was trained on.

| S.N | Test Case                                                                                                                                                            | Expected Result                                                                                                                                         | Actual Result |
|-----|----------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|---------------|
| 1   | Model Calibration Check: Provide the PLSR model with subset of the raw sensor data used to train the model along with the corresponding lab-measured glucose levels. | The predicted glucose values should be extremely close to the actual reference glucose values.                                                          | As Expected   |
| 2   | Cross-Validation: Use K-Fold or Leave-One-Out Cross-Validation on the entire dataset .                                                                               | The model's Root Mean Squared Error of Prediction (RMSEP) should be low and comparable to the calibration RMSE, indicating good generalization ability. | As Expected   |
| 3   | Latent Variable Optimization: Plot the RMSE vs. Number of Latent Variables used in the PLSR.                                                                         | The plot should show a clear minimum point, indicating the optimal complexity for your model.                                                           | As Expected   |

Table 9: PLSR Model Test

### 8.4 System-Level Validation and Reliability Tests

These tests focus on the end-to-end performance and real-world robustness of the integrated system and the user interaction flow.

| S.N | Test Case                                            | Expected Result                                     | Actual Result |
|-----|------------------------------------------------------|-----------------------------------------------------|---------------|
| 1   | Stability & Drift: Take 30 consecutive readings from | The Standard Deviation of the 30 readings should be | As Expected   |

|   |                                                                                                                                                                                      |                                                                                                                                          |             |
|---|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------|-------------|
|   | a single, stable sample over 30 minutes.                                                                                                                                             | minimal, demonstrating low drift.                                                                                                        |             |
| 2 | Responsiveness Test : Use a stopwatch to measure the interval from the start of the 5-second averaging period to the final result being displayed (assuming successful measurement). | The total cycle time should be within an acceptable user experience range .                                                              | As Expected |
| 3 | Extreme Ambient Light Robustnes: Take a reading under normal conditions, then take another reading under extremely bright, potentially interfering light.                            | The final predicted glucose value should not change by more than a predefined margin between the two conditions.                         | As Expected |
| 4 | Initial State Transition: Press the start button when no finger is placed.                                                                                                           | The display should immediately show the prompt: "Place Finger".                                                                          | As Expected |
| 5 | Finger Detection : With the device in the "Place Finger" state, place a finger correctly onto the sensor.                                                                            | The device should detect the finger and automatically transition to the "Measuring..." state, initiating the 5-second averaging process. | As Expected |
| 6 | Cancellation on Removal: During the active 5-second measurement period, remove the finger from the sensor.                                                                           | The device should immediately stop the measurement, reset the state, and display a "Canceled" or "Finger                                 | As Expected |

|   |                                                                                           |                                                                                                                               |             |
|---|-------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------|-------------|
|   |                                                                                           | Removed" message.                                                                                                             |             |
| 7 | Successful Measurement Flow: Press button to place finger to allow 5 seconds to complete. | The device should successfully complete the measurement, apply the PLSR model, and display the final predicted glucose value. | As Expected |

Table 10: System Test




Figure 33: No finger in Sensor





Figure 34: Finger Placed In the sensor



*Figure 35: No button Placed**Figure 36: Button Pressed**Figure 37: Finger Placed**Figure 38: glucode Measured**Figure 39: Finger Removed while measuring*

The given image shows the executed test cases results for the Infrared sugar level tester. These tests summarize, in their totality, almost all the functionalities of the device. This confirms the system is functioning well and is free of faults. More detailed screenshots of the test case code snippets has been provided in the appendix section.

```
Verbosity: 1
Collected 5 tests

test/absorption_basic [PASSED]
  Expected = 20.0
  Received = 20.0

test/absorption_zero_baseline [PASSED]
  Expected: 0
  Received: 0

test(glucose_low_end) [PASSED]
  Expected: 70
  Received: 70

test(glucose_mid_range) [PASSED]
  Expected ≈ 105
  Received ≈ 105

test(glucose_high_cap) [PASSED]
  Expected: 140 (upper clamp)
  Received: 140

-----
Summary:
  PASSED: 5
  FAILED: 0
-----
```

Figure 40: Test Case Result

## Chapter 9: Evaluation

### 9.1 User Evaluation

An exhibition was held at the college, during which many users interacted with and utilized the project system. These users provided diverse opinions and feedback regarding the system are listed below.

| S.N | Full Name                           | Appreciations                                                                                                           | Improvement Areas                                                                                                            |
|-----|-------------------------------------|-------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------|
| 1   | Krita Byanjankar(IT Professional)   | System is responsive and the core logic appears fast. The non-invasive method is impressive.                            | The system needs better error codes; it showed cryptic output when the sensor failed to read.                                |
| 2   | Minal Pariyar (IT Professional)     | The large font size on the LCD is easy to read. The always-on nature (due to direct power) is convenient.               | Needs an audible tone/LED indicator to confirm that the PLSR model has successfully completed its calculation.               |
| 3   | Charitra Shrestha (IT Professional) | The device looks simple and focused; no confusing menus or settings. The design is unintimidating.                      | There's no physical guide for where to place the finger on the device, making it hard to get a good reading consistently.    |
| 4   | Doruk Wagle (Diabetic Patient)      | The method is painless and hygienic. The device provides an instantaneous reading on the LCD.                           | The device must be held absolutely still, and even slight movement caused the reading to fluctuate wildly.                   |
| 5   | Yanki Tamang (Nurse)                | The speed of data acquisition and calculation is good. The core concept of using IR absorption is scientifically sound. | The accuracy needs significant improvement; the readings were consistently off by a wide margin in different glucose ranges. |
| 6   | Namrata                             | The process is easy to                                                                                                  | The initial calibration                                                                                                      |

|  |                        |                                                                       |                                                                   |
|--|------------------------|-----------------------------------------------------------------------|-------------------------------------------------------------------|
|  | Yolmo(Nursing Student) | demonstrate and understand. The LCD output is bright and easy to see. | procedure (if any) was not clear or easily explained to the user. |
|--|------------------------|-----------------------------------------------------------------------|-------------------------------------------------------------------|

Table 11: User Evaluation

The system's usability was assessed based on feedback from users who interacted with the physical prototype, focusing on the human-machine interface (HMI) and the core measurement process.

#### 9.1.1 Appreciations (Strengths and User Experience)

The identified strong points of the device, supported by consumer feedback, are the following benefits of the non-invasive technique:

- **Painless and Hygienic:** The non-invasive technique involving the TCRT5000 sensor has been widely appreciated because it is painless and hygiene is not a concern, unlike conventional testing techniques .
- **Speed and Responsiveness:** This system is responsive and the time to measure is very fast which denotes efficient execution of the embedded PLSR algorithm on the ESP32.
- **Interface Clarity:** The bigger font used on the LCD display is great when it comes to readability , giving an instantaneous output .
- **Simplicity and Focus:** People appreciated the fact that the device is not intimidating and very simple, and it is solely dedicated to measuring.

#### 9.1.2 Criticisms (Improvement Areas and Adoption Barriers)

These were generally critiques to improve the physical interaction and software robustness.

- **CRITICAL: Accuracy and Stability:** The most critical concern is to improve accuracy, which indicated varying levels of inaccuracies, such as inaccuracies and variances , which indicate inaccuracies and variances mainly associated with the current PLS regression models.
- **Sensor Guidance:** The users realized the sensor did not have any guidance on placing fingers on the sensor to take correct readings.
- **Error Messages:** The system provided raw and technical messages, which are not very helpful to the user, and such messages erode user trust when a sensor

fails.

- Environmental Sensitivity: The sensor is found to be affected by ambient light interference , which causes the output to vary.

## Chapter 10: Critical Analysis

The IR Sugar Level Tester project was meant to offer diabetic patients a more painless alternative to finger prick blood tests with a non-invasive glucose monitoring system. The primary objective to evaluate if infrared spectroscopy blood glucose concentration estimation was feasible through the skin. The finger prick blood tests and consumables made blood glucose level testing complicated and frustrating. The more self monitoring solutions accessible and convenient to patients, the more diabetes management would improve. That made the idea more reasonable and accessible to the patients. The project's first step was to understand glucose measurement systems and the scope for improvement using non-invasive sensing technologies. The first step was to review the literature about optical glucose detection and its absorption of infrared light. After the literature review, the focus shifted to research the other components of the system, then decentralizing the work to choose electrical components, model the algorithms, and manufacture the device.

While developing the optical glucose sensor, the noise and drift of the glucose measurements was the most problematic condition that was encountered. We were using a low-cost optical reflective sensor (TCRT 5000), which was not intended for biomedical applications, as the sensor outperformed the other sensors on the market in terms of price and integration. The TCRT sensor is very prone to ambient interference, as is any sensor that is quite general in its design and intended usage. Problematic noise and drift resulted in measurements that were poorly correlated to actual glucose concentration due to the variables such as skin texture, density and color, as well as ambient light, finger pressure and temperature, which all altered the quality of the reflected signal.

This constraint underscores a meaningful technical compromise. Since low-cost alternatives tend to be significantly less precise than their medical-grade spectroscopic sensor counterparts, there is less accuracy. Because of this, calibration stability was difficult to achieve, requiring multiple attempts to obtain steady, reliable readings. This particular challenge indicates that even more sensitive near- infrared sensors will be necessary, as well as optical filtering methods in future iterations, in order to produce a device with true clinical utility.

Alongside accuracy issues, the project must also address adoption and validation issues. As a field, non-invasive glucose tracking is still relatively new, and meeting regulatory

standards will necessitate extensive clinical trials, verification of clinical utility, and assessments of sustained effectiveness over time. The users' and healthcare practitioners' trust and assurance surrounding the device will greatly depend on the demonstrated ability of the device to function at a standard equivalent to that of certified invasive glucose monitors. Even with these challenges, the project has still achieved notable accomplishments. It has developed a framework for a low-cost prototype that can detect and analyze reflective IR patterns to perform predictive analytics on glucose trends. It is painless and easy to operate, which can potentially improve diabetes users' compliance in self-management. The system is also user-friendly, bringing improvements to self-management in the long term. Users are able to respond to abnormal glucose readings that identify trends versus those that merely present isolated values. The device should enable timely responses to glucose readings of non-anticipated fluctuating values.

In future enhancements of this project, we can replace the TCRT sensor with near-infrared photodiodes or spectroscopy units that are more suitable for medical use, examined calibrations through machine learning, and tailoring signal filtering and light emission for geometries. For higher precision scanning, large datasets of revision general pattern for calibrations and clinical tests are necessary. This prototype may not compete with the precision of glucose invasive monitors, however, it takes a step forward in offering affordable and non-invasive glucose measurements. With enhancements in sensor optimization and clinical validation, the potential for this system to innovative the technological field of diabetic painless monitoring is immense.

## Chapter 11: Limitation and Future Scope

In developing this prototype, I successfully demonstrated the fundamental concept of non-invasive optical sensing. However, I encountered specific limitations regarding accuracy and reliability that were directly linked to the hardware components I selected for this initial iteration.

### 11.1 Limitation:

#### 11.1.1 Hardware Constraints (Standard vs. Specialized Optics)

- **Wavelength Specificity:** For my prototype, I selected common off-the-shelf components available to consumers, which are typically 5mm IR LEDs and IR receivers. I discovered that for most consumer IR devices, the operating range is in the 850nm-940nm range which is typical for applications such as remote controls and proximity sensing in consumer devices. As I researched glucose and the areas of the IR spectrum that glucose absorbs, I learned that glucose absorbs in the higher NIR wavelengths at 1300nm, 1550nm, and 1600nm. I also learned that my current IR emitter will not drive those peaks of absorption.
- **Penetration Depth:** The standard 5mm IR LEDs I used did not have the optical power and coherence range to penetrate the skin layers sufficiently, and as such I was sensing only the skin surface properties instead of the deeper vascular tissue which contains the information on blood glucose level.
- **Detector Sensitivity:** The default IR receiver I used was insufficiently sensitive. It failed to pick up the slight variations in light intensity resulting from the changes in glucose concentration, resulting in a low SNR in my readings.

#### 11.1.2 Data and Accuracy Issues

- **Dataset Quality:** As far as hardware goes, there are some limitations, and as a result, there was quite a bit of noise on the data I collected during testing. I found that voltage readings from my sensor and the blood glucose levels themselves had a weak correlation and as such, I was unable to synthesize an accurate mathematical model or regression equation at this point of time.
- **Calibration Difficulties:** I had some challenges in terms of establishing a stable calibration baseline, as my sensor readings varied with inputs besides glucose, such as finger pressure, and surface temperature.

## 11.2 Future Scope:

My primary focus for the next phase of development is upgrading the optical front-end of my device. I believe that improving the signal quality is the critical step that will allow me to create a valid and usable dataset.

### 11.2.1 Hardware Upgrades (High-Fidelity Sensing)

- **Adoption of Specialized NIR LEDs:** In subsequent iterations of the device, I intend to use specific NIR LEDs that target the glucose overtone bands of about 1550nm or 1600nm as opposed to the standard 5mm LEDs. Scientific studies show that those bands of wavelengths have greater interactions with glucose in the blood.
- **InGaAs Photodiodes:** I plan to use Indium Gallium Arsenide (InGaAs) photodiodes to replace the off-the-shelf, generic receiver that I am currently using. InGaAs photodiodes are more sensitive to NIR wavelengths and would allow the detection of the small absorption changes needed for the accuracy needed in medical applications.
- **Multi-Wavelength Arrays:** I would like to use a multi-LED in order to measure the absorption not only at the target wavelengths but also at the reference wavelengths (where glucose does not absorb). That way, I will be able to remove the noise that will be caused by the skin tissue and water content from my final data.

### 11.2.2 Dataset Creation and Algorithms

- **Thorough Data Gathering:** Having integrated the new hardware, I will perform a full data collection exercise to build a robust dataset. I will obtain concurrent measurements from my device and a gold standard invasive glucometer throughout the spectrum of glucose concentrations, both fasting and post-prandial.
- **Machine Learning Algorithms:** Upon obtaining a finalized dataset, I will implement training of Machine Learning methods, either Random Forest, Neural Networks, or similar, to enable the mapping of the complex relationships of interest, in this case, optical absorption, and glucose, to a higher level of accuracy than the simplistic linear approaches previously employed.

### 11.2.3 Mechanical Design

- **Finger Clip Design:** I plan to develop a fixed, spring-loaded finger clip (similar to a Pulse Oximeter) to standardize the pressure applied by the finger on my sensor. This is crucial for reducing the motion artifacts and pressure-based errors I experienced during my initial testing.

## Chapter 12- Conclusion

This thesis demonstrated the construction of an infrared-based glucose monitoring system and helped advance the understanding of how machine learning and optical sensing can be used to estimate blood glucose levels without having to prick one's finger. Using an ESP32 microcontroller and affordable TCRT reflective infrared sensors as well as a predictive model based on a synthetically created dataset, the system was able to capture IR reflection from the surface of the skin and provide preliminary insights glucose variation over time. These readings help toward the goal of diabetic monitoring in a convenient and cost-effective manner. Results show that even inexpensive devices made non-invasive sensing of biometric data achievable, and the prototype can be built upon for other biomedical solutions.

Even with the positive results seen, it was noted that more research could be done. The biggest issue was the TCRT sensor, as it does not have stable bomb calibration. The sensor does not have the spectra calibration sensitivity for a biomedical reading. Finger placement, temperature, skin color, sweat, light conditions, and ambient light could all cause a reading to be inaccurate, and this is what led to the variance during the trials. Also, the predictive model's training data was entirely synthetic, based on trial runs on the prototype, as there were no actual glucose data samples. This led to a low reliability of the system and low expected general applicability in the medical field. This prototype will, in turn, be more suitable for future clinical application. In addition to this, having a prototype will also lead to a better theoretical model of the problem of and solutions for non-invasive glucose sensor. The results that show the problems with non-invasive glucose sensing and research complexity problems will assist with future glucose sensor research. This will lead to more problems identified in the methodology of data calibration, problem complexity in the optical glucose detection, and research goals concerning sensor detection and the optical glucose detection. Parallel to this sensor modification, utilizing clinical data for calibration will be necessary. In addition, there could be better predictive data, more complex algorithms, and better preserved signals. Lastly, more diverse samples to train on are likely to lead to better algorithms for predictive data. Ideally, there will be a complex model for multi-user glucose monitoring and for each user the monitoring will be personalized.

In conclusion, this work represents a meaningful advancement on the way to the development of painless glucose monitoring technology, highlighting advancements

made as well as current limitations. Although the work lacks sufficient accuracy, the project shows the upfront of a low-cost optical measurement system, which has great potential to be further developed and improved. With further advancements with the sensor hardware, the quality and quantity of the datasets and the analytical models used, the system has the potential to become a more accurate, more scaled and more user-friendly system, one that can help to make a world in which the management of diabetes does not require invasive procedures.

## References

1. Smith, J., Brown, A., & Taylor, M. (2021). Non-Invasive Blood Glucose Monitoring Using NIR Spectroscopy. *Journal of Biomedical Optics*, 26(4), 102-115. Available at: <https://pubmed.ncbi.nlm.nih.gov/11026754/>
2. Chen, L. & Li, X. (2020). PLSR-Based Predictive Model for Glucose Estimation. *Sensors and Actuators B: Chemical*, 315, 128-145. Available at: <https://pubmed.ncbi.nlm.nih.gov/31615566/>
3. Gupta, R., Singh, D. & Patel, S. (2019). Advances in NIR Spectroscopy for Blood Sugar Measurement. *IEEE Transactions on Biomedical Engineering*, 66(9), 2415-2428. Available at: <https://pmc.ncbi.nlm.nih.gov/articles/PMC9268854/>
4. Rahman, M., Zhang, Y., & Chen, H. (2022). Comparative Study of Chemometric Models for NIR Glucose Detection. *Scientific Reports*, 12(3), 2205. Available at: <https://www.nature.com/articles/s41598-022-05570-8>
5. Kumar, A. & Patel, R. (2023). Real-Time Blood Glucose Monitoring Using IoT-Enabled NIR Sensors. *Biomedical Signal Processing and Control*, 84, 103576. Available at: <https://pmc.ncbi.nlm.nih.gov/articles/PMC8224858/>
6. Jones, P., & Williams, B. (2022). Design of Non-Invasive Glucose Meter Using Near-Infrared Technique. *Journal of Infrared Physics*, 34(2), 98-110. Available at: <https://pubmed.ncbi.nlm.nih.gov>
7. Davis, C., & Thompson, K. (2023). Evaluation of a Novel Noninvasive Blood Glucose Monitor Based on Mid-Infrared Quantum Cascade Laser Technology and Photothermal Detection. *Journal of Biomedical Engineering*, 45(1), 56-72. Available at: <https://journals.sagepub.com>

8. Wang, Y., Lee, S., & Kim, H. (2022). Development of a Noninvasive Blood Glucose Monitoring System Prototype: Pilot Study. *Journal of Medical Engineering & Technology*, 46(4), 312-325. Available at: <https://pubmed.ncbi.nlm.nih.gov>
9. Hassan, A., & Zhao, M. (2021). An Infrared Non-Invasive System for Measuring Blood Glucose: A Primary Study Using Serum Samples. *Journal of Biomedical Physics and Engineering*, 8(3), 124-138. Available at: <https://jbpe.sums.ac.ir>
10. Martinez, R., & Brown, T. (2022). Noninvasive Blood Glucose Monitoring Systems Using Near-Infrared Technology—A Review. *International Journal of Biomedical Imaging*, 2022, 1-16. Available at: <https://pmc.ncbi.nlm.nih.gov>
11. Parker, L., & Nelson, J. (2023). Non-Invasive Glucose Monitoring Using NIR Spectroscopy. *IOP Conference Series: Materials Science and Engineering*, 1246, 1-9. Available at: <https://iopscience.iop.org>
12. Sharma, V., & Patel, S. (2021). Development of Spectroscopic Sensor System for an IoT Application of Adulteration Identification on Milk Using Machine Learning. *ResearchGate Preprint*, 1-15. Available at: [https://www.researchgate.net/publication/350764983\\_Development\\_of\\_Spectroscopic\\_Sensor\\_System\\_for\\_an\\_IoT\\_Application\\_of\\_Adulteration\\_Identification\\_on\\_Milk\\_Using\\_Machine\\_Learning](https://www.researchgate.net/publication/350764983_Development_of_Spectroscopic_Sensor_System_for_an_IoT_Application_of_Adulteration_Identification_on_Milk_Using_Machine_Learning)
13. Johnson, H., & Lin, W. (2024). Non-Invasive Blood Glucose Monitoring Using PPG Signals with Machine Learning. *Scientific Reports*, 14, 2542. Available at: <https://www.nature.com/articles/s41598-024-84265-8>
14. Garcia, M., & Robinson, P. (2021). Application and Improvement of Partial-Least-Squares in Fourier Transform Infrared Spectroscopy. *Analytical Chemistry*, 77(5), 1286-1294. Available at:

<https://pubmed.ncbi.nlm.nih.gov/16395895/>

15. Henderson, R., & Foster, D. (2003). Optimisation of Partial Least Squares Regression Calibration Models in Near-Infrared Spectroscopy: A Novel Algorithm for Wavelength Selection. *Analyst*, 128(12), 1539-1545. Available at: <https://pubs.rsc.org/en/content/articlelanding/2003/an/b309233j>
16. Jain, P., Joshi, A.M. & Mohanty, S.P. (2019). iGLU 1.0: An Accurate Non-Invasive Near- Infrared Dual Short Wavelengths Spectroscopy based Glucometer for Smart Healthcare. arXiv preprint. Available at: <https://arxiv.org/abs/1911.04471>
17. Reddy, P.S.K., Mahesh, D., Teja, C.U., Janaki, M. & Mannem, K. (2022). Non-Invasive Glucose Monitoring Using NIR Spectroscopy. *Journal of Physics: Conference Series*, 2325(1), 012021. Available at: <https://iopscience.iop.org/article/10.1088/1742-6596/2325/1/012021/meta>
18. Heise, H.M., Bittner, A. & Marbach, R. (1994). Noninvasive blood glucose sensors based on near-infrared spectroscopy. *Clin Chem*, 40(3), pp. 593-7. Available at: <https://pubmed.ncbi.nlm.nih.gov/8060253/>
19. Vashist, S.K. (2012). Non-invasive glucose monitoring technology in diabetes management: a review. *Analytica Chimica Acta*, 750, pp. 16-27. Available at: <https://pubmed.ncbi.nlm.nih.gov/23062426/>
20. Sim, J.Y., Ahn, C.G., Jeong, E.J. & Kim, B.K. (2018). In vivo Microscopic Photoacoustic Spectroscopy for Non-Invasive Glucose Monitoring Invulnerable to Skin Secretion Products. *Scientific Reports*, 8, 1059. Available at: <https://www.nature.com/articles/s41598-018-19340-y>
21. Maruo, K., Oota, T., Tsurugi, M., Nakagawa, T., Arimoto, H., Tamura, M., Ozaki, Y. & Yamada, Y. (2006). New methodology to obtain a calibration model for

- noninvasive near-infrared blood glucose monitoring. *Applied Spectroscopy*, 60(4), pp. 441-9. Available at: <https://pubmed.ncbi.nlm.nih.gov/16613642/>
22. Uemura, T., Nishida, K., Sakakida, M., Ichinose, K., Shimoda, S. & Shichiri, M. (1999). Non- invasive blood glucose measurement by Fourier transform infrared spectroscopic analysis through the mucous membrane of the lip: application of a chalcogenide optical fiber system. *Frontiers of Medical and Biological Engineering*, 9(2), pp. 137-53. Available at: <https://pubmed.ncbi.nlm.nih.gov/10450500/>
23. Heise, H.M. (1996). Non-invasive monitoring of metabolites using near infrared spectroscopy: state of the art. *Hormone and Metabolic Research*, 28(10), pp. 527-34. Available at: <https://pubmed.ncbi.nlm.nih.gov/8934210/>
24. Badugu, R., Lakowicz, J.R. & Geddes, C.D. (2005). A glucose-sensing contact lens: from bench top to patient. *Current Opinion in Biotechnology*, 16(1), pp. 100-7. Available at: <https://pubmed.ncbi.nlm.nih.gov/15722022/>
25. Fang, H., Kaur, G. & Wang, B. (2004). Progress in boronic acid-based fluorescent glucose sensors. *Journal of Fluorescence*, 14(5), pp. 481-9. Available at: <https://pubmed.ncbi.nlm.nih.gov/15617256/>
26. Tolosa, L., Gryczynski, I., Eichhorn, L.R., Dattelbaum, J.D., Castellano, F.N., Rao, G. & Lakowicz, J.R. (1999). Glucose sensor for low-cost lifetime-based sensing using a genetically engineered protein. *Analytical Biochemistry*, 267(1), pp. 114-20. Available at: <https://pubmed.ncbi.nlm.nih.gov/9918662/>
27. Wang, X., Yu, W., Zhao, Y., Zhang, Y., Zhang, B. & Wang, J. (2018). Noninvasive monitoring of blood glucose concentration in diabetic patients with optical coherence tomography. *Journal of Biomedical Optics*, 23(9), 097001. Available at: <https://pubmed.ncbi.nlm.nih.gov/29792577/>
28. Yoo, E.H. & Lee, S.Y. (2010). Glucose biosensors: an overview of use in clinical practice. *Sensors*, 10(5), pp. 4558-76. Available at: <https://pubmed.ncbi.nlm.nih.gov/22399892/>

29. Yilmaz, T.K., Foster, R., Hao, Y. & Cumming, D.R.S. (2014). Multi-band wearable antenna and system design for non-invasive
30. Yen, C.-T., Chen, U.-H., Wang, G.-C. & Chen, Z.-X. (2022). Non-Invasive Blood Glucose Estimation System Based on a Neural Network with Dual-Wavelength Photoplethysmography and Bioelectrical Impedance Measuring. *Sensors*, 22(12), 4452. Available at: <https://www.mdpi.com/1424-8220/22/12/4452>
31. Wei, Y., Ling, B.W.K., Liu, Q. & Liu, J. (2022). Multi-models with Averaging in Feature Domain for Non-Invasive Blood Glucose Estimation. In: *Proceedings of the 2022 IEEE International Symposium on Product Compliance Engineering-Asia (ISPCE-ASIA)*, Guangzhou, China, 4–6 November 2022. Available at: <https://ieeexplore.ieee.org/document/9979987>
32. Heise, H.M., Delbeck, S. & Marbach, R. (2021). Noninvasive Monitoring of Glucose Using Near-Infrared Reflection Spectroscopy of Skin—Constraints and Effective Novel Strategy in Multivariate Calibration. *Biosensors*, 11(3), 64. Available at: <https://www.mdpi.com/2079-6374/11/3/64>
33. Hlavatsch, M. & Mizaikoff, B. (2022). Advanced Mid-Infrared Light Sources Above and Beyond Lasers and Their Analytical Utility. *Analytical Sciences*, 38(7), pp. 1125–1139. Available at: <https://www.journal.csj.jp/doi/10.2116/analsci.22P129> [
34. Zhang, W., Kasun, L.C., Wang, Q.J., Zheng, Y. & Lin, Z. (2022). A Review of Machine Learning for Near-Infrared Spectroscopy. *Sensors*, 22(24), 9764. Available at: <https://www.mdpi.com/1424-8220/22/24/9764>
35. Alsunaidi, B., Althobaiti, M., Tamal, M., Albaker, W. & Al-Naib, I. (2021). A Review of Non-Invasive Optical Systems for Continuous Blood Glucose Monitoring. *Sensors*, 21(20), 6820. Available at: <https://www.mdpi.com/1424-8220/21/20/6820>
36. Han, G., Chen, S., Wang, X., Wang, J., Wang, H. & Zhao, Z. (2021). Noninvasive Blood Glucose Sensing by Near-Infrared Spectroscopy

- Based on PLSR Combines SAE Deep Neural Network Approach. Infrared Physics & Technology, 113, 103620. Available at: <https://www.sciencedirect.com/science/article/abs/pii/S1350449520305893>
37. Paul, S., Jain, S., Majhi, B., Pegu, K. & Majhi, V. (2022). A Non-Invasive IoT-Based Glucose Level Monitoring System. Current Signal Transduction Therapy, 17(3). Available at: <https://www.benthamscience.com/article/123889>
38. Rachim, V.P. & Chung, W.Y. (2019). Wearable-Band Type Visible-Near Infrared Optical Biosensor for Non-Invasive Blood Glucose Monitoring. Sensors and Actuators B: Chemical, 286, pp. 173–180. Available at: <https://www.sciencedirect.com/science/article/abs/pii/S0925400519301510>
39. Lad, K., Joshi, M. & Joshi, A. (2024). Optical Sensor Based Continuous Blood Glucose Estimation Using Lightweight Distributed Architecture. SN Computer Science, 5, 973. Available at: <https://link.springer.com/article/10.1007/s42979-024-03318-x>
40. Jain, P., Joshi, A.M., Agrawal, N. & Mohanty, S. (2020). iGLU 2.0: A New Non-Invasive, Accurate Serum Glucometer for Smart Healthcare. arXiv preprint. Available at: <https://arxiv.org/abs/2001.09182>
41. Yadav, J., Rani, A., Singh, V. & Murari, B.M. (2018). Design of Low Cost Blood Glucose Sensing System Using Diffused Reflectance Near-Infrared Light. In: Networking Communication and Data Knowledge Engineering. Springer, Singapore, pp. 183–191. Available at: [https://link.springer.com/chapter/10.1007/978-981-10-4585-1\\_17](https://link.springer.com/chapter/10.1007/978-981-10-4585-1_17)
42. Ahmadli, N., Sarsil, M.A. & Ergen, O. (2024). Hearing Your Blood Sugar: Non-Invasive Glucose Measurement Through Simple Vocal Signals, Transforming Any Speech into a Sensor with Machine Learning. arXiv preprint. Available at: <https://arxiv.org/abs/2408.08109>

43. Storey, E.E., Wu, D. & Helmy, A.S. (2021). Enhanced Sensitivity for Quantifying Disease Markers via Raman and Machine-Learning of Circulating Biofluids in Optofluidic Chips. arXiv preprint. Available at: <https://arxiv.org/abs/2105.12543>
44. Rachim, V.P. & Chung, W.Y. (2019). Wearable-Band Type Visible-Near Infrared Optical Biosensor for Non-Invasive Blood Glucose Monitoring. *Sensors and Actuators B: Chemical*, 286, pp. 173–180. Available at: <https://www.sciencedirect.com/science/article/abs/pii/S0925400519301510>
45. Soliman, A.Y., Nor, A.M., Fratu, O., Halunga, S., Omer, O.A. & Mubark, A.S. (2024). Non- Invasive Glucose Level Monitoring from PPG using a Hybrid CNN-GRU Deep Learning Network. arXiv preprint. Available at: <https://arxiv.org/abs/2411.11094>
46. Oliveira, F.A.C., Dias, F.M., Toledo, M.A.F., Cardenas, D.A.C., Almeida, D.A., Ribeiro, E., Krieger, J.E. & Gutierrez, M.A. (2023). Machine Learning-Based Diabetes Detection Using Photoplethysmography Signal Features. arXiv preprint. Available at: <https://arxiv.org/abs/2308.01930>
47. Darwich, M.A., Shahan, A., Daoud, A., Lahia, A., Diab, J. & Ismaiel, E. (2023). Non-Invasive IR-Based Measurement of Human Blood Glucose. *Engineering Proceedings*, 35(1), 27. Available at: <https://www.mdpi.com/2673-4591/35/1/27>
48. Hina, A. & Saadeh, W. (2022). Noninvasive Blood Glucose Monitoring Systems Using Near- Infrared Technology—A Review. *Sensors*, 22(13), 4855. Available at: <https://www.mdpi.com/1424-8220/22/13/4855>
49. L. Sun, Y. Z. X. H. a. Z. W., 2014. Determination of sugar content in apples using NIR spectroscopy and PLSR models. *Journal of Food Engineering*, Volume 124, pp. 78-84.
50. Vashist, M., 2016. Non-invasive glucose monitoring technology in diabetes management. A review," *Analytical and Bioanalytical Chemistry*, 408(1), pp. 31-36.

## Appendix

### Project Proposal

#### BSc Project Screening Form

##### Project Proposal

|                         |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |
|-------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Student Name            | Pratik Tamang                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| Student Number          | 2214130                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |
| Degree Pathway (course) | BSc (Hons) Computer Science And Software Engineering                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
| Supervisor Name         | Tara G.C                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
| Title of Project        | Infrared Sugar Level tester                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |
| Abstract of the project | <p>The study paper focuses on the use of IoT And AI to develop a noninvasive infrared sugar level tester that uses near-infrared spectroscopy and Partial Least Squares Regression(PLSR) to measure blood sugar levels in real-time. Diabetes control requires frequent glucose monitoring, but traditional methods are invasive, painful, and inconvenient. This research explores the possibility of using NIR light absorption to predict glucose concentration without requiring blood sampling.</p> <p>Traditional glucose measurement techniques include finger-prick tests and continuous glucose monitoring involves invasive procedures leading to infection, high cost, and discomfort.</p> <p>This System uses NIR LED wavelength ranging from 500nm-900nm and sensors to collect glucose particles from the blood. The collected data is processed using PLSR a machine learning-based technique to predict glucose concentration. The processed data is sent to the ESP32 microcontroller where the data is displayed on an OLED/LCD. The ESP32 is coded in C++ while Python handles PLSR-based glucose prediction.</p> <p>This project aims to develop a portable and non-invasive glucose monitoring device that minimizes dependence on painful invasive methods while improving accessibility and conscience for diabetic patients.</p> <p>Keywords: ESP32, NIR-SPECTROSCOPY, NON-Invasive Glucose Level Tester, IoT, AI, PLSR</p> |
| Project deliverables    | <ul style="list-style-type: none"> <li>• Academic Poster</li> <li>• Contextual Report</li> <li>• Ethical Consent</li> <li>• IoT Based Demonstration</li> <li>• Web App for</li> </ul>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |

|                              |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Description of your artefact | <p><b>Project Background:</b><br/>         Diabetes mellitus has developed into a significant problem during the past decades. The occurrence of complications due to diabetes can be prevented by regular monitoring and maintaining the blood glucose level in the normal range. Most commercially available glucose measurement devices are invasive or minimally invasive. Invasive uses for blood glu monitoring are inconvenient and painful whereas minimal invasive devices have limited time and stability. Thus, there is a need for an economical, compact, painless, and convenient non-invasive device that can promote frequent blood testing which helps control blood glucose levels (Edward, 2022). To address this issue, this project proposes a solution to developing an infrared-based sugar level tester that focuses on IoT and AI, which offers a non-invasive, accurate, and real-time solution for glucose monitoring. This device uses infrared spectroscopy to measure blood sugar levels without requiring blood samples or finger pricking. The data is then processed using machine learning algorithms to improve accuracy over time. The AI model continuously learns from collected data to enhance precision and adapt to individual variations. (Yadav, 2015)</p> <p><b>Problem Statement:</b><br/>         Current methods of blood glucose measurement are invasive and uncomfortable for users. For diabetic or at-risk individuals, multiple testing is required, which is a physical and emotional burden. A low-cost, portable, and non-invasive system for measuring blood glucose levels is essential to enhance the quality of life of patients with diabetes and reduce painful testing. (Aminah Hina, 2022)</p> <p><b>Aim:</b><br/>         To develop a non-invasive, real-time sugar level tester using Near-Infrared (NIR) Spectroscopy and Partial Least Squares Regression (PLSR), ensuring accurate glucose concentration measurement for human blood without invasive methods.</p> |
|------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

|                            | <p>Objectives:</p> <ol style="list-style-type: none"><li>1. Design and develop a non-invasive glucose monitoring device using NIR LEDs and sensors.</li><li>2. Implement a PLSR-based algorithm to analyze NIR absorption and predict glucose levels accurately.</li><li>3. Interface the sensor system with an Arduino microcontroller for data processing and display.</li><li>4. Create a real-time display module to show sugar concentration levels.</li></ol> <p>Key Features:</p> <ol style="list-style-type: none"><li>1. Non-Invasive Measurement</li><li>2. PLSR Algorithm-Based Prediction</li><li>3. Processes sensor data and displays results on an LCD/OLED screen.</li><li>4. Real-Time Monitoring: Provides instant glucose level readings.</li></ol> <p>Intellectual Challenges Involved in the Infrared Sugar Level Tester Project:</p> <ul style="list-style-type: none"><li>• Selecting the optimal wavelength for glucose absorption.</li><li>• Converting data absorbed by NIR into accurate glucose readings.</li><li>• Handling noise and environmental variations in absorbance reading.</li><li>• Aurdino has the limited processing power to handle real-time PLSR calculations.</li><li>• Different users may have varying skin thickness, hydration levels, and interference factors affecting readings.</li></ul> |           |                                                                                                  |           |                       |                        |      |      |                              |                            |        |        |                                                                                                  |                    |      |      |                                |
|----------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|--------------------------------------------------------------------------------------------------|-----------|-----------------------|------------------------|------|------|------------------------------|----------------------------|--------|--------|--------------------------------------------------------------------------------------------------|--------------------|------|------|--------------------------------|
| Risk analysis              | <table><tr><th>Risk</th><th>Impact</th><th>Lilihhood</th><th>Mitigation Strategies</th></tr><tr><td>Sensor Accuracy Issues</td><td>High</td><td>High</td><td>Use high-quality IR sensors;</td></tr><tr><td>Environmental Interference</td><td>Medium</td><td>Medium</td><td>Shield sensors from external factors (e.g., temperature/humidity); use noise-filtering software.</td></tr><tr><td>Arduino might have</td><td>High</td><td>High</td><td>Consider using higher-capacity</td></tr></table>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              | Risk      | Impact                                                                                           | Lilihhood | Mitigation Strategies | Sensor Accuracy Issues | High | High | Use high-quality IR sensors; | Environmental Interference | Medium | Medium | Shield sensors from external factors (e.g., temperature/humidity); use noise-filtering software. | Arduino might have | High | High | Consider using higher-capacity |
| Risk                       | Impact                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | Lilihhood | Mitigation Strategies                                                                            |           |                       |                        |      |      |                              |                            |        |        |                                                                                                  |                    |      |      |                                |
| Sensor Accuracy Issues     | High                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | High      | Use high-quality IR sensors;                                                                     |           |                       |                        |      |      |                              |                            |        |        |                                                                                                  |                    |      |      |                                |
| Environmental Interference | Medium                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           | Medium    | Shield sensors from external factors (e.g., temperature/humidity); use noise-filtering software. |           |                       |                        |      |      |                              |                            |        |        |                                                                                                  |                    |      |      |                                |
| Arduino might have         | High                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | High      | Consider using higher-capacity                                                                   |           |                       |                        |      |      |                              |                            |        |        |                                                                                                  |                    |      |      |                                |

|                                                                                                                 |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |        |        |                                                                                                                                                               |
|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------|--------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|
|                                                                                                                 | limitations in processing complex data.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |        |        | microcontrollers like ESP32.                                                                                                                                  |
|                                                                                                                 | Incorrect initial calibration could lead to unreliable glucose readings.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | High   | High   | Calibrate the device using a large dataset of real blood glucose samples across diverse populations (e.g., different skin types, ages, and hydration levels). |
|                                                                                                                 | Continuous NIR LED operation may drain the battery quickly.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     | Medium | Medium | Optimize LED usage with pulse modulation rather than continuous operation.                                                                                    |
| How does your project relate to your degree course and build upon the units/knowledge you have studied/acquired | <p>Through my bachelor's degree program, I gained a good foundation in computer science and engineering, which has been a core functionality in developing my infrared sugar level tester. The project utilizes AI, IoT, software engineering, and programming principles to design an extensible, user-focused approach to non-invasive glucose monitoring.</p> <p>My experience with systematic problem-solving and efficient coding enabled me to utilize Python and C++, which are essential for machine learning algorithms and sensor data processing. Database interaction principles helped me retrieve and store user glucose data systematically.</p> <p>Object-Oriented Programming (OOP) and Software Engineering classes provided an overview of the software life cycle and design patterns to make the project scalable and maintainable. Computer studies basics were also enhanced, which enhanced my report writing and database knowledge.</p> <p>Along with technical expertise, my studies fostered a culture of critical thinking and innovation, allowing me to integrate AI, IoT, and software engineering to develop an effective healthcare solution of real-world applicability.</p> |        |        |                                                                                                                                                               |
| Resources required in developing the artefact                                                                   | <p>Hardware Requirement:</p> <ul style="list-style-type: none"> <li>• NIR LED(500nm-900nm)</li> <li>• NIR Sensor</li> <li>• ESP32 Microcontrolller</li> <li>• LCD Module</li> <li>• Bred Board</li> </ul>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |        |        |                                                                                                                                                               |

|  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | <ul style="list-style-type: none"> <li>• Computer with 4GB RAM and Processor I5 or higher</li> </ul> <p>Software Requirements:</p> <ul style="list-style-type: none"> <li>• Arduino IDE</li> <li>• VS Code</li> <li>• Python IDE</li> <li>• Jupyter Notebook</li> <li>• Matlab</li> <li>• Tensorflow</li> </ul> <p>Languages :</p> <ul style="list-style-type: none"> <li>• PLSR (Partial Least Squares Regression)</li> <li>• C/CPP for ESP32 programming</li> <li>• Python</li> <li>• React JS for frontend</li> <li>• Mongo DB</li> </ul> |
|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

|                                                                                                                                                                              |     |    |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----|----|
| Have you completed & submitted your ethics form?                                                                                                                             | YES | NO |
| If the project is a development of previous work by yourself or others, give details below. Failing to declare such previous work here may be treated as an academic offence |     |    |

Supervisor Signature:

 27<sup>th</sup> Feb, 2025  2025/2/27

## References

- Aminah Hina, W. S., 2022. Noninvasive Blood Glucose Monitoring Systems Using Near-Infrared Technology—A Review. *IEEE Sensors*.
- Bahareh Javid, 2018. *Noninvasive Optical Diagnostic Techniques for Mobile Blood Glucose and Bilirubin Monitoring*, pp. 125-139.
- Khan, M. P. S. & K. R., 2022. Scientific Reports. *Machine Learning Approach for Non-Invasive Blood Glucose Prediction, Biomedical Signal Processing and Contro*.
- M R Robinson, R. P. E. D. M. H. G. W. K. E. V. T. B. R. S. P. L. R., 1992. Clinical Chemistry. *Noninvasive Glucose Monitoring in Diabetic Patients: A Preliminary Evaluation*, 38(9), p. 1618-1622.
- m.thompson, j., 2018. *Infrared Spectroscopy*. 1st ed. s.l.:enny Stanford Publishing.
- Yukihiro Ozaki PhD, . W. F. M. P. A. A. C. P., 2006. *Near-Infrared Spectroscopy in Food Science and Technology*. s.l.: John Wiley & Sons.

## Part 2 – List of relevant resources

Fill in this section *after* your project proposal has been approved by your supervisor. Use Harvard referencing (see <https://lrweb.beds.ac.uk/a-guide-to-referencing>). Modify the list below as appropriate. This list is part of Assignment 1 and will be submitted with the Project Proposal.

### 1. Books :

- Ozaki, Y., McClure, W. F., & Davies, A. M. C. (2007) *Near-Infrared Spectroscopy in Food Science and Technology*. Hoboken: Wiley. Available at: <https://onlinelibrary.wiley.com/doi/book/10.1002/0470047704>
- Thompson, J. (2019) *Infrared Spectroscopy*. Boca Raton: CRC Press. Available at: <https://www.taylorfrancis.com/books/mono/10.1201/9781351206037>

### 2. Journal Papers:

- IEEE Sensors Journal (2021) *Non-invasive Glucose Monitoring Using Near-Infrared Spectroscopy: A Review*, 21(5), pp. 3412-3423. Available at: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC9268854/>
- Khan, M., Patel, S., & Kumar, R. (2023) *Machine Learning Approach for Non-Invasive Blood Glucose Prediction, Biomedical Signal Processing and Control*, 78, pp. 103456. Available at: <https://www.nature.com/articles/s41598-022-05570-8>

### 3. Web Sites with relevant information

- ResearchGate (2025) *NIR Spectroscopy Resources*. Available at: <https://www.researchgate.net/>.
- TensorFlow (2025) *Open Source Machine Learning Framework*. Available at: <https://www.tensorflow.org/>.

### 5. Relevant Software:

- MathWorks (2025) *MATLAB - The Language of Technical Computing*. Available at: <https://www.mathworks.com/products/matlab.html>
- Arduino (2025) *Arduino IDE - Open Source Development Environment*. Available at: <https://www.arduino.cc/en/software>

### 6. Relevant Hardware:

- Vishay (2025) *VEML7700 - High Accuracy IR Light Sensor*. Available at: <https://www.vishay.com/optical-sensors/ambient-light-sensors/veml7700/>
- Hamamatsu (2025) *G1738-01 NIR Photodetector for Glucose Monitoring*. Available at: <https://www.hamamatsu.com/us/en/product/type/G1738-01/index.html>

## Ethics Form

|                                                    |                                                                                   |
|----------------------------------------------------|-----------------------------------------------------------------------------------|
| FACULTY OF CREATIVE ARTS, TECHNOLOGIES AND SCIENCE |                                                                                   |
| Form for Research Ethics Projects (Ethics Form)    |                                                                                   |
| Student Name                                       | Pratik Tamang                                                                     |
| Student Number                                     | 2214130                                                                           |
| Degree Pathway                                     | BSc (Hons)Computer Science and Software Engineering                               |
| Supervisor name                                    | Tara G.C                                                                          |
| Supervisor Signature                               |  |
| Title of project                                   | Infrared Sugar Level Tester                                                       |

**SECTION A**  
All the data related to sugar level measurements were collected from freely available sources such as scientific journals and NIR spectroscopy research databases. These data will only be used for research purposes

## SECTION B Check List

Please answer the following questions by circling YES or NO as appropriate:

|                                                                                                                                                                |                                            |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------|
| Does the study involve vulnerable participants or those unable to give informed consent (e.g. children, people with learning disabilities, your own students)? | YES<br><input checked="" type="radio"/> NO |
| Will the study require permission of a gatekeeper for access to participants (e.g. schools, self-help groups, residential homes)?                              | YES<br><input checked="" type="radio"/> NO |
| Will it be necessary for participants to be involved without consent (e.g. covert observation in non-public places)?                                           | YES<br><input checked="" type="radio"/> NO |
| Will the study involve sensitive topics (e.g. obtaining information about sexual activity, substance abuse)?                                                   | YES<br><input checked="" type="radio"/> NO |
| Will blood, tissue samples or any other substances be taken from participants?                                                                                 | YES<br><input checked="" type="radio"/> NO |
| Will the research involve intrusive interventions (e.g. the administration of drugs, hypnosis, physical exercise)?                                             | YES<br><input checked="" type="radio"/> NO |
| Will financial or other inducements be offered to participants (except reasonable expenses or small tokens of appreciation)?                                   | YES<br><input checked="" type="radio"/> NO |
| Will the research investigate any aspect of illegal activity (e.g. drugs, crime, underage alcohol consumption or sexual activity)?                             | YES<br><input checked="" type="radio"/> NO |
| Will participants be stressed beyond what is considered normal for them?                                                                                       | YES<br><input checked="" type="radio"/> NO |
| Will the study involve participants from the NHS (patients or staff) or will data be obtained from NHS premises?                                               | YES<br><input checked="" type="radio"/> NO |

If the answer to any of the questions above is "Yes", or if there are any other significant ethical issues, then further ethical consideration is required. Please document carefully how these issues will be addressed:

Signed (student):   
Date: 29<sup>th</sup>, Feb 2025

Countersigned (Supervisor):   
Date: 27<sup>th</sup>, Feb, 2025

## Weekly Reports from UG




UNIVERSITY OF BEDFORDSHIRE  
 DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
 FINAL YEAR UG PROJECT  
 WEEKLY PROGRESS REPORT FORM

|                                                         |                                    |
|---------------------------------------------------------|------------------------------------|
| Student's Name: <u>Pratik Tamang</u>                    | Supervisor's Name: <u>Pawanice</u> |
| Project Title: <u>Infrared Blood Sugar level Tester</u> |                                    |
| Week No: _____                                          | Report No: <u>07</u>               |

|                                                 |                                                                                                                                                                |
|-------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"> <li>- Creating Circuit Diagram</li> <li>- Created Api for integration</li> <li>- Folder setup and started coding</li> </ul> |
| Plan for next week                              | <ul style="list-style-type: none"> <li>- continue coding</li> <li>- Assemble device</li> </ul>                                                                 |
| Supervisor Comment                              | *Finalize the project and work on the Report.                                                                                                                  |

Student's Signature [Signature]  
 Date 2/1/11

Supervisor's Signature [Signature]  
 Date 2/1/11

When signed this form must be scanned and submitted via the relevant link on BREO



UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                   |                             |
|---------------------------------------------------|-----------------------------|
| Student's Name: Pratik Tamang                     | Supervisor's Name: Pawan KC |
| Project Title : Infrared Blood Sugar Level Tester |                             |
| Week No: 5                                        | Report No: 3                |

|                                                 |                                                                                                                                                                                                  |
|-------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"><li>- Started Training Plsr Model</li><li>- Validated accuracy of collected data.</li></ul>                                                                    |
| Plan for next week                              | <ul style="list-style-type: none"><li>- Continue refining model integration.</li><li>- Improve data handling and processing.</li><li>- Plan API for model to communicate with the app.</li></ul> |
| Supervisor Comment                              | <ul style="list-style-type: none"><li>- start the development of the artifact.</li><li>- Progress looks good.</li></ul>                                                                          |

Student's Signature .....

Supervisor's Signature .....

Date 07/11 .....

Date 07/11 .....


When signed this form must be scanned and submitted via the relevant link on BREO

UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                   |                             |
|---------------------------------------------------|-----------------------------|
| Student's Name: Pratik Tamang                     | Supervisor's Name: Pawan KC |
| Project Title : Infrared Blood Sugar Level Tester |                             |
| Week No: 6                                        | Report No: 4                |

|                                                 |                                                                                                                                                                                        |
|-------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"><li>- Finalizing the model and improving accuracy</li><li>- Improved data handling and sensor reading.</li><li>- Start to code the aurdino</li></ul> |
| Plan for next week                              | <ul style="list-style-type: none"><li>- Continue coding Arduino for device functionality.</li><li>- Test integration with PLSR model in the aurdino.</li></ul>                         |
| Supervisor Comment                              | - Finalize the model.                                                                                                                                                                  |

Student's Signature  ..... Supervisor's Signature  .....

Date 07/11 ..... Date 07/11  .....

When signed this form must be scanned and submitted via the relevant link on BREO

UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                   |                             |
|---------------------------------------------------|-----------------------------|
| Student's Name: Pratik Tamang                     | Supervisor's Name: Pawan KC |
| Project Title : Infrared Blood Sugar Level Tester |                             |
| Week No: 4                                        | Report No: 2                |

|                                                 |                                                                                                                                                                           |
|-------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"><li>- Started training the pls model.</li><li>- Wrote basic Arduino codes for hardware testing (not related to main project).</li></ul> |
| Plan for next week                              | <ul style="list-style-type: none"><li>- Finalize and fine tune thePLSR model.</li><li>- Folder setup for the project.</li></ul>                                           |
| Supervisor Comment                              | - check the hard ware necessary for the development.                                                                                                                      |

Student's Signature .....  .....  
Date 07/11 .....

Supervisor's Signature .....  .....  
Date 07/11 ..... 

UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                   |                             |
|---------------------------------------------------|-----------------------------|
| Student's Name: Pratik Tamang                     | Supervisor's Name: Pawan KC |
| Project Title : Infrared Blood Sugar Level Tester |                             |
| Week No: 3                                        | Report No: 1                |

|                                                 |                                                                                                                                                                                                                                                    |
|-------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"><li>- Collected dataset for training and fine-tuning the PLSR model.</li><li>- Ensured dataset was clean and comprehensive.</li><li>- Spent 1 week on data preparation and understanding requirements.</li></ul> |
| Plan for next week                              | <ul style="list-style-type: none"><li>- Complete any remaining dataset cleaning.</li><li>- Begin training the plsr model.</li></ul>                                                                                                                |
| Supervisor Comment                              | <ul style="list-style-type: none"><li>- Check the clarity of the dataset.</li><li>- Fine tune the model</li></ul>                                                                                                                                  |

Student's Signature  .....  
Date 07/11 .....

Supervisor's Signature  .....  
Date 07/11 .....



When signed this form must be scanned and submitted via the relevant link on BREO



UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                   |                             |
|---------------------------------------------------|-----------------------------|
| Student's Name: Pratik Tamang                     | Supervisor's Name: Pawan KC |
| Project Title : Infrared Blood Sugar Level Tester |                             |
| Week No: 7                                        | Report No: 5                |

|                                                 |                                                                                                                                                   |
|-------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"> <li>- Tuned the model</li> <li>- Started the Aurdino Code</li> <li>- Started Self Reflective Report</li> </ul> |
| Plan for next week                              | <ul style="list-style-type: none"> <li>- Continue coding Arduino</li> <li>- Start assembling the Iot Devices</li> </ul>                           |
| Supervisor Comment                              | <p>* Complete the reflective report</p> <p>* Tune the model for further accuracy</p>                                                              |

Student's Signature .....

Date 11/13 .....

Supervisor's Signature .....

Date 11/13 .....

When signed this form must be scanned and submitted via the relevant link on BREO



UNIVERSITY OF BEDFORDSHIRE  
DEPARTMENT OF COMPUTER SCIENCE AND TECHNOLOGY  
FINAL YEAR UG PROJECT  
WEEKLY PROGRESS REPORT FORM

|                                                         |                                   |
|---------------------------------------------------------|-----------------------------------|
| Student's Name: <u>Pratik Tamang</u>                    | Supervisor's Name: <u>Pawanke</u> |
| Project Title: <u>Infrared Blood Sugar level tester</u> |                                   |
| Week No: <u>08</u>                                      | Report No: <u>6</u>               |

|                                                 |                                                                                                                       |
|-------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------|
| Summary of progress<br>(including any problems) | <ul style="list-style-type: none"> <li>- Completed Self Reflective</li> <li>- Gathered all the IOT Devices</li> </ul> |
| Plan for next week                              | <ul style="list-style-type: none"> <li>- Code the arduino</li> <li>- Create Circuit Diagram of Device.</li> </ul>     |
| Supervisor Comment                              | * Wrap up the project                                                                                                 |

Student's Signature [Signature]


Date 11/11


Supervisor's Signature [Signature]

Date 14/11/2025

When signed this form must be scanned and submitted via the relevant link on BREO

## Testing Results





$$\hat{y}_{\text{glucose}} = \sum_{i=1}^p (\text{Coefficient}_i \times X_i) + \text{Intercept}$$

## Model Training

```
// --- PLSR model parameters embedded in firmware ---
const int nFeatures = 1;
float pls_coeff[nFeatures] = {1.00141187}; // The extracted beta (beta)
float pls_intercept = 69.97973763;        // The extracted intercept (beta_0)

// Example calculation snippet (simplified for illustration):
// float predicted_glucose = (scaled_ir_input[0] * pls_coeff[0]) + pls_intercept;
```

```
# Optional: split into train/test to check performance
X_train, X_test, y_train, y_test = train_test_split(X, y, test_size=0.2, random_state=42)
```

```
pls = PLSRegression(n_components=1)
pls.fit(X_train, y_train)
```

PLSRegression ⓘ ?  
PLSRegression(n\_components=1)

```
y_pred = pls.predict(X_test)
mse = mean_squared_error(y_test, y_pred)
rmse = np.sqrt(mse)
r2 = r2_score(y_test, y_pred)
print(f"Test RMSE: {rmse:.2f}")
print(f"Test R^2: {r2:.3f}")
```


Test RMSE: 4.93

Test R^2: 0.746

```
coef = pls.coef_.flatten()
# Calculate the intercept using the mean of the training data
# For a simple linear relationship y = X * coef + intercept,
# the intercept can be derived as: intercept = mean(y_train) - mean(X_train) * coef
intercept = y_train.mean() - np.dot(X_train.mean(axis=0), coef)

print("\n=== PLSR Model Parameters ===")
print("Coefficient(s):", coef)
print("Intercept:", intercept)
```

```
=== PLSR Model Parameters ===
Coefficient(s): [1.00141187]
Intercept: 69.97973762723899
```



## Testing Code Snippets

```
#include <Arduino.h>
#include <LiquidCrystal.h>

LiquidCrystal lcd(12, 14, 33, 27, 26, 25);

Tabnine | Edit | Test | Explain | Document
void setup() {
  lcd.begin(16, 2);
  lcd.print("Hello, World!");
}
int dummyPower = 32;

Tabnine | Edit | Test | Explain | Document
void loop() {
  lcd.setCursor(0, 1);
  lcd.print(millis() / 1000);
}
```

```

void loop() {
  int rawValue = analogRead(tcrtPin);
  String distanceLevel;
  String detection;

  if (rawValue < 100) {
    distanceLevel = "VERY CLOSE";
    detection = "OBJECT DETECTED";
  }
  else if (rawValue < 300) {
    distanceLevel = "CLOSE";
    detection = "OBJECT DETECTED";
  }
  else if (rawValue < 600) {
    distanceLevel = "MEDIUM";
    detection = "MAYBE";
  }
  else {
    distanceLevel = "FAR / NO OBJECT";
    detection = "NO OBJECT";
  }
  Serial.print("Raw: ");
  Serial.print(rawValue);
  Serial.print(" | Level: ");
  Serial.print(distanceLevel);
  Serial.print(" | ");
  Serial.println(detection);
}

```

```

#ifndef GLUCOSE_FUNCTIONS_H
#define GLUCOSE_FUNCTIONS_H

float getAbsorptionScaled(float raw, float baseline);
float getSmoothedAbsorption(float newValue);
float getExperimentalGlucose(float absorption_scaled);
String getGlucoseStatus(float glucose);

#endif

```

```
#include "glucose_functions.h"

#define AVG_SAMPLES 5
static float lastReadings[AVG_SAMPLES] = {0};
static int readingIndex = 0;

float getAbsorptionScaled(float raw, float baseline) {
    if(baseline <= 0) return 0;
    return ((baseline - raw) / baseline) * 100.0;
}

float getSmoothedAbsorption(float newValue) {
    lastReadings[readingIndex] = newValue;
    readingIndex = (readingIndex + 1) % AVG_SAMPLES;
    float sum = 0;
    for(int i=0; i<AVG_SAMPLES; i++) sum += lastReadings[i];
    return sum / AVG_SAMPLES;
}

float getExperimentalGlucose(float absorption_scaled) {
    const float glucoseMin = 70;
    const float glucoseMax = 140;
    float glucose = glucoseMin + (absorption_scaled / 10.0) * (glucoseMax - glucoseMin);
    if(glucose < glucoseMin) glucose = glucoseMin;
    if(glucose > glucoseMax) glucose = glucoseMax;
    return glucose;
}

String getGlucoseStatus(float glucose) {
    if(glucose < 70) return "LOW!";
    else if(glucose > 140) return "HIGH!";
    else return "NORMAL";
}
```

```
#include <Arduino.h>
#include <../pio/libdeps/esp32dev/AUnit/src/AUnit.h>
#include "../src/glucose_functions.h" // Make sure path is correct

test(absorption_scaled_test) {
    assertEquals(getAbsorptionScaled( raw: 50, baseline: 100), 50.0f);
}

test(smoothed_absorption_test) {
    // reset smoothing buffer if necessary
    for(int i=0;i<5;i++) getSmoothedAbsorption( newValue: 0);
    assertEquals(getSmoothedAbsorption( newValue: 10), 2.0f);
}

test(experimental_glucose_test) {
    assertEquals(getExperimentalGlucose( absorption_scaled: 5), 105.0f);
}

test(glucose_status_test) {
    assertEquals(getGlucoseStatus(60), "LOW!");
    assertEquals(getGlucoseStatus(120), "NORMAL");
    assertEquals(getGlucoseStatus(150), "HIGH!");
}
```

## Testing Results and Formula

```
=====
TCRT5000 Sensor Test
=====
Move your hand or white/black paper in front of the sensor
Format: RawValue | DistanceLevel | Detection

Raw: 812 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 798 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 315 | Level: CLOSE | OBJECT DETECTED
Raw: 158 | Level: VERY CLOSE | OBJECT DETECTED
Raw: 89 | Level: VERY CLOSE | OBJECT DETECTED
Raw: 712 | Level: FAR / NO OBJECT | NO OBJECT
Raw: 620 | Level: MEDIUM | MAYBE
Raw: 421 | Level: CLOSE | OBJECT DETECTED
Raw: 798 | Level: FAR / NO OBJECT | NO OBJECT
```

```
*** ===== First 5 Rows =====
      raw  baseline  reflection_scaled  glucose_mgdl  reflection_norm  \
0  2660      2126          62.56      136.83      0.919613
1  1843      2459          37.47      112.37      0.082722
2  2280      1860          61.29      137.88      0.877252
3  2365      2294          51.55      121.98      0.552368
4  1701      2130          39.93      114.51      0.164777

      absorption_synthetic
0      0.080387
1      0.917278
2      0.122748
3      0.447632
4      0.835223
```

```
// --- PLSR parameters (absorption-only) ---
const float PLSR_INTERCEPT = 120.17;
const float PLSR_COEF = -30.0223;
```

$$y = \text{intercept} + \sum (\text{coef}_i \times x_i)$$

$$\text{raw\_model} = \text{raw} \times \frac{\text{mean\_raw}}{110}$$

$$\text{baseline\_model} = \text{baseline} \times \frac{\text{mean\_baseline}}{110}$$

$$x_{\text{std}} = \frac{x - \text{mean}_x}{\text{std}_x}$$

So for each variable:

$$\text{raw\_s} = \frac{\text{raw\_model} - \text{mean\_raw}}{\text{std\_raw}}$$

$$\text{baseline\_s} = \frac{\text{baseline\_model} - \text{mean\_baseline}}{\text{std\_baseline}}$$

$$\text{reflection\_scaled\_s} = \frac{\text{reflection\_scaled} - \text{mean\_reflection\_scaled}}{\text{std\_reflection\_scaled}}$$

$$\text{reflection\_norm\_s} = \frac{\text{reflection\_norm} - \text{mean\_reflection\_norm}}{\text{std\_reflection\_norm}}$$

$$\text{absorption\_s} = \frac{\text{absorption} - \text{mean\_absorption}}{\text{std\_absorption}}$$

$$\begin{aligned} \text{Glucose} = & \text{pls\_intercept} + \text{coef\_raw} \cdot \frac{\text{raw} \cdot \text{mean\_raw}/110 - \text{mean\_raw}}{\text{std\_raw}} \\ & + \text{coef\_baseline} \cdot \frac{\text{baseline} \cdot \text{mean\_baseline}/110 - \text{mean\_baseline}}{\text{std\_baseline}} \\ & + \text{coef\_reflection\_scaled} \cdot \frac{\frac{\text{raw} - \text{baseline}}{\text{baseline}} \cdot 100 - \text{mean\_reflection\_scaled}}{\text{std\_reflection\_scaled}} \\ & + \text{coef\_reflection\_norm} \cdot \frac{\frac{\text{raw} - \text{baseline}}{\text{baseline}} - \text{mean\_reflection\_norm}}{\text{std\_reflection\_norm}} \\ & + \text{coef\_absorption} \cdot \frac{\text{constrain} \left( \frac{\text{baseline} - \text{raw}}{\text{baseline}} \cdot 100, -5, 25 \right) - \text{mean\_absorption}}{\text{std\_absorption}} \end{aligned}$$

## Project Poster: