

INSTALLATION REPORT

WORK ORDER DETAILS

Ref Work Order No.	ULPL-F/CG/SVP/AUTOCLAVE/26270346A
Date	29/08/2026
Subject	SVP Autoclave PLC Upgradation

PERSONNEL

No.	Name	Department	Sign.
1	Mr Chandrashekar	Production	
2	Mr Chandrashekhar	Engineering	
3		QA	

ACTIVITY PERFORMED

- 1) Supplied and Installed the Control panel with PLC ,HMI and necessary components
- 2) Developed the PLC HMI Program as per Requirements
- 3) Verified the Vacuum Leak Test Cycle
- 4) Verified the Standard cycle
- 5) Verified the HPHV Cycle
- 6) Verified the Bowie & Dick Test cycle

For SENSOTECH

Bapu Patil

For ULTRA LABORATORIES PRIVATE LIMITED

1. Mr Chandrashekhar [Production]
2. Mr Chandrashekhar [Engineering]
3. [QA]