Test Clean Tables

Contents

```
threshold_df <- data.frame(
  Decision_Context = c(
    "Head-to-Head Parity (Risk > 6%)",
    "Head-to-Head Parity (Risk > 4%)",
    "Head-to-Head Parity (Observed)",
    "Practice Transition (S-ICD > 6% vs. TV-ICD Obs)"
  ),
  Break_Even_25k = c("£12,761 (-26.8%)", "£12,753 (-26.8%)", "£12,620 (-27.6%)", "£22,230 (+27.5%)"),
  Break_Even_35k = c("£12,773 (-26.7%)", "£12,765 (-26.8%)", "£12,631 (-27.5%)", "£25,469 (+46.1%)"),
  Current_ICER = c("Dominated", "Dominated", "Dominated", "£10,198 / QALY"),
  Economic_Mechanism = c(
    "Must be discounted ~£2,550 below TV-ICD list price (£15,317) to offset 8-yr vs 10-yr battery repla
    "Must offset higher 30-year replacement frequency across 1,103 implanted patients.",
    "Must offset replacement burden under unguided historic selection.",
    "Risk > 6% clinical efficiency (avoiding low-risk implants) cross-subsidizes S-ICD premium."
  )
)
colnames(threshold_df) <- c("Decision Context", "Break-Even (£25k)", "Break-Even (£35k)", "Current ICER

format_table(
  threshold_df,
  caption = "Price threshold analysis for S-ICD hardware and replacement costs (Current S-ICD list pric
  font_size = 8.5,
  col_widths = list("5" = "6cm")
)
```

Table 1: Price threshold analysis for S-ICD hardware and replacement costs (Current S-ICD list price: £17,434).

Decision Context	Break-Even (£25k)	Break-Even (£35k)	Current ICER	Economic Rationale
Head-to-Head Parity (Risk > 6%)	£12,761 (-26.8%)	£12,773 (-26.7%)	Dominated	Must be discounted list price (£15,317) battery replacement
Head-to-Head Parity (Risk > 4%)	£12,753 (-26.8%)	£12,765 (-26.8%)	Dominated	Must offset higher 3 frequency across 1,1
Head-to-Head Parity (Observed)	£12,620 (-27.6%)	£12,631 (-27.5%)	Dominated	Must offset replacen unguided historic se
Practice Transition (S-ICD > 6% vs. TV-ICD Obs)	£22,230 (+27.5%)	£25,469 (+46.1%)	£10,198 / QALY	Risk > 6% clinical e low-risk implants) c premium.