SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 & ANSI/NCSL Z540-1-1994

ATC INC.
4037 Guion Lane
Indianapolis, IN 46268
Hemi Sagi Phone: 317 328 8492

CALIBRATION

Valid To: March 31, 2017 Certificate Number: 2197.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations[1]:

I. Dimensional

Parameter/Equipment	Range	CMC[2] (±)	Comments
Optical Gauging	10 µm to 1.5 mm	1.0 µm	Microscope

II. Fluid Quantities

Parameter/Equipment	Range	CMC[2] (±)	Comments
Gas Flow	(15 to 5000) gm/min (0.03 to 120) gm/min (1 to 40) mg/min (0.1 to 3000) µg/min	0.33 % of reading 0.31 % of reading 0.39 % of reading 0.67 % of reading	Gas Cal-5000L primary gas flow standard PVTt

III. Mechanical

Parameter/Equipment	Range	CMC[2] (±)	Comments
Pressure	Up to 5000 psia 4 in H_20 to 10 psi	0.013 % Full Scale 0.01 % Full Scale	Pressure standards
Vacuum	Up to 20 Torr	0.25 % Full Scale	Vacuum standard

IV. Thermodynamics

Parameter/Equipment	Range	CMC[2] (±)	Comments
Temperature	0 ºC to 100 ºC	0.2 ºC	Comparison with thermistor probe

---

[1] This laboratory offers commercial calibration service.

[2] Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

American Association for Laboratory Accreditation

Accredited Laboratory

A2LA has accredited

ATC INC.

Indianapolis, IN

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 19th day of March 2015.

President & CEO
For the Accreditation Council
Certificate Number 2197.01
Valid to March 31, 2017

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.