SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

AER SAMPLING SDN BHD
49-01 Jalan Sena 2
Masai Malaysia 81750
Ed (Employee ID: 15) Phone: +60 7 380 5344

CALIBRATION

Valid To: February 28, 2017 Certificate Number: 3704.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations[1]:

I. Mechanical

Parameter/Equipment	Range	CMC[2, 3] (±)	Comments[4]
Dry Gas Meter	(0.0050 to 0.0400) m^3/min	1.6 %	USEPA method 5, section 10.3 (Y) @ 20 °C, 760 mmHg
Meter Orifice –			
Orifice Tube/Meter, Metering System Orifice	0.0212 m^3/min ± 0.0071 m^3/min (0.75 ft^3/min ± 0.25 ft^3/min)	2.4 %	USEPA method 5, section 9.2.1 & 10.3 (ΔH@) @ 20 °C, 760 mmHg (or 68 °F, 29.92 inHg)
Critical/Sonic/Choke Orifice	(0.0050 to 0.0400) m^3/min	2.2 %	USEPA method 5, section 16.2.2 (K') @ 20 °C, 760 mmHg

II. Thermodynamic

Parameter/Equipment	Range	CMC[2] (±)	Comments[4]
Temperature Sensor – Thermocouple Potentiometer System			
Type K	(-20 to 900) °C	0.6 °C	USEPA method 5, section 10.5 & USEPA method 2, section 10.3.1.

[1] This laboratory does not normally offer commercial calibration service.

[2] Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

[3] In the statement of CMC, the value is defined as the percentage of reading, unless otherwise noted.

[4] USEPA Method 2 (or 5): Title 40, Code of Federal Regulations (United States of America), Part 60, Appendix A, Test Method 2 (or 5).

Accredited Laboratory

A2LA has accredited

AER SAMPLING SDN BHD

Masai, MALAYSIA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 1st day of January, 2015.

President & CEO
For the Accreditation Council
Certificate Number 3704.01
Valid to February 28, 2017

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.