SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

BALANZAS Y EQUIPOS S.R.L (BALECA)
Jose Amado Soler #9 Ens. Serralles
Santo Domingo, Dominican Republic
Angelo Vincenzi Phone: 809 563 1735
Email: angelo@baleca.com

CALIBRATION

Valid To: February 28, 2018 Certificate Number: 3655.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations[1]:

I. Mechanical

Parameter/Equipment	Range	CMC[2] (±)	Comments
Scales & Balances[3]			ASTM E898
Class I[4]	(0 to 500) g (500 to 1000) g	0.40 mg 1.2 mg	Mass standards OIML R111 Class E2
Class II[4]	(0 to 100) g (0 to 200) g (0 to 500) g (0 to 1 000) g (0 to 10 000) g (0 to 32 000) g (0 to 50 000) g	0.98 mg 1.9 mg 4.2 mg 8.8 mg 0.088 g 0.40 g 0.42 g	OIML R111 Class F1
Class III[4]	(0 to 5) lb (0 to 10) lb (0 to 20) lb	0.00040 lb 0.00086 lb 0.0017 lb	ASTM E617 Class 4
	(0 to 50) lb (0 to 100) lb (0 to 200) lb (0 to 500) lb (0 to 1 000) lb (0 to 5 000) lb	0.0047 lb 0.01 lb 0.02 lb 0.045 lb 0.094 lb 0.42 lb	NIST 105-1 Class F
	(0 to 10 000) lb (0 to 20 000) lb (0 to 40 000) lb	0.94 lb 1.9 lb 4.2 lb	NIST 547 Class T

Parameter/Equipment	Range	CMC[2] (±)	Comments
Scales & Balances[3] (cont) -			
Class IIIL[5]	(0 to 50 000) lb (0 to 100 000) lb (0 to 200 000) lb	4.1 lb 8.5 lb 17 lb	NIST 547 Class T
Class III[4]	(0 to 5) kg (0 to 10) kg (0 to 25) kg (0 to 50) kg	0.00040 kg 0.00084 kg 0.0040 kg 0.0040 kg	OIML R111 Class F1
	(0 to 150) kg (0 to 250) kg (0 to 500) kg (0 to 2 500) kg (0 to 5 000) kg	0.019 kg 0.041 kg 0.044 kg 0.42 kg 0.42 kg	NIST 105-1 Class F
	(0 to 10 000) kg (0 to 20 000) kg	0.86 kg 1.7 kg	NIST 547 Class T
Class IIIL[5]	(0 to 25 000) kg (0 to 50 000) kg (0 to 100 000) kg	4.1 kg 4.1 kg 8.3 kg	NIST 547 Class T
Class IIII[4]	(0 to 50) kg (0 to 250) kg (0 to 2 500) kg (0 to 5 000) kg	0.04 kg 0.4 kg 4.0 kg 4.0 kg	OIML R111 Class F1 NIST 105-1 Class F
Mass	(1 to 50) g	0.10 mg	OIML R111 Class F2
	(100 to 200) g	0.12 mg	OIML R111 Class F1, C2 ASTM E617-97
	500 g to 1 kg (20 to 25) kg (10 to 50) lb	0.55 mg 98 mg 98 mg	OIML R111 Class F2 F NIST 105-1 F NIST 105-1

[1] This laboratory offers commercial calibration service and field calibration service.

[2] Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

[3] Field calibration service is available for this calibration and this laboratory meets A2LA *R104 – General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

[4] Classes identified as per OIML R 76-1

[5] Classes identified as per Handbook 44-2012

Accredited Laboratory

A2LA has accredited

BALANZAS Y EQUIPOS S.R.L (BALECA)

Santo Domingo, DOMINICAN REPUBLIC

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and *R205 – Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 19th day of February 2016.

President and CEO
For the Accreditation Council
Certificate Number 3655.01
Valid to February 28, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.