



Calibration Certificate

SpanDust™ 1.28



Information contained herein is classified as EAR99 under the U.S. Export Administration Regulations. Export, reexport or diversion contrary to U.S. law is prohibited.



Test conditions

Humidity:	39 %
Air pressure:	1009 mbar
Ambient air temperature:	29 ° C

NIST Reference Material

Manufacturer:	Thermo Scientific
Lot number:	193291
Certified Mean Diameter:	0.994 μm
Standard Deviation:	0.01 μm
Expanded Uncertainty:	0.015 μm

SpanDust™ 1.28

Lot. Number:	100128-050-026
Valid until:	16.09.2021

Test results

Certified Mean Diameter:	1.287 $\mu\text{m} \pm 0.0009 \mu\text{m}$
Standard Deviation:	0.0004 μm
Expanded Uncertainty:	0.0180 μm

with coverage factor K=2 providing a level of confidence of approximately 95%.