



PORTABLE MEASURING SYSTEMS

The **Gasmeter™** DX 4000 Portable Emissions Measuring System incorporates all the benefits of an FTIR CEM system into a portable package. By combining hot extractive sampling and a high-performance FTIR gas analyser components such as HF, HCl, NH₃, N₂O, formaldehyde and others can be measured continuously in harsh conditions. **Gasmeter™** analysers are developed and manufactured by Gasmeter Technologies, the leading company in FTIR emissions monitoring applications.





DX4000 FTIR system for gas analysis

- ✚ Rugged, yet lightweight design
- ✚ Easy and quick on-site set-up
- ✚ Measurement ranges from ppm to %
- ✚ Different application areas
- ✚ Typical components measured: H_2O , CO , CO_2 , SO_2 , NO , NO_2 , N_2O , HCl , HF , NH_3 , methane, ethane, ethene, hexane, propane, formaldehyde
- ✚ Additional components can be easily added (hydrocarbons, organic compounds, inorganic compounds)



APPLICATION AREAS

- ✚ Continuous Emission Monitoring
- ✚ Stack Testing
- ✚ Process Monitoring
- ✚ Quality Control
- ✚ Engine Exhaust Gas Monitoring
- ✚ Workplace Air Quality Monitoring
- ✚ Combustion Research
- ✚ Emergency Rescue Services



Portable sampling unit



- ✚ For extracting hot and wet gas emissions and removing particles
- ✚ Fine particle filter heated to 180 °C
- ✚ Sample pump heated to 180 °C
- ✚ Automatic zero gas valve for test and calibration on site
- ✚ Sample gas flow ~ 4 l/min
- ✚ Temperature controllers for two heated lines up to 9 m + 1 m
- ✚ Automatic operation controlled by computer, including automatic zero calibration and zero gas flushing in error situations
- ✚ Manual operation possible
- ✚ Optional O₂ measurement with additional sensor




Calibrator

- ✚ Generate water vapour up to 50 vol. %
- ✚ Generate wet calibration gases
- ✚ Portable
- ✚ Easy to use
- ✚ One of the most important tools in generating new applications
- ✚ The injected liquid is vaporized rapidly in the stainless steel injection chamber, which can be heated up to 180 °C
- ✚ Special need for service - water calibrations



Portable Gas sampling probe with heated sample line

- ✚ For portable applications
- ✚ For continuous gas sampling in processes with dust and humidity
- ✚ Removing aerosols and dust from gas sample
- ✚ Ready to use
- ✚ Compact
- ✚ Quick set up
- ✚ Easy operation
- ✚ Tool-less connectors
- ✚ No external parts needed



PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

GASMET MULTI-GAS ANALYSER
CX4000 with Oxitec 500 E SME 5 Oxygen Analyser,
DX4000 & EN4000 Multi-gas Analyser

manufactured by:

Gasmet Technologies OY

Pullitie 8 A 1
FIN-00880 Helsinki
Finland

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

MCERTS Performance Standards for Continuous Emission
Monitoring Systems, Version 3.4 dated July 2012,
EN15267-3:2007,
& QAL 1 as defined in EN 14181: 2004

Certification Ranges :

CO	0 to 75 mg/m ³	0 to 300mg/m ³	0 to 1500mg/m ³
NO	0 to 200 mg/m ³	0 to 600mg/m ³	0 to 2000mg/m ³
SO ₂	0 to 75 mg/m ³	0 to 300mg/m ³	0 to 1500mg/m ³
NO ₂	0 to 200 mg/m ³	0 to 500 mg/m ³	
N ₂ O	0 to 100 mg/m ³	0 to 500mg/m ³	
HCl	0 to 15 mg/m ³	0 to 90mg/m ³	
NH ₃	0 to 15 mg/m ³	0 to 50mg/m ³	
H ₂ O	0 to 30 % vol.	0 to 40 % vol.	
HF	0 to 3 mg/m ³	0 to 10 mg/m ³	
CO ₂	0 to 25 % vol.		
O ₂	0 to 25 % vol.**		

* Certification range only applies to system with CX4001 FTIR analyser option.

** Certification range only applies to CX4000 analyser with Oxitec 500 E SME 5 oxygen analyser option.

Project No: 674/0062 & 674/0334
Certificate No: Sira MC 030014/07
Initial Certification: 04 July 2003
This Certificate Issued: 04 July 2013
Renewal Date: 03 July 2018

Technical Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK DA1 4AL
Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501

This certificate may only be reproduced in its entirety and without change
Registered Office: Rake Lane, Eccleston, Chester, UK CH4 3UN
To authenticate the validity of this certificate please visit www.siracertification.com/mcerts

MCERTS Performance Standards for Continuous Emission Monitoring Systems, Version 3.4 dated July 2012, EN15267-3:2007, & QAL 1 as defined in EN 14181: 2004

Certification Ranges :

CO	0 to 75 mg/m ³	0 to 300mg/m ³	0 to 1500mg/m ³
NO	0 to 200 mg/m ³	0 to 600mg/m ³	0 to 2000mg/m ³
SO ₂	0 to 75 mg/m ³	0 to 300mg/m ³	0 to 1500mg/m ³
NO ₂	0 to 200 mg/m ³	0 to 500 mg/m ³	
N ₂ O	0 to 100 mg/m ³	0 to 500mg/m ³	
HCl	0 to 15 mg/m ³	0 to 90mg/m ³	
NH ₃	0 to 15 mg/m ³	0 to 50mg/m ³	
H ₂ O	0 to 30 % vol.	0 to 40 % vol.	
HF	0 to 3 mg/m ³	0 to 10 mg/m ³	
CO ₂	0 to 25 % vol.		
O ₂	0 to 25 % vol.**		