



1502 East Warner Avenue, Suite B  
Santa Ana, CA 92705  
USA  
[sales@delta-phase.us](mailto:sales@delta-phase.us)  
+1-714-866-8070

July 03, 2024

Letter confirmation,

By this letter, we confirm the accuracy and response time of AM7 and SA9 (5mm) sensors as below:

|               | COD     | TSS      | Color   | Ammonium | pH        | Temp    |
|---------------|---------|----------|---------|----------|-----------|---------|
| Accuracy      | ± 2%    | ± 2%     | ± 2%    | ± 3%     | ± 0.02 pH | ± 3%    |
| Response time | ≤ 1 min | ≤ 10 sec | ≤ 1 min | ≤ 1 min  | ≤ 5 sec   | ≤ 5 sec |

Sincerely,

  
Peter A. DeBarber  
Director of Business Development

