



Online Ammonia Nitrogen Sensor

Modbus Protocol

MODBUS RTU transmission mode

32-bit floating point numbers are ABCD first, DCBA then parsed.

Encoding system	8-bit binary Each 8-bit byte in the protocol contains two 4-bit hexadecimal characters (0-9, A-F)
Bit per byte	1 start bit 8 data bits, the least significant bit is sent first No parity check bit 1 stop bit
Baud rate	9600bps
Set slave ID 01 changed to 02	01 10 08 00 00 01 02 02 00 2E F0 01 10 08 00 00 01 03 A9
Get slave ID 01	FF 03 08 00 00 01 93 B4 FF 03 02 01 00 90 00
Get pH value 6.84	01 03 20 00 00 04 4F C9 01 03 08 B8 CB 02 43 4B BD DA 40 7C B4
Get NH4-N value 10 mg/L	01 03 13 00 00 06 C1 4C 01 03 0C 00 00 20 41 00 00 00 BE 00 00 80 3F A7 82
Get NH4 user calibration parameters K= 1 ,B= 0	01 03 05 00 00 04 44 C5 01 03 08 00 00 80 3F 00 00 00 00 9E 12
Set NH4 user calibration parameters K= 1 ,B= 0	01 10 05 00 00 04 08 00 00 80 3F 00 00 00 00 B1 BA 01 10 05 00 00 04 C1 06