



Fluorescence Dissolved Oxygen Sensor Modbus protocol

MODBUS RTU Transmission Mode:

32 bit floating point number analysis method (adjust ABCD first , then analyze DCBA)

Encoding system	8-bit binary Each 8-bit byte in the protocol contains two 4-bit hexadecimal characters (0-9, A-F)
Bit per byte	1 start bit 8 data bits, the least significant bit is sent first No parity check bit 1 stop bit
Baud rate	9600bps
Temperature: 17.5℃ Percentage: 95.8% DO Value: 8.72mg/L	SEND: 01 03 20 00 00 06 C4 70 GET: 01 03 0C 00 00 8D 41 83 5B 75 3F E8 88 0B 41 F6 6B
Get user calibration parameters (K=1,B=0)	01 03 22 00 00 04 4E 42 01 03 08 00 00 80 3F 00 00 00 00 91 5
Set user calibration parameters (K=1,B=0)	SEND: 01 10 22 00 00 04 08 00 00 80 3F 00 00 00 00 86 9C GET: 01 10 22 00 00 04 CB 81
Get slave ID 01	FF 03 08 00 00 01 93 B4 FF 03 02 01 00 90 00 Get address: 01
Set slave ID	Original address 01 after setting address 02 01 10 08 00 00 01 02 02 00 2E F0 01 10 08 00 00 01 03 A9