



Doosan VentureDigm #827, Heungandaero 415, Dongan-gu, Anyang-si, Gyeonggi-do, Korea (14059)
Tel) 82-31-478-3420~8 Fax) 82-31-478-3449 Homepage) www.korbi.com

Specification

· Online TN Analyzer ·

Model	HATN-3000
<Measurement specification>	
Measurement item	Total Nitrogen (TN)
Measurement method	UV Absorption photometry
Measurement range	0~100mg/L (Adjustable)
Wavelength	220nm (275nm for compensation)
Measurement interval	30 minutes (Adjustable)
Detection limit	1mg/L F.S.
Accuracy	±2% F.S.
Reproducibility	±2% F.S.
Linearity	±3% F.S.
Zero/Span drift	±2% F.S.
Oxidation method	UV persulfate oxidation
Calibration	Auto/manual calibration by standard solution
Sample/Reagent measuring method	Step motor pulse control
Sample/Reagent distribution method	8-port distribution control method (Fixed quantity injection with multi-channel valve system)
<Instrument specification>	
Operation system	Windows 7 (CPU process) or more
Output	DC 4~20mA, RS-232C
Status display	Normal operation, on calibration, error
Display	Color LED 10.4" (Touch screen)
Power	AC 220V, 50/60Hz
Power consumption	Average 70W
Voltage regulation	3% or less
Data storage capacity	HDD 100G or more
Data output	USB terminal
Enclosure	Steel, 1.5T (Except paint)
Size	645(W) × 439(D) × 1600(H) mm
<Measurement environment>	
Operating temp.	5~50℃
Operating humidity	RH 95% or less

※ The specification may be subject to change according to the manufacturer's updates without prior notice.