

Inspection certificate

DIN EN 10204/3.1

ZERTIFIZIERT
DIN ISO 9001
NR 70100 222
CERTIFIED



Compact Weather Station

Model Type	WS500	
Serial Number	258.0713.0813.035	

This is to certify, that this Lufft product has been tested according to the QM of the G. LUFFT Mess- und Regeltechnik GmbH manual in accordance with DIN EN ISO 9001. Ordering specifications are complied with. Execution of instruments / systems as well as testing of accuracy was carried out following LUFFT quality assurance procedures. Quality inspection was successfully passed.

Measurements

	Reference Value	Actual Value	Status
Relative Humidity	50,0 %	50,6 %	✓
Temperature	5,99 °C	5,83 °C	✓
Air Pressure	986,8 hPa	986,8 hPa	✓


Wind Direction and Speed

Angular Deviation

	2,0 m/s	5,0 m/s	10,0 m/s	20,0 m/s	50,0 m/s	Status
RMSE	1,0°	0,4°	0,6°	0,8°	0,9°	✓

Wind Speed Deviation

	2,0 m/s	5,0 m/s	10,0 m/s	20,0 m/s	50,0 m/s	Status
RMS	2,0 m/s	5,0 m/s	10,0 m/s	20,0 m/s	50,0 m/s	✓

Date	Inspector	Quality Management
18.07.2013	 i. A. Bernhard Lenz	 i. V. Rolf Großmann