

Serinus 10 O₃ Analyzer Data Sheet

ID Number

12-1953

Date

10/05/2013

Service Number

NA

Engineer Initials

CY

General Settings

Firmware Version	2.03.0000
Concentration Units	PPM
Temperature Units	C
Pressure Units	torr

Measurement Settings

Average Period	1 Min
Filter Type	Kalman

Serial Communication

Main Gas ID		10
RS232 #1		RS232 #2
Baud	38400	38400
Protocol	Advanced	Advanced

Analog Outputs

Output Type	Voltage
Range	0.5 ppm
Over-Ranging	Disabled
Offset	0.25 V
5.0V CAL	3150
0.5V CAL	632

Voltages

Concentration Voltage	1.23
Reference Voltage	2.96
Analog Supply	12.09
Digital Supply	4.97

Pressures

Gas Flow	0.49
Pressure Units	torr
Ambient	765.53
Cell	744.07
Cal Pressure	742.56

Temperatures

Lamp	50
Chassis	30

Calculation Factors

Instrument Gain	1.00
P/T/F Correction	0.99
O ₃ Background	0.02
Zero Offset	0.00

Digital Pots

Input Pot	076
Lamp Adjust	169
Test Measure	0
Measure Coarse Zero	116
Measure Fine Zero	152
PGA Gain	64

Conc. Voltage (RAW)	1.23
Conc. Voltage	0.02
Lamp Current	9.91
Reference Voltage	2.96

Notes