



Met One Instruments, Inc.
1600 NW Washington Blvd.
Grants Pass, Oregon 97526
Telephone 541-471-7111
Facsimile 541-471-7116

Regional Service
3206 Main St. Suite 106
Rowlett, Texas 75088
Telephone 972-412-4715
Facsimile 972-412-4716

Test Certificate

Model: Temp Sensor BX-592

Sensor Serial No:

P11936

Job Number:

Customer:

Test Date: 2-25-2013

Tested by:

KUTSCHKE

Room Temperature: 23.7 °C

Room Relative Humidity:

31.7 %

Recommended calibration interval is 12 months from the first day of use.

Calibration Standards

Standards	Manufacturer	Model	SN	Cal Due
DMM	HEWLETT PACKARD	3468B	2231A01057	April 13, 2013
DMM	HEWLETT PACKARD	3468B	2231A02768	September 11, 2013
AMBIENT TEMPERATURE	ROSEMONT	T-200	1486222 (N7426)	May 17, 2013
ICE POINT TEMPERATURE	ROSEMONT	T-200	1524475	November 1, 2013
RH/TEMP Sensor	Met One Instruments	083E-1-35	N6200	May 24, 2013

Test Data

TEST	EXPECTED	ACTUAL	ERROR	SPEC	NOTES
0°C	0.375V	0.376 V	0.001 V	±0.005V	
AMBIENT	0.614 V	0.614 V	0.000 V	±0.005V	19.12
ID VOLTAGE	1.800V	1.798 V	-0.002 V	±0.015V	

Test Procedure #592-6100

The standards used for this calibration have accuracies equal to or greater than the instruments tested. These standards are on record and traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated hereon, all instruments are calibrated to meet the manufacturer's published specifications. The Calibration system complies with MIL-STD-45662A.