

ID Number

13-0367

Date

25-Jun-13

Service Number

N/A

Engineer Initials

PRS

General Settings

Firmware Version	2.08.0004
Concentration Units	PPM
Temperature Units	°C
Pressure Units	torr

Measurement Settings

Average Period	1 Min
Filter Type	KALMAN
Background Interval	24 Hrs

Serial Communication

Main Gas ID	55
RS232 #1	RS232 #2
Baud	38400
Protocol	Advanced

Analog Outputs

Output Type	Voltage
Range	0.5
Over-Ranging	Disabled
Offset	0.25
5.0V CAL	3159
0.5V CAL	668

Voltages

Conc Voltage (Raw)	0.15
Reference Voltage	2.57
Analog Supply	11.96
Digital Supply	4.94
High Voltage	712.18

Pressures

Gas Flow	0.71
Pressure Units	torr
Ambient	766.83
Cell	717.20
Cal Pressure	717.47

Temperatures

Heater set point	50
Cell	49.99
Chassis	34.09
Cooler	13.00

Calculation Factors

Instrument Gain	5.45
P/T/F Correction	1.00
Background	0.0009
Zero Offset	0.0000

Digital Pots

High Voltage Adjust	154
High Voltage	712.18
Lamp Adjust	49
Lamp Current	34.73
Input Pot	128
Measure zero	139

Conc Voltage (Raw)	0.15
Conc Voltage	0.00
PGA Gain	128
Reference Zero	0
Reference Gain	17
Reference Voltage	2.57

Notes

Converter Efficiency (No scrubber)	100.00%
H2S Efficiency (With scrubber)	97.90%

SO2 Scrubber Eff	99.00%
------------------	--------

1. This testing result is traceable to NMI Australia via NATA or NIST laboratories.
2. Calibration gas Ecotech reference: CG-1007 (SO2) / CG-1058 (H2S)
3. Datasheets are created based on SO2 testing, then H2S efficiency is measured. This instrument will need to be re-calibrated with H2S upon installation.