



A-2000 CONSOLE SENSOR CALIBRATION

Calibration Date: February 9, 2018

Console S/N: 8077

Reference: 10" manometer, 1" incline

Ashcroft CP24960 - S/N 1071076 (0.25%)

Sensor	Range	Units	Output	Gain	Offset
Pitot	0 - 127	mm H ₂ O	0 - 5 Volts	25.1206	0.6096
Static	-635 - 635	mm H ₂ O	0 - 5 Volts	251.4854	-621.6396
Venturi	0 - 254	mm H ₂ O	0 - 5 Volts	50.2158	0.2377
Barometer	0 - 776	mm Hg	0 - 5 Volts	195.7578	637.1696
Vacuum	+/- 776	mm Hg	0 - 50 mVolts	38963.6000	-962.9140

Calibration performed by: