

**A2000 VENTURI CALIBRATION**

Venturi Calibration with Critical Orifice - Summary

Date: February 9, 2018

Venturi S/N	1653
A2000 Console S/N	8077
Reference Orifice	COK-3013

Bar Pressure	763.8	(mm Hg)
Temperature	23.8	(deg C)

starting n	0.4948
starting k	0.7195

Test Date	Orifice ID	Orifice K'	Vacuum (mm Hg)	Flow Rate CFM	Flow Rate SCFM	Venturi dP (mm H ₂ O)	Y
2/9/2018	3033	0.338	469.9	0.44	0.44	15.6	0.5626
2/9/2018	3033	0.469	463.6	0.61	0.6	29.2	0.7803
2/9/2018	3033	0.551	469.9	0.71	0.706	40.6	0.9160
2/9/2018	3033	0.691	469.9	0.89	0.878	63.1	1.1477
2/9/2018	3033	0.817	457.2	1.06	1.034	87.9	1.3569

Calibration Results	
Venturi S/N 1653	
English Units	Metric Units
k = 0.7229	k = 26.745
n = 0.5082	n = 0.5082

Calibrated by