

Remote mount design

Sensor					
FEP321 (without explosion protection)		FEP325 (Zone 2 / Division 2)		FEP325 (Zone 1 / Division 1)	
1) 	2)  G01083-02	1) 	G00489-01		
Transmitters					
FET321 (without explosion protection)		FET325 (Zone 2, Division 2)		FET325 (Zone 1, Division 1)	
3) 	4)  G01084-02	3) 	4)  G01084-02	4)  G00863-02	
		3) 	4)  G01084-02	3) 	
				4)  G01084-02	3) 
				4)  G01084-02	

Sensor	
Measured error	Default: 0.4 % of measured value, 0.2 % of measured value
Nominal diameter range	DN 3 ... 2000 (1/10 " ... 80 ")
Process connection ⁵⁾	Flange in accordance with DIN 2501 / EN 1092-1, ASME B16.5 / B16.47, JIS 10K
Nominal pressure	PN 10 ... 100, ASME CL 150, 300, 600, 900, 1500, 2500
Liner	Hard rubber (DN 15 ... 2000), soft rubber (DN 50 ... 2000), PTFE (DN 10 ... 600), PFA (DN 3 ... 200), ETFE (DN 25 ... 600), Linatex (DN 50 ... 600)
Conductivity	> 5 µS/cm (20 µS/cm for demineralized water)
Electrodes	Stainless steel, Hastelloy B, Hastelloy C, platinum-iridium, tantalum, titanium, tungsten carbide
Process connection material	Steel, stainless steel
IP rating	IP 65, IP 67, IP 68, (NEMA 4X)
Fluid temperature	-25 ... 180 °C (-13 ... 356 °F)
Transmitters	
Power supply	100 ... 230 V AC (-15 / +10 %), 24 V AC (-30 / +10 %), 24 V DC (-30 / +30 %)
current output	4 ... 20 mA, active or passive
Pulse output	Can be configured locally as active or passive using software
Switch output / switch input	Optoelectronic coupler, programmable function
Display	Graphical display, configurable
Housing	Field mount housing: choice of single-compartment housing or dual-compartment housing
Communication	HART protocol (standard), PROFIBUS PA, FOUNDATION Fieldbus (option)
Approvals	
Explosion protection approvals	• ATEX / IECEx zone 1, 2, 21, 22 • NEPSI zone 1, 2 • FM / cFM Cl 1 Div 1 (≤ DN 300), Cl 1 Div 2 • GOST zone 1, 2
Pressure Equipment Directive 97/23/EC	Conformity assessment in accordance with category III, fluid group 1
CRN (Canadian Reg.Number)	On request

1) Design level "B" sensor.

2) Design level "C" sensor, DN 25 ... 600.

3) Single-compartment housing.

4) Dual-compartment housing.

5) For information on flange thicknesses, see the chapters entitled "Dimensions for sensor design level „B"" on page 37 and "Dimensions for sensor design level „C"" on page 43.