

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

AUTOLIV AAM
Brigham Facility, Metrology Laboratory
250 American Way
Brigham City, UT 84302
Mark Bertagnolli Phone: 810 510 0612

CALIBRATION

Valid To: April 30, 2017

Certificate Number: 403.15

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Length Standards	Up to 11 in	(40 + 1.5L) μin	Super mic w/gage blocks
OD – Gages, (Pin, Plugs)	(0.01 to 2) in	(45 + 1.5L) μin	Laser mic w/master plugs
Micrometers	Up to 12 in	(59 + 1.5L) μin	Gage blocks Gage blocks (long block set)
Calipers	Up to 12 in	(560 + 1.5L) μin	Gage blocks
Indicators	Up to 4 in	(100 + 1.5L) μin	Gage blocks

II. Dimensional Testing/Calibration¹

Parameter/Equipment	Range	CMC ^{2, 3} ($\pm$)	Comments
Length ⁴ – 3D X – Axis Y – Axis Z – Axis	Up to 24 in Up to 38 in Up to 22 in	(1000 + 10L) μ in	CMM: Mitutoyo BHN710

III. Mechanical

Parameter/Equipment	Range	CMC ² ($\pm$)	Comments
Torque – Tools / Wrenches	(4 to 50) in·lbf (30 to 400) in·lbf (80 to 1000) in·lbf (20 to 250) ft·lbf (50 to 600) ft·lbf	1 % of reading 1 % of reading 1 % of reading 1.2 % of reading 1.2 % of reading	Torque system CDI – 600 TL with transducers CDI 2000-400-02 and 2000-12-02

¹ This laboratory offers commercial dimensional testing and calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches.

⁴ This laboratory meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program for the types of dimensional tests listed above and is considered equivalent to that of a calibration.



Accredited Laboratory

A2LA has accredited

AUTOLIV AAM

Brigham City, UT

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 12th day of May 2015.

A handwritten signature in blue ink, appearing to read 'J. C. Bunt'.

Senior Director of Quality and Communications
For the Accreditation Council
Certificate Number 403.15
Valid to April 30, 2017
Revised March 23, 2016

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.