

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

ACUMENT GLOBAL TECHNOLOGIES - NORTH AMERICA - CALIBRATION LABORATORY

826 E. Madison St
Belvidere, IL 61008
David Goss Phone: 815-544-7589

CALIBRATION

Valid To: May 31, 2018

Certificate Number: 1155.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ² (±)	Comments ³
Calipers	Up to 6 in	800 µin	Class 2 gage blocks, master rings
Micrometers	Up to 1 in (1 to 6) in	75 µin 110 µin	Class 2 gage blocks
Indicators	Up to 1 in	200 µin	Indicator calibrator, class 2 gage blocks
Inside Diameter	(0.04 to 1) in	35 µin	P&W Labmaster Universal™, master rings
Outside Diameter	Up to 1 in	35 µin	Class 2 gage blocks, Supermicrometer™
Thread Pitch Diameter, O.D.	Up to 1 in	100 µin	3 wire measurement: master thread wires, Supermicrometer™

Parameter/Equipment	Range	CMC ² (±)	Comments ³
V-Micrometer	(0.052 to 1) in	110 µin	Master gage pins
Gage Blocks – Shop Grade Only	Up to 1 in (1 to 4) in	35 µin 40 µin	P&W Labmaster Universal™
Torx Configuration – Circumscribed Diameter	Up to 1 in	35 µin	Supermicrometer™
Inscribed Diameter	Up to 1 in	40 µin	Height gage comparator
Hex Configuration – Across Corners	Up to 1 in	35 µin	Supermicrometer™
Across Flats	Up to 1 in	35 µin	

II. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Torque Wrenches	(0 to 500) ft·lbf (0 to 200) in·lbf	1.0 % full scale 1.0 % full scale	A.K.O. torque wrench calibrator

¹ This laboratory is not normally available for commercial calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ “Labmaster Universal” and “Supermicrometer” are registered trademarks with a last listed owner of Pratt & Whitney Measurement Systems, Inc., Connecticut U.S.A.



Accredited Laboratory

A2LA has accredited

ACUMENT GLOBAL TECHNOLOGIES - NORTH AMERICA - CALIBRATION LABORATORY

Belvidere, IL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 9th day of June, 2016.

A handwritten signature in blue ink, appearing to read 'A. C. Burt'.

Senior Director of Quality and Communications
For the Accreditation Council
Certificate Number 1155.01
Valid to May 31, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.