



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

AMSLER TESTING INC.
28364 South Western Ave. # 439
Rancho Palos Verdes, CA 90275
Sven Neumann Phone: 310 221 0037

CALIBRATION

Valid To: July 31, 2017

Certificate Number: 1261.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Alignment Static Verification ³	(200 to 240 000) lbf	1.5 % of indicated value	NASM 1312
Dynamic Load Verification ³	(100 to 240 000) lbf	0.5 % of indicated value	NASM 1312
Static Load Calibration ³	(200 to 240 000) lbf	0.25 % of indicated value	ASTM E4

¹ Commercial calibration service and field calibration service is sometimes available for this laboratory.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

A handwritten signature in blue ink, appearing to be 'L. L. L.', is written over the page number.



Accredited Laboratory

A2LA has accredited

AMSLER TESTING INC.

Rancho Palos Verdes, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 13th day of July 2015.

A handwritten signature in black ink, appearing to be 'L. J. ...', is written over a horizontal line.

President & CEO
For the Accreditation Council
Certificate Number 1261.01
Valid to July 31, 2017
Revised December 08, 2016

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.