



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

BCI TEST AND CALIBRATION LABORATORY
1328 Azalea Garden Road
Norfolk, VA 23502
J. Lyle Bagley Phone: 757 943 1348
Ryan Sposito Phone: 757 855 6006 X 2238

CALIBRATION

Valid To: April 30, 2018

Certificate Number: 2688.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ^{2,3} ($\pm$)	Comments
Pressure –			
High Range Piston	(200 to 10 000) psig	0.005 % IV + 0.03 psig	Ametek deadweight tester
Low Range Piston	(20 to 1500) psig	0.0056 % IV + 0.02 psig	
Pressure Relief Valve Setting	(36 to 950) psig (>950 to 10 000) psig	1 % IV + 0.5 psig 1 % IV + 5 psig	Druck/GE DPI0-104

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, IV is defined as the indicated value.



Accredited Laboratory

A2LA has accredited

BCI TEST AND CALIBRATION LABORATORY

Norfolk, VA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 16th day of November 2016.

A handwritten signature in black ink, appearing to be 'L. S.', written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 2688.01
Valid to April 30, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.