

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

ATEQ CORPORATION  
35980 Industrial Road  
Livonia, MI 48150  
Bill Whitlow Phone: 734 838 3100

## CALIBRATION

Valid To: September 30, 2018

Certificate Number: 3049.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

## I. Fluid Quantities

| Parameter/Equipment | Range            | CMC <sup>2</sup> (±)   | Comments       |
|---------------------|------------------|------------------------|----------------|
| Gas Flow            | (0.2 to 2) sccm  | 3.1 % rdg + 0.011 sccm | ATEQ CDF F17-0 |
|                     | (2 to 20) sccm   | 2.0 % rdg + 0.08 sccm  | ATEQ CDF F17-1 |
|                     | (20 to 250) sccm | 3.6 % rdg + 1.3 sccm   | ATEQ CDF F17-2 |

## II. Mechanical

| Parameter/Equipment | Range                                                                                                                       | CMC <sup>2</sup> (±)                                                                  | Comments |
|---------------------|-----------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------|----------|
| Pressure – Gas      | (0 to 1245) Pa<br>(0 to 1) psig<br>(0 to 3) psig<br>(0 to 30) psig<br>(0 to 100) psig<br>(0 to 300) psig<br>(0 to -13) psig | 1.7 Pa<br>0.003 psig<br>0.02 psig<br>0.06 psig<br>0.15 psig<br>0.36 psig<br>0.04 psig | Gauge    |

---

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



# ***Accredited Laboratory***

A2LA has accredited

**ATEQ CORPORATION**

*Livonia, MI*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 17<sup>th</sup> day of August 2016.

A handwritten signature in blue ink, appearing to read 'J. C. Bunt'.

Senior Director of Quality and Communications  
For the Accreditation Council  
Certificate Number 3049.01  
Valid to September 30, 2018

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*