



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

ACETRONIC INDUSTRIAL CONTROLS, INC.

7015 Ordan Drive
Mississauga, Ontario
Canada L5T 1Y2

Kim Thiara Phone: 905 564 7227

CALIBRATION

Valid To: March 31, 2017

Certificate Number: 3665.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouple Indicators –			
Type E	(0 to 1000) °C	1.0 °C	Fluke 741B
Type J	(0 to 1200) °C	1.0 °C	
Type K	(0 to 1272) °C	1.0 °C	
Type R	(0 to 1767) °C	1.7 °C	
Type S	(0 to 1767) °C	1.7 °C	
Type T	(0 to 400) °C	1.0 °C	
Type E	(32 to 1832) °F	1.4 °F	
Type J	(32 to 2192) °F	1.2 °F	
Type K	(32 to 2500) °F	1.4 °F	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouple Indicators – (cont.)			
Type R	(32 to 3213) °F	3.0 °F	Fluke 741B
Type S	(32 to 3213) °F	2.9 °F	
Type T	(32 to 752) °F	1.4 °F	
Electrical Calibration of RTD Indicators & Indicating Systems –			
Pt(385), 100 Ω	(0 to 800) °C	1.0 °C	Fluke 741B
Pt(385), 1000 Ω	(0 to 630) °C	1.0 °C	
Pt(385), 100 Ω	(32 to 1472) °F	1.2 °F	
Pt(385), 1000 Ω	(32 to 1166) °F	1.2 °F	
Ni(672), 120 Ω	(0 to 260) °C	1.0 °C	
Ni(672), 120 Ω	(32 to 500) °F	1.0 °F	

II. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Pressure –			
Pneumatic	(0 to 500) psi	0.7 psi	Fluke 700P07
Hydraulic	(0 to 10 000) psi	25 psi	Fluke 700P31

III. Thermodynamics

Parameter/Equipment	Range	CMC ² (±)	Comments
Temperature Sensors – Types J and K	(35 to 260) °C	1.3 °C	Versatile 4-wire RTD Probe
Types J and K	(95 to 500) °F	2.3 °F	
Temperature Sensors Type RTD –			
Pt(385), 100 Ω	(35 to 260) °C	1.2 °C	Versatile 4-wire RTD Probe
Pt(385), 100 Ω	(95 to 500) °F	2.3 °F	

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



Accredited Laboratory

A2LA has accredited

ACETRONIC INDUSTRIAL CONTROLS INC.

Mississauga, Ontario, CANADA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 22nd day of October 2014.

A handwritten signature in black ink, appearing to be 'L. L. L.', written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 3665.01
Valid to March 31, 2017
Revised February 21, 2017

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.