

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

APPLIED AUTOMATION TECHNOLOGIES, INC.

1688 Star Batt Drive
Rochester Hills, MI 48309
Kenon Skidmore Phone: 248 656 4930

CALIBRATION

Valid To: September 30, 2018

Certificate Number: 3879.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ² (±)	Comments
CMM ³ –			ASME B89.4.10360.2-2008 is used as a guideline
Repeatability	Up to 45 mm	0.0005 µm	Master sphere
Volumetric Accuracy	(300 to 1000) mm	(1.1 + 0.001) µm	Ball bar

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



Accredited Laboratory

A2LA has accredited

APPLIED AUTOMATION TECHNOLOGIES, INC.

Rochester Hills, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 28th day of September 2016.

A handwritten signature in blue ink, likely of the Senior Director of Quality and Communications.

Senior Director of Quality and Communications
For the Accreditation Council
Certificate Number 3879.01
Valid to September 30, 2018
Revised February 3, 2017

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.