

# Air quality — Certification of automated measuring systems

**Part 2: Initial assessment of the AMS  
manufacturer's quality management  
system and post certification  
surveillance for the manufacturing  
process**

ICS 13.040.99

# National foreword

This British Standard is the UK implementation of EN 15267-2:2009.

The UK participation in its preparation was entrusted to Technical Committee EH/2/1, Stationary source emission.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2009  
© BSI 2009

ISBN 978 0 580 57915 8

## Amendments/corrigenda issued since publication

| Date | Comments |
|------|----------|
|      |          |
|      |          |
|      |          |
|      |          |

EUROPEAN STANDARD

**EN 15267-2**

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2009

ICS 13.040.99

English Version

**Air quality - Certification of automated measuring systems - Part  
2: Initial assessment of the AMS manufacturer's quality  
management system and post certification surveillance for the  
manufacturing process**

Qualité de l'air - Certification des systèmes de mesurage automatisés - Partie 2 : Evaluation initiale du système de gestion de la qualité des fabricants d'AMS et surveillance après certification du procédé de fabrication

Luftbeschaffenheit - Zertifizierung von automatischen Messeinrichtungen - Teil 2: Erstmalige Beurteilung des Qualitätsmanagementsystems des Herstellers und Überwachung des Herstellungsprozesses nach der Zertifizierung

This European Standard was approved by CEN on 14 February 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

Page

|                                                                                     |    |
|-------------------------------------------------------------------------------------|----|
| Foreword.....                                                                       | 3  |
| Introduction .....                                                                  | 4  |
| 1 Scope .....                                                                       | 5  |
| 2 Normative references .....                                                        | 5  |
| 3 Terms and definitions .....                                                       | 5  |
| 4 Quality management systems .....                                                  | 7  |
| 4.1 General requirements.....                                                       | 7  |
| 4.2 Documentation requirements.....                                                 | 7  |
| 5 Management responsibility .....                                                   | 8  |
| 5.1 Management commitment .....                                                     | 8  |
| 5.2 Customer focus.....                                                             | 8  |
| 5.3 Quality policy .....                                                            | 8  |
| 5.4 Planning.....                                                                   | 8  |
| 5.5 Responsibility, authority and communication.....                                | 9  |
| 5.6 Management review.....                                                          | 9  |
| 6 Resource management .....                                                         | 10 |
| 6.1 Provision of resources.....                                                     | 10 |
| 6.2 Human resources.....                                                            | 10 |
| 6.3 Infrastructure.....                                                             | 10 |
| 6.4 Work environment .....                                                          | 10 |
| 7 Product realisation .....                                                         | 10 |
| 7.1 Planning of product realisation.....                                            | 10 |
| 7.2 Customer-related processes .....                                                | 10 |
| 7.3 Design and development .....                                                    | 11 |
| 7.4 Purchasing .....                                                                | 12 |
| 7.5 Production and service provision.....                                           | 13 |
| 7.6 Control of monitoring and measuring devices.....                                | 13 |
| 8 Measurement, analysis and improvement .....                                       | 13 |
| 8.1 General.....                                                                    | 13 |
| 8.2 Monitoring and measurement .....                                                | 13 |
| 8.3 Control of non-conforming product.....                                          | 14 |
| 8.4 Analysis of data .....                                                          | 14 |
| 8.5 Improvement .....                                                               | 14 |
| 9 Assessment.....                                                                   | 15 |
| 9.1 Initial assessment.....                                                         | 15 |
| 9.2 Continuing surveillance .....                                                   | 15 |
| Annex A (informative) Process for the control and assessment of design changes..... | 16 |
| Bibliography .....                                                                  | 17 |

## Foreword

This document (EN 15267-2:2009) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2009, and conflicting national standards shall be withdrawn at the latest by September 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document is Part 2 of a series of European Standards:

EN 15267-1, *Air quality — Certification of automated measuring systems — Part 1: General principles*

EN 15267-2, *Air quality — Certification of automated measuring systems — Part 2: Initial assessment of the AMS manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN 15267-3, *Air quality — Certification of automated measuring systems — Part 3: Performance criteria and test procedures for automated measuring systems for monitoring emissions from stationary sources*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

## Introduction

The product certification of automated measuring systems (AMS) supports the requirements of certain Directives of the European Union (EU), which require, either directly or indirectly, that AMS comply with performance criteria, maximum permissible measurement uncertainties and testing requirements. These Directives include the Directive on the limitation of emissions of certain pollutants into the air from large combustion plants [1], the Directive on the incineration of waste [2] and the Framework Directive on ambient air quality assessment and management [3] and the associated daughter directives [4], [5], [6] and [7].

An AMS will typically undergo design changes during its product life. It is essential to ensure that such changes do not alter the AMS such that it no longer conforms with its certified performance. In order to control such design changes for the product certification of AMS this European Standard specifies the requirements for

- the manufacturer's quality management system,
- the initial assessment of an AMS manufacturer's production control, and
- the continuing surveillance of the effect on performance of certified AMS of subsequent design changes.

This European Standard follows the structure of EN ISO 9001:2000 such that clause numbers in Clauses 4 to 8 of this standard coincide with those of EN ISO 9001:2000. However, this European Standard does not preclude the use of other quality management systems that are compatible with the objectives of EN ISO 9001:2000.

## 1 Scope

This European Standard specifies the requirements for the manufacturer's quality management system, the initial assessment of the manufacturer's production control and the continuing surveillance of the effect of subsequent design changes on the performance of certified automated measuring systems.

This European Standard also serves as a reference document for auditing the manufacturer's quality management system.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 9001:2000, *Quality management systems — Requirements (ISO 9001:2000)*

EN ISO/IEC 17021, *Conformity assessment — Requirements for bodies providing audit and certification of management systems (ISO/IEC 17021:2006)*

EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **automated measuring system**

##### **AMS**

entirety of all measuring instruments and additional devices for obtaining a result of measurement

NOTE 1 Apart from the actual measuring device (the analyser), an AMS includes facilities for taking samples (e.g. probe, sample gas lines, flow meters and regulator, delivery pump) and for sample conditioning (e.g. dust filter, pre-separator for interferences, cooler, converter). This definition also includes testing and adjusting devices that are required for functional checks and, if applicable, for commissioning.

NOTE 2 The term "automated measuring system" (AMS) is typically used in Europe. The terms "continuous emission monitoring system" (CEM) and "continuous ambient-air-quality monitoring system" (CAM) are also typically used in the UK and USA.

[EN 15267-1:2008, 3.1]

### 3.2

#### **relevant body**

competent authority or certification body, nominated by a competent authority or EU member state, that carries out the certification of automated measuring systems

[EN 15267-1:2008, 3.2]

### 3.3

#### **competent authority**

organisation which implements the requirements of EU Directives and regulates installations, which must comply with the requirements of applicable European Standards

[EN 15267-1:2008, 3.3]

**3.4**  
**certification body**  
any body operating a product certification system or any body accredited to EN ISO/IEC 17021 for the certification of quality management systems

[EN 15267-1:2008, 3.4]

**3.5**  
**manufacturer**  
organisation, situated at a stated location or locations, that carries out or controls such stages in the manufacture, assessment, handling and storage of a product that enables it to accept responsibility for continued compliance of the product and its certification, and undertakes all obligations in that connection

NOTE The term "manufacturer" is used instead of "organisation" as used in EN ISO 9001. For the purpose of this document they are interchangeable.

[EN 15267-1:2008, 3.5]

**3.6**  
**test laboratory**  
laboratory accredited to EN ISO/IEC 17025 for carrying out performance tests on automated measuring systems in accordance with applicable European Standards

[EN 15267-1:2008, 3.6]

**3.7**  
**product**  
automated measuring system

[EN 15267-1:2008, 3.7]

**3.8**  
**technical documentation**  
information for the operation of automated measuring systems, such as manuals

**3.9**  
**technical file**  
record of the reference documents and design changes to the reference documents

[EN 15267-1:2008, 3.8]

**3.10**  
**reference document**  
document that controls the manufacture and design of an AMS and is referenced in the test report

NOTE Reference documents can include drawings, specifications, instructions and computer code.

[EN 15267-1:2008, 3.9]

**3.11**  
**related document**  
document not referenced in the test report

NOTE A related document can be used, for example, for the detailed manufacture of component parts.

[EN 15267-1:2008, 3.10]



### 3.12

#### **certification range**

range over which the automated measuring system is tested and certified for compliance with the relevant performance criteria

NOTE 1 The lower limit is typically the detection limit of the AMS and often considered to be zero.

NOTE 2 Generally, the lower the certification range, the better the performance of the AMS. Also an AMS typically performs satisfactorily at higher values over the measurement range.

[EN 15267-1:2008, 3.11]

### 3.13

#### **surveillance**

systematic iteration of conformity assessment activities as a basis for maintaining the validity of the statement of conformity

NOTE For the purposes of this European Standard surveillance focuses on the manufacturer's quality management system to ensure that automated measuring systems continue to comply with the standard to which they are certified.

[EN 15267-1:2008, 3.12]

### 3.14

#### **legislation**

directives, acts, ordinances and regulations

[EN 15267-1:2008, 3.13]

## **4 Quality management systems**

### **4.1 General requirements**

The manufacturer's quality management system shall ensure compliance of all automated measuring systems, of the type that has been product certified, with the performance criteria specified in the relevant certificate.

### **4.2 Documentation requirements**

#### **4.2.1 General**

The requirements specified in 4.2.1 of EN ISO 9001:2000 apply.

The manufacturer shall have methods to keep up-to-date with the requirements of legislation, standards and guidelines that are relevant to its certified AMS and AMS for which it seeks certification in accordance with the EN 15267 series of standards.

#### **4.2.2 Quality manual**

The requirements specified in 4.2.2 of EN ISO 9001:2000 apply.

The quality management system manual shall ensure that no factor (e.g. measured components, scope) defined within the AMS certificate and technical documentation (e.g. reference documents) is modified without appropriate evaluation of its impact on the performance of the AMS.

#### **4.2.3 Control of documents**

The requirements specified in 4.2.3 of EN ISO 9001:2000 apply.

Documents that control the manufacture and design of AMS are referred to as reference documents. They can include drawings, specifications, instructions and computer code.

The AMS manufacturer shall ensure that all related documents are cross-referenced to the relevant reference documents for the AMS.

Where there are common reference documents associated with more than one AMS certificate, the manufacturer shall have methods to ensure simultaneous supplementary action in the event of an amendment to such documents.

Where a manufacturer also has documents for AMS not intended for testing and certification, then the manufacturer shall have a method that enables both the related documents and reference documents to be clearly identified.

The AMS manufacturer shall either retain, or have access to any legislation and CEN standards which contains performance criteria for AMS.

#### **4.2.4 Control of records**

The requirements specified in 4.2.4 of EN ISO 9001:2000 apply.

### **5 Management responsibility**

#### **5.1 Management commitment**

The requirements specified in 5.1 of EN ISO 9001:2000 apply.

The manufacturer shall make a commitment to know and understand the monitoring requirements and applicable performance criteria for AMS, and the impact of these requirements on instrument design, manufacturing and certification.

#### **5.2 Customer focus**

The requirements specified in 5.2 of EN ISO 9001:2000 apply.

#### **5.3 Quality policy**

The requirements specified in 5.3 of EN ISO 9001:2000 apply.

#### **5.4 Planning**

##### **5.4.1 Quality objectives**

The requirements specified in 5.4.1 of EN ISO 9001:2000 apply.

##### **5.4.2 Quality management system planning**

The requirements specified in 5.4.2 of EN ISO 9001:2000 apply.

## **5.5 Responsibility, authority and communication**

### **5.5.1 Responsibility and authority**

The requirements specified in 5.5.1 of EN ISO 9001:2000 apply.

The manufacturer shall designate persons for:

- the effective co-ordination of processes with respect to AMS intended for use at installations and monitoring activities which require certified products,
- determining the need to liaise with the testing laboratories and relevant bodies responsible for AMS testing and certification, with respect to any proposed change to the design defined in the AMS certificate and the technical documentation, and
- authorising and controlling reference documents.

### **5.5.2 Management representative**

The requirements specified in 5.5.2 of EN ISO 9001:2000 apply.

The manufacturer shall have a representative who has overall responsibility for:

- ensuring the compliance of all AMS, of the type that has been product certified, with the performance criteria specified in the relevant certificate, and
- design changes to the type of AMS that has been product certified.

### **5.5.3 Internal communication**

The requirements specified in 5.5.3 of EN ISO 9001:2000 apply.

## **5.6 Management review**

### **5.6.1 General**

The requirements specified in 5.6.1 of EN ISO 9001:2000 apply.

The maximum interval between management reviews, normally every 12 months, has to be defined by the manufacturer to assure an appropriate, adapted and efficient quality management system.

The persons responsible for the activities as detailed in 5.5.1 and 5.5.2 shall participate in the review.

### **5.6.2 Review input**

The requirements specified in 5.6.2 of EN ISO 9001:2000 apply.

The review shall include the overall effectiveness of the quality management system with respect to AMS intended for use at installations requiring certified products.

A review of AMS conformity shall be used as an indicator of the effectiveness of the quality management system.

### **5.6.3 Review output**

The requirements specified in 5.6.3 of EN ISO 9001:2000 apply.

## **6 Resource management**

### **6.1 Provision of resources**

The requirements specified in 6.1 of EN ISO 9001:2000 apply.

### **6.2 Human resources**

#### **6.2.1 General**

The requirements specified in 6.2.1 of EN ISO 9001:2000 apply.

#### **6.2.2 Competence, awareness and training**

The requirements specified in 6.2.2 of EN ISO 9001:2000 apply.

Appropriate staff shall be trained in the requirements of the applicable European standards for monitoring and measurement, and AMS performance criteria.

### **6.3 Infrastructure**

The requirements specified in 6.3 of EN ISO 9001:2000 apply.

### **6.4 Work environment**

The requirements specified in 6.4 of EN ISO 9001:2000 apply.

## **7 Product realisation**

### **7.1 Planning of product realisation**

The requirements specified in 7.1 of EN ISO 9001:2000 apply.

### **7.2 Customer-related processes**

#### **7.2.1 Determination of requirements related to the product**

The requirements specified in 7.2.1 of EN ISO 9001:2000 apply.

The manufacturer shall:

- determine the type of AMS and its application,
- identify the performance criteria for each measured component and the site conditions, which affect AMS performance, and
- identify any special conditions, which could impact upon the AMS's ability to meet the applicable performance criteria.

#### **7.2.2 Review of requirements related to the product**

The requirements specified in 7.2.2 of EN ISO 9001:2000 apply.

### 7.2.3 Customer communication

The requirements specified in 7.2.3 of EN ISO 9001:2000 apply.

## 7.3 Design and development

### 7.3.1 Design and development planning

The requirements specified in 7.3.1 of EN ISO 9001:2000 apply.

An illustration of the process for the control and assessment of design changes is given in Annex A.

### 7.3.2 Design and development inputs

The requirements specified in 7.3.2 of EN ISO 9001:2000 apply.

The manufacturer shall maintain a technical file of design changes for each type of AMS. The technical file shall contain at least the following information:

- component changes;
- dates of the changes;
- assessment of the impact of the changes, with respect to the applicable performance criteria;
- data supporting the assessment of the impact of the changes;
- assessment of the significant changes by the test laboratory;
- references to linked information, such as communications with the certification body.

When the manufacturer makes design changes, then the manufacturer shall notify the test laboratory and relevant body, unless the manufacturer has evidence to show that the AMS still meets the performance stated in the original certificate.

The following three classes of changes to certified AMS are defined:

- Type 0: changes that have no measurable influence to the performance of the AMS;
- Type 1: changes that can have an influence on the performance of the AMS, but where subsequent tests prove that such changes do not have a significant influence;
- Type 2: changes that have a significant influence on the performance of the AMS.

A significant influence is considered to be one that reduces the performance of the AMS compared to that recorded in the certificate for the stipulated performance characteristics.

The manufacturer shall evaluate all changes to a certified AMS. Where Type 2 changes are identified the manufacturer shall consult the relevant body and test laboratory to confirm whether any additional testing is required to determine the impact of any design changes. The manufacturer shall document all changes and evaluations in accordance with the requirements of this European Standard and in such a way that they can be audited.

## 7.4 Purchasing

### 7.4.1 Purchasing process

The requirements specified in 7.4.1 of EN ISO 9001:2000 apply.

The manufacturer's quality management system shall have the following provisions for the control of the purchasing process.

While the manufacture, test and final inspection may be sub-contracted, the responsibility for ensuring conformance with the product certificate shall not be sub-contracted. This is not for products, processes or services where the manufacturer fully verifies each item for conformance.

A supplier providing a product, process, or service that can affect the AMS's compliance with the performance stated in the AMS certificate shall only be selected after an evaluation has demonstrated that they have the capability of ensuring compliance with all specified requirements.

The evaluation shall be made by one or more of the following methods:

- the supplier has third party management system certification to the appropriate standard and scope issued by an accredited body, which can demonstrate that it operates in compliance EN ISO/IEC 17021;
- a documented evaluation, which provides objective evidence that the supplier can provide product, process or service that are fit for purpose;
- a documented site assessment by competent personnel to ensure that all relevant controls are available, documented, understood and effective;
- inspection and test data relevant to the performance criteria, where such test data is produced in facilities which meet the requirements of EN ISO/IEC 17025.

Where the features affecting the performance can not be verified at a later stage, then the evaluation shall include initial and periodic site assessments at the supplier's premises to ensure relevant controls are available, documented, understood and effective.

Suppliers not used for a period exceeding one year shall be re-evaluated prior to the placing of the contract. This is not mandatory for products, processes or services where the manufacturer fully verifies each item for conformance.

The ongoing ability of the supplier to provide conforming product, process or service shall be assessed at periods not exceeding one year.

### 7.4.2 Purchasing information

The requirements specified in 7.4.2 of EN ISO 9001:2000 apply.

The purchasing documents shall clearly specify the requirements for a sub-contracted product or service described in the AMS certificate and within the manufacturer's documentation for the AMS.

NOTE Manufacturer's documents can be instructions, related documents, data sheets and sales literature.

For items where conformance cannot be verified after manufacture, the purchasing information shall set out the specific management procedures, resources and sequence of activities relevant to the particular item. This shall include inspection and test data to the appropriate accreditation standard.

### 7.4.3 Verification of purchased product

The requirements specified in 7.4.3 of EN ISO 9001:2000 apply.

For purchased products that are components of the AMS, the manufacturer shall verify their suitability and quality by assessing the evidence which shows that the supplied products meet the required specifications. This shall be achieved by at least one or more of the following methods:

- the components supplied are independently tested at a suitably-qualified third-party test laboratory;
- the supplier evaluates the components and the manufacturer has verified that the suppliers monitoring and measurement processes are performed to a suitable level of quality;
- the manufacturer performs an appropriate number of tests on representative batches of the supplied product.

## **7.5 Production and service provision**

### **7.5.1 Control of production and service provision**

The requirements specified in 7.5.1 of EN ISO 9001:2000 apply.

### **7.5.2 Validation of processes for production and service provision**

The requirements specified in 7.5.2 of EN ISO 9001:2000 apply.

### **7.5.3 Identification and traceability**

The requirements specified in 7.5.3 of EN ISO 9001:2000 apply.

The manufacturer shall establish and maintain procedures to identify and trace types of components used during all stages of production, testing, final inspection and placing of AMS on the market. This is in addition to the requirement for traceability of the final and complete AMS.

### **7.5.4 Customer property**

The requirements specified in 7.5.4 of EN ISO 9001:2000 apply.

### **7.5.5 Preservation of product**

The requirements specified in 7.5.5 of EN ISO 9001:2000 apply.

## **7.6 Control of monitoring and measuring devices**

The requirements specified in 7.6 of EN ISO 9001:2000 apply.

## **8 Measurement, analysis and improvement**

### **8.1 General**

The requirements specified in 8.1 of EN ISO 9001:2000 apply.

### **8.2 Monitoring and measurement**

#### **8.2.1 Customer satisfaction**

The requirements specified in 8.2.1 of EN ISO 9001:2000 apply.

### **8.2.2 Internal audit**

The requirements specified in 8.2.2 of EN ISO 9001:2000 apply.

### **8.2.3 Monitoring and measurement of processes**

The requirements specified in 8.2.3 of EN ISO 9001:2000 apply.

### **8.2.4 Monitoring and measurement of AMS**

The requirements specified in 8.2.4 of EN ISO 9001:2000 apply.

The manufacturer shall have the means for inspection and testing that verify that the AMS meets the required performance criteria.

The manufacturer shall refer to the applicable requirements when testing products, but may choose a representative form of inspection provided that the manufacturer has verifiable data to demonstrate that such simplified testing is representative of the test procedures.

## **8.3 Control of non-conforming product**

The requirements specified in 8.3 of EN ISO 9001:2000 apply.

The manufacturer shall maintain a system such that in the event of AMS not complying with the product certificate and having been supplied, then the manufacturer's customer can be identified. Where a non-conforming AMS is supplied to a customer, the manufacturer shall take appropriate action, to the effects, or potential effects of the non-conformity. The manufacturer shall also inform in writing, the customer and the relevant body.

For all non-conforming AMS that has been supplied to a customer, the manufacturer shall maintain, for a minimum period of 5 years, records of the actions taken

- to inform customers and the relevant body, and
- to implement any necessary corrective and preventative actions.

## **8.4 Analysis of data**

The requirements specified in 8.4 of EN ISO 9001:2000 apply.

## **8.5 Improvement**

### **8.5.1 Continual improvement**

The requirements specified in 8.5.1 of EN ISO 9001:2000 are not within the scope of this standard.

### **8.5.2 Corrective action**

The requirements specified in 8.5.2 of EN ISO 9001:2000 apply.

### **8.5.3 Preventive action**

The requirements specified in 8.5.3 of EN ISO 9001:2000 apply.



## 9 Assessment

### 9.1 Initial assessment

The manufacturer shall have a method of demonstrating compliance to the requirements of this European Standard.

NOTE Compliance can be demonstrated by an audit carried out by the relevant body (or a party acting on behalf of the relevant body). Alternatively, where a manufacturer has incorporated the requirements of this European Standard into a certified quality management system (e.g. EN ISO 9001:2000) the relevant body can accept an assessment carried out by a certification body accredited to EN ISO/IEC 17021 provided that the scope of such an assessment also includes the requirements of this European Standard.

When applying for certification of an AMS, the manufacturer shall include a copy of the report on the initial assessment.

### 9.2 Continuing surveillance

Following certification of an AMS, the manufacturer shall have a method of demonstrating continued compliance with the performance criteria stated in the certificate.

This post certification surveillance should focus in particular on any design changes to the reference documents for the certified AMS, their reporting to the test laboratory and relevant body, and the results of any associated performance evaluation.

NOTE Surveillance is carried out by or on behalf of the relevant body (see EN 15267-1).

## Annex A (informative)

### Process for the control and assessment of design changes




Figure A.1 — Illustration of the process for the control and assessment of design changes

## Bibliography

- [1] Directive 2001/80/EC of the European Parliament and of the Council of 23 October 2001 on the limitation of emissions of certain pollutants into the air from large combustion plants, OJ L 309, p. 1
- [2] Directive 2000/76/EC of the European Parliament and of the Council of 4 December 2000 on the incineration of waste, OJ L 332, p. 91
- [3] Council Directive 96/62/EC of 27 September 1996 on ambient air quality assessment and management, OJ. L 296, p. 55
- [4] Council Directive 1999/30/EC of 22 April 1999 relating to limit values for sulphur dioxide, nitrogen dioxide and oxides of nitrogen, particulate matter and lead in ambient air, OJ L 163, p. 41
- [5] Directive 2000/69/EC of the European Parliament and of the Council of 16 November 2000 relating to limit values for benzene and carbon monoxide in ambient air, OJ L 313, p. 12
- [6] Directive 2002/3/EC of the European Parliament and of the Council of 12 February 2002 relating to ozone in ambient air, OJ L 67, p. 14
- [7] Directive 2004/107/EC of the European Parliament and the Council of 15 December 2004 relating to arsenic, cadmium, mercury, nickel and polycyclic aromatic hydrocarbons in ambient air, OJ L 23, p. 3

## BSI - British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

### Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

### Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001 Email: [orders@bsigroup.com](mailto:orders@bsigroup.com) You may also buy directly using a debit/credit card from the BSI Shop on the Website <http://www.bsigroup.com/shop>

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

### Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact Information Centre. Tel: +44 (0)20 8996 7111 Fax: +44 (0)20 8996 7048 Email: [info@bsigroup.com](mailto:info@bsigroup.com)

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002 Fax: +44 (0)20 8996 7001 Email: [membership@bsigroup.com](mailto:membership@bsigroup.com)

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsigroup.com/BSOL>

Further information about BSI is available on the BSI website at <http://www.bsigroup.com>.

### Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright and Licensing Manager. Tel: +44 (0)20 8996 7070 Email: [copyright@bsigroup.com](mailto:copyright@bsigroup.com)