


Quick Reference Guide


INDICATIONS FOR USE: The DERMACLOSE Continuous External Tissue Expander is indicated for use in assisting with the closure of moderate to large surgical or traumatic acute full thickness wounds of the skin by approximating and reducing the size of the wound.

CONTRAINDICATIONS: The DERMACLOSE Continuous External Tissue Expander should not be used on ischemic, infected, or acute burned tissue. It should not be used on fragile tissue at the edges of a wound.




1. Undermine wound edges

Undermine or elevate wound margins on a supra-fascial plane by approximately half the width of the wound.




2. Close wound ends

Prior to applying DERMACLOSE device, close as much of the distal and proximal wound ends as possible.




See DERMACLOSE package inserts for full instructions, warnings, contraindications and precautions.




3. Position skin anchors

Space skin anchors a maximum of 2 cm apart from each other. Position the tips of the skin anchors approximately 0.5 - 1 cm from the wound edge. The DERMACLOSE device is intended to be used with three skin anchors per side of wound and one tension controller for each 10 cm of length of wound. Use multiple devices for longer wounds. Pearl: Some surgeons put a strip of loban™ or thin DuoDERM™ along wound edges and place anchors through the loban or DuoDERM.




4. Staple skin anchors in place

Affix each skin anchor with a minimum of two skin staples. Press stapler firmly against each anchor and staple between each of the staple guide tabs found on the skin anchors.




5. Protect the wound bed

Protect the wound bed with petrolatum impregnated or similar non-adhering wound dressing, such as prior to applying the tension line to the anchors. The Xeroform or Adaptic can be extended under the elevated wound margins to prevent tissue adherence during the tissue expansion process.




6. Lace the line around the skin anchors

Seat the distal end of the bridge tubing on the center anchor by firmly pressing the lumen of the tubing into the top of the skin anchor tab.



When lacing the DERMACLOSE device think... very manageable wound


V
VERY




Using both hands, grasp the tension line near the front of the tension controller and with the first movement form the letter **V** by going around the tabs of the top two outer anchors.

M
MANAGEABLE


With the next movement form the letter **M** by guiding the tension line around the tabs of the opposing two outer anchors.



W
WOUND



Holding the tension line with one hand, guide line over the final anchor tab opposite the tension controller, completing the letter **W**. Gently pull back on the tension controller to remove any slack in the line.



7. Wind tension controller

Apply tension to line by turning tension controller knob clockwise (approximately 22 rotations) until multiple audible clicking sounds are heard. A built-in clutch automatically prevents over-tensioning. At this point the spring motor is fully engaged and no additional tightening is required.



8. Push in the lock button

Lock the device to prevent accidental tension line release by pushing in the locking button at the rear of the tension controller.

9. Post op follow-up

Evaluate tissue movement after 48 to 72 hours. Typically wounds are re-approximated to within 1 cm in 3 to 7 days and can then be sutured or stapled closed. Use standard wound therapy.

Adaptic and Xeroform are registered trademarks of their respective owners.

Dermaclose and GEM design are trademarks of Baxter International Inc., or its subsidiaries.

PHONE: 800-510-3318 | FAX: 205-941-1522 | EMAIL: smca_information@baxter.com | 439 Industrial Lane, Birmingham, AL 35211 | www.synovismicro.com

Synovis Micro Companies Alliance, Inc. (a subsidiary of Baxter International Inc.)

MS-14270/A