

Details of XRR/XLayers

Function Category: XRR

Function: XLayers

Written by: Mrinal Bera (mrinalkb@cars.uchicago.edu) & Wei Bu (bu@cars.uchicago.edu)

Details Last Updated: June 15, 2021

Description

Calculates X-ray reflectivity from a system of multiple layers using Parratt formalism. Here are the description about all the parameters:

Fixed Parameters

Parameters	Units	Description	Default values
x	\AA^{-1}	Array of wave-vector transfer along z-direction i.e Q_z	
E	keV	Energy of the X-rays in keV (optional)	10.0
Minstep	\AA	The thickness of each of the layers in	0.5
rrf		'True' for Fresnel normalized reflectivity (R/Rf) and 'False' for just reflectivity	'True'
qoff	\AA	Q_z offset to correct the $Q_z=0$ of the instrument (zero angle correction)	0.0
yscale		A scale factor for R or R/Rf	1.0
Bkg		In-coherent background	0.0

Single Fitting Parameters

Parameters	Units	Description	Default values
qoff	\AA	Q_z offset to correct the $Q_z=0$ of the instrument (zero angle correction)	0.0
yscale		A scale factor for R or R/Rf	1.0
Bkg		In-coherent background	0.0

Multiple Fitting Parameters

Parameters	Units	Description	Default
------------	-------	-------------	---------

			values
Layers		Layer description	['top', 'bottom']
d	Å	Thicknesses of each of the layers	[0.0,1.0]
rho	el/Å ³	Electron density of each of the layers	[0.0,0.33]
beta		Absorption coefficient of each of the layers	[0.0,0.0]
Sig	Å	Roughness of interfaces between the layers	[0.0,3.0]