



Science and  
Technology  
Facilities Council



National Centre for  
Atmospheric Science  
NATURAL ENVIRONMENT RESEARCH COUNCIL

## AMOF CHILBOLTON ATMOSPHERIC OBSERVATORY CALIBRATION CERTIFICATE

|  |                         |
|--|-------------------------|
| Instrument type                        | Drop-counting raingauge |
| Serial number                          | BN-010 rg001dc_ch       |
| Collector diameter (mm)                | 150                     |
| Calibration (mm rain per drop counted) | 0.00331                 |
| Calibration range (mm/hour)            | 0 – 70                  |
| Calibration date                       | 22 June 2020            |
| Calibrated by                          | Judith Jeffery          |