



Science and
Technology
Facilities Council



National Centre for
Atmospheric Science
NATURAL ENVIRONMENT RESEARCH COUNCIL

AMOF CHILBOLTON ATMOSPHERIC OBSERVATORY CALIBRATION CERTIFICATE

Instrument type	Drop-counting raingauge
Serial number	BW-005 rg006dc_ch
Collector diameter (mm)	150
Calibration (mm rain per drop counted)	0.00329
Calibration range (mm/hour)	0 – 70
Calibration date	22 June 2020
Calibrated by	Judith Jeffery