



Science and  
Technology  
Facilities Council



National Centre for  
Atmospheric Science  
NATURAL ENVIRONMENT RESEARCH COUNCIL

## AMOF CHILBOLTON ATMOSPHERIC OBSERVATORY CALIBRATION CERTIFICATE

Instrument type	Drop-counting raingauge
Serial number	BX-015 rg008dc_ch
Collector diameter (mm)	200
Calibration (mm rain per drop counted)	0.00186
Calibration range (mm/hour)	0 – 40
Calibration date	23 June 2020
Calibrated by	Judith Jeffery