CERTIFICATE OF CONFORMITY



MONITORING THE WORLD AROUND US SINCE 1864

Munro Instruments Ltd, Gilbert House, 406 Roding Lane South, Woodford Green, Essex IG8 8EY Telephone: 020 8551 7000 Fax: 020 8551 4565 Email: info@munroinstruments.co.uk

STFC To.	Purchase Order No	4070137342			
C/O UK SHARED BUISNESS SERVICES LTD	Works Order No	4221			
POLARIS HOUSE	Certificate No. 15968				
NORTH STAR AVENUE					
SWINDON SN21UH					

	Description and Part No.		Quantity	References	Batch No. and Test Report No.
SHA	ALLOW COVER MODEL		1	R100ST	4221
TIPS	VOLUME OF WATER	EQUAL DISCHARGE	-		
250	1620 ml	50/50	17.00		
	SHA CAI TIPS	TIPPING BUCKET RAIN GAUG SHALLOW COVER MODEL CALIBRATED FOR 0.2mm PER TIPS VOLUME OF WATER	TIPPING BUCKET RAIN GAUGE SHALLOW COVER MODEL CALIBRATED FOR 0.2mm PER TIP TIPS VOLUME OF WATER EQUAL DISCHARGE	TIPPING BUCKET RAIN GAUGE SHALLOW COVER MODEL CALIBRATED FOR 0.2mm PER TIP TIPS VOLUME OF WATER EQUAL DISCHARGE	TIPPING BUCKET RAIN GAUGE SHALLOW COVER MODEL CALIBRATED FOR 0.2mm PER TIP TIPS VOLUME OF WATER EQUAL DISCHARGE

REMARKS

CERTIFIED THAT THE ABOVE UNIT OPERATES WITHIN THE LIMITS AS REQUIRED BY THE BRITISH METEOROLOGICAL OFFICE

inspecte respects control a	that the whole o d, tested and unle with the requirer grrangements add	ess otherwise s ments of the co opted in respec	stated above, contract or order. tof these supplies	onform in The quali iee heve	all
accorded	d with the conditi	ions of our Qua	lity Managepre	ng Byshalan,	
	2	5	NUNRO	2	ST S
Signed			1/2		× //
	For and on beha	alf of MUNRO II	NSTRUMENTS	WSPECT!	27
	16	/02/2016			

16/02/2016