NeedleFinder 2026: Sub-Millimeter, Sub-Second Catheter Reconstruction on MRI for Interstitial Gynecologic Brachytherapy

Guillaume Pernelle, Paolo Zaffino, Lauren Barber, Alireza Mehrtash, Christiane Yeung, Tina Kapur, and Collaborators

October 2026 (Target Journal: Physics in Medicine & Biology*)*

ABSTRACT

Objective. Interstitial gynecologic brachytherapy planned on MRI alone requires locating 15 to 30 thin plastic catheters on 3T T2-weighted images. Tracing them by hand takes tens of minutes while the patient is under anesthesia. Automatic methods based on voxel segmentation break catheters into fragments and mistake blood vessels for catheters. We set out to make catheter reconstruction automatic, accurate to below a millimeter, and fast enough to run during the procedure.

Approach. NeedleFinder 2026 works in three stages. Stage 0 is an analytical filter that responds only to small, round, dark spots. Stage 1 is a 3D U-Net that predicts catheter centerlines, directions, and the vaginal obturator. Stage 2 traces candidate trajectories and uses two graph neural networks: one scores how slice detections link into 3D trajectories, and one judges whether the whole set of trajectories forms a plausible implant. Stage 2 is pretrained on 300 simulated implant bundles (SynthBrachy3T). We evaluated the method on 54 clinical 3T MRI scans containing 1,063 expert-traced catheters with patient-level 5-fold cross-validation, against re-implementations of three published methods, against a second team of human observers on 40 scans, and on 48 public needle-free pelvic MRI scans.

Main results. NeedleFinder 2026 found 91.3% of catheters with 85.3% precision (F1 88.2%), compared with 81.3% F1 for our 2019 3D U-Net and 53.7% for a 2026 structure-aware tracker ($p < 10^{-4}$). The median shaft error was 0.46 mm and the median tip error 2.61 mm. The method agreed with the expert reference more closely (0.56 mm mean shaft discrepancy) than two human teams agreed with each other (0.72 mm). With a single annotated patient, synthetic pre-training reached 82.0% F1. On needle-free scans the method produced 0.38 false catheters per scan. Full-volume reconstruction takes 2.74 s on an 8-core CPU (1.20 s on standard 320 × 256 scans, 3.47 s on high-resolution 640 × 416 / 768 × 768 scans) or 1.25 s on a GPU.

Significance. Automatic catheter reconstruction on MRI now matches expert agreement, needs very little training data, and runs in 1 to 3 seconds per patient. It turns a manual task of tens of minutes into a short verification step.

PLAIN-LANGUAGE SUMMARY

Some cancers of the cervix, vagina, and uterus are treated with brachytherapy: a tiny radioactive source is moved, for a few minutes, through thin plastic tubes placed inside the tumor. To aim the radiation, the treatment team must know exactly where each tube is. They find out with an MRI scan taken while the patient is still under anesthesia, and then trace every tube on the images by hand. With 15 to 30 tubes per patient this takes tens of minutes, and a position error of one or two millimeters can change the dose to the tumor or to nearby organs by about ten percent.

On an MRI scan a tube does not glow. It is a small dark dot, a few pixels wide, on each image slice. Blood vessels, air pockets, and the edges of muscles are also dark, so a computer can easily confuse them with tubes.

This paper presents NeedleFinder 2026, software that finds the tubes automatically. It works in three steps. First, a simple filter keeps only dark spots that are dark all the way around, the way a tube cross-section looks, and discards one-sided edges. Second, a neural network looks at the 3D scan and marks the center of each tube on each slice. Third, two "graph" networks connect the marks from slice to slice into 3D curves, pull apart tubes that touch, and then look at the whole bundle together to throw out curves that do not fit a real implant.

We tested it on 54 patients with 1,063 tubes. It found 91 percent of them and was right 85 percent of the time when it reported one. Its traced positions were, on average, half a millimeter from the expert's, which is closer than two experts are to each other. It runs in under three seconds on an ordinary computer (about one second on standard-resolution scans) and needs no graphics card. In practice, the physicist clicks one button, checks the result, and fixes the rare miss, instead of tracing everything from scratch.

1 INTRODUCTION

1.1 *The Clinical Problem*

Brachytherapy delivers radiation from inside the body. For locally advanced cervical, vaginal, and recurrent endometrial cancers it is the cornerstone of curative radiotherapy [1,2]. The EMBRACE-I study reported 92% local control at five years with a median dose of 90 Gy (EQD2) to the high-risk target volume [2], and EMBRACE-II now prescribes planning aims of 90 to 95 Gy while sparing the bladder, rectum, and sigmoid colon [1]. Reaching such doses in large or irregular tumors requires *interstitial* implants: in addition to a central tandem and a vaginal obturator or ring, 10 to 30 thin plastic catheters or needles are inserted into the tumor through a perineal template or freehand [3–5].

The tumor itself is best seen on T2-weighted (T2W) MRI. Historically, clinics took both an MRI and a CT scan and registered the two, which adds its own uncertainty to the reconstruction [6]. The AAPM Task Group 303 report describes

MRI-based workflows and asks for applicator reconstruction accuracy of 2 mm or better [7], the Medical Physics Practice Guideline 13.a lists MRI among the accepted planning modalities [8], and MRI-only interstitial workflows have been validated clinically [9,10].

Finding the catheters on MRI is the hardest step of an MRI-only workflow [4,11]. The dose from the ^{192}Ir source falls steeply with distance: organ-at-risk dose parameters change by up to 5 to 6% per millimeter of applicator displacement [12], a 1.5 mm shift can change the rectal D2cc by 10% [13], and in MRI-only cervical implants a systematic 1 mm error along the source path already changes target and organ doses significantly [14]. Swapping two catheters is worse still [15]. Yet a 2 mm plastic catheter contains no mobile water protons and shows up on T2W MRI only as a small dark hole, a few pixels across, surrounded by anatomy full of other dark structures: blood vessels, air, calcifications, and the fascial planes between muscles. Tracing 15 to 30 catheters slice by slice therefore takes tens of minutes, 46 minutes on average in one MRI-only gynecologic series [16], while the patient waits under anesthesia or sedation.




FIGURE 1 **Overview of NeedleFinder 2026.** The input is a clinical 3T T2W pelvic MRI (Case 017: 26 reference trajectories, including the obturator channels). Stage 0 computes an analytical dark-core symmetry map S_0 in native Rust (0.41 s cohort mean). Stage 1 applies a 3D residual U-Net with implant-region gating to predict centerline peaks, local directions, and the obturator (0.96 s on CPU, 0.22 s on GPU). Stage 2 links slice detections into 3D trajectories with native C++17/Rust seed tracking, a triplet message-passing network, and adhesion-aware duplicate merging; the last tile shows the out-of-fold predictions (vermilion) over the expert reference (grey) (1.22 s on CPU, 0.54 s on GPU), then prunes false trajectories with a bundle graph transformer and smooths each curve with a cubic polynomial (0.15 s on CPU, 0.08 s on GPU). The full patient volume is processed end-to-end in 2.74 s on an 8-core CPU (1.20 s on standard scans, 3.47 s on high-resolution scans) or 1.25 s on a GPU; the bar shows the stage breakdown.

1.2 *Key Terms*

Readers outside medical physics may find the following definitions useful.

- **Catheter, needle, shaft, tip.** We use *catheter* and *needle* interchangeably for the thin plastic tubes. The *shaft* is the length of the tube visible on the scan; the *tip* is its deepest end.
- **Digitization or reconstruction.** Tracing the 3D path of each catheter on the images so the planning software knows where the source will travel.
- **Slice, voxel, in-plane.** An MRI volume is a stack of 2D images (slices). Each 3D picture element is a voxel. *In-plane* means within one slice; here the in-plane pixel is 0.31 to 0.88 mm while slices are 1.2 to 6.0 mm thick.
- **Recall, precision, F1.** Recall is the fraction of true catheters the method found. Precision is the fraction of reported catheters that were real. F1 is their harmonic mean, a single score that penalizes both misses and false alarms.
- **Shaft error and tip error.** The distance, in millimeters, between the automatic trajectory and the expert's trajectory, measured in-plane along the shaft and in 3D at the tip.
- **Cross-validation, out-of-fold.** The 54 patients were split into 5 groups. Each group was tested by a model trained on the other four, so no patient was ever tested by a model that had seen them. All reported results are *out-of-fold* in this sense.
- **Inter-observer variability.** How much two expert teams disagree when they trace the same scan independently. It is the practical floor for how accurately any method can be measured.

1.3 *Previous Work*

Automatic catheter digitization has gone through three generations. Two recent systematic reviews cover the field [11,17].

Model-based tracking (2013 to 2019). Our first work on the Brigham and Women's Hospital (BWH) cohort seeded each catheter with one click at its tip, searched slice by slice inside an expanding cone, and fitted a Bézier curve; 93 of 101 catheters in 10 patients were traced correctly at 0.3 s per catheter [18]. Mastmeyer *et al.* [3] replaced the curve search with a mechanical rod-and-spring catheter model, reaching 1.49 mm mean Hausdorff distance on 760 catheters in 54 patients, still from a manual tip click. Related model-to-image registration reconstructs rigid applicators on MRI within inter-observer variability [19], and line fitting by RANSAC digitizes straight needles on CT [20]. These methods are interpretable but need manual initialization and fail when catheters bend sharply or cross dark anatomy.

Voxel segmentation with heuristic post-processing (2019 to 2026). To remove manual initialization, Zaffino *et al.* [4] trained an ensemble of 3D U-Nets to segment catheter voxels on 50 clinical T2W scans with 826 catheters, then clustered the voxels with DBSCAN and removed small islands. It reached a voxel Dice of 0.60, a mean distance error of 2.0 mm, 6.7% false-positive and 1.5% false-negative clusters, in 2 to 9 minutes per patient. Convolutional segmentation, often nnU-Net based

[21], has since been applied to CT brachytherapy of the cervix, prostate, and breast [22–29], to MRI-guided prostate implants [30–32], to interventional MRI [33], to transrectal ultrasound [34–38], and to gynecologic MRI with line markers [16] or with applicator mesh alignment [39]. Even with topology-preserving losses such as clDice [40], pure voxel segmentation has two failure modes on T2W MRI: a one-slice signal dropout splits one catheter into two to four fragments, and dark vessels and tendons are segmented as short false catheters. The ultrasound literature documents the same pattern: accuracy collapses on artifact-laden catheters [37] and tips are much harder than shafts [36]. Electromagnetic tracking avoids the imaging problem altogether but requires dedicated hardware in the operating room [41–44].

Structure-aware tracking and graph networks (2022 to 2026). On CT, Xiao *et al.* [45] identified two further problems: *catheter adhesion*, where shafts touch or overlap on consecutive slices so that greedy trackers swap or merge them, and the scarcity of annotated clinical data. They proposed an implant-region gating module (IRGM) that suppresses features outside the implant, a direction-constrained 3D Hough transform for initialization, synchronous cost-based inward tracking across slices, cubic polynomial smoothing, and synthetic-to-real training, on 38 patients and 203 fractions. Their companion work handles needle adhesion in dense CT seed implants by detecting tips and handles [46]. Outside brachytherapy, message-passing networks have replaced greedy linking in tracking-by-assignment [47], and synthetic-to-real training has been shown to transfer to real needle detection [48]. Applied to MRI masks, however, a mask-driven tracker still turns vessels inside the implant region into false catheters, and nothing in it asks whether the full set of trajectories makes sense as one implant.

Open data. No public needle or catheter dataset for brachytherapy exists [49], and the AI literature in brachytherapy remains largely retrospective and single-center [17]. The public pelvic MRI scans we use for the false-alarm test come from the Medical Segmentation Decathlon [50].

1.4 *Contributions*

This paper combines the strengths of all three generations on the full 54-patient, 1,063-catheter BWH cohort, to our knowledge the largest annotated MRI cohort for interstitial gynecologic brachytherapy.

1. **A three-stage architecture.** An analytical dark-core symmetry prior (Stage 0), a 3D residual U-Net with five output heads and implant-region gating (Stage 1), and two graph networks (Stage 2): native seed tracking with a triplet message-passing network and adhesion-aware duplicate merging that keeps touching catheters apart, and a tip-aware bundle graph transformer that judges all trajectories together and refines their tips. Stage 2 is pre-trained on SynthBrachy3T, a simulator of candidate-trajectory bundles.
2. **A controlled comparison.** We re-implemented three published approaches and evaluated all methods on identical patient-level 5-fold splits, on the whole cohort and on the 23 patients with severe catheter adhesion.

3. **Three clinically relevant checks.** Agreement with experts compared with agreement between experts on 40 doubly traced scans; accuracy as a function of how many annotated patients were available for training, from zero to all; and false alarms on 48 public pelvic MRI scans containing no catheters.

2 MATERIALS AND METHODS

2.1 *Patient Cohorts and Public Benchmark*

Table 1 summarizes the data. The primary cohort is 54 intraoperative 3T T2W turbo spin-echo MRI scans acquired during MRI-guided interstitial gynecologic brachytherapy in the AMIGO suite at Brigham and Women's Hospital (Harvard Medical School, Boston, MA, USA) with Institutional Review Board approval [3,4,18]. It consolidates the patients of our 2013 and 2019 studies, of which 48 scans could be recovered with their reference traces, and adds 6 new implants. Expert medical physicists traced 1,063 reference trajectories (9 to 42 per patient): 759 full-depth interstitial catheters from the primary annotation set, which we call *long shafts*, and 304 short obturator and partial-field-of-view tracks from the secondary set. In-plane pixel size ranges from 0.31 to 0.88 mm and slice thickness from 1.2 to 6.0 mm.

Xiao *et al.* [45] analyzed separately the fractions in which catheters touch or overlap. We define a **severe-adhesion subgroup** in the same spirit: the 23 of 54 patients in which at least 4 long, or at least 6 of all, reference catheters pass within 3.0 mm of another catheter on shared axial slices. All 571 catheters of these patients are evaluated; 191 of them (166 long) lie within 3.0 mm of a neighbor, and 233 do across the whole cohort.

TABLE 1 **Clinical imaging cohorts and evaluation protocols.** Primary evaluation uses 5-fold patient-level out-of-fold (OOF) cross-validation on 54 clinical 3T T2W MRI datasets from Brigham and Women's Hospital (BWH AMIGO).

Cohort / Subset	Source	Pts	Cath.	XY (mm)	Δz (mm)	Evaluation role
BWH GYN full cohort[a]	BWH (3T T2W)	54	1063	0.31–0.88	1.2–6.0	5-fold OOF primary benchmark
Long shafts (primary set)	BWH (3T T2W)	54	759	0.31–0.88	1.2–6.0	Full-depth active dwell shafts
Short / obturator (secondary set)	BWH (3T T2W)	54	304	0.31–0.88	1.2–6.0	Partial-FOV & template references
Adhesion subgroup[b]	BWH (3T T2W)	23	571	0.31–0.88	1.2–6.0	Touching / crossing shaft stress test
Dual-observer subset[c]	BWH (3T T2W)	40	505	0.31–0.88	1.2–6.0	Inter-observer floor (`CY` vs. `LB/AMe`)
MSD Task05_Prostate[d]	Radboudumc	48	0	0.60–0.75	3.0–4.0	Out-of-distribution 0-needle specificity

[a] Consolidated superset of the cohorts of Pernelle *et al.* (2013, 10 patients) and Zaffino *et al.* (2019, 50 patients), extended with new implants; 9–42 reference trajectories per patient. Spacings are the ranges of the 54 image headers.

[b] Severe catheter-adhesion subgroup: the 23/54 patients in which $\geq$ 4 long or $\geq$ 6 total reference catheters pass within $\leq$ 3.0 mm of another catheter on shared axial slices. All 571 catheters of these patients are evaluated; 191 of them (166 long) lie within 3.0 mm of a neighbour (233 across the whole cohort).

[c] Forty clinical scenes independently contoured by two expert medical physics teams (`CY`: 544 trajectories on disk, 505 active in the planning scenes; `LB/AMe`: 513 trajectories; 481 Hungarian 1-to-1 matched pairs within 3.5 mm).

[d] Public CC-BY-SA 4.0 pelvic axial T2W MRI benchmark containing prostate, rectum, bladder, obturator internus, and dense periprostatic vasculature with zero implanted catheters.

Two further datasets serve as checks. Forty scans were traced independently by two expert teams (CY, 505 trajectories; LB/AMe, 513 trajectories); the 464 pairs matched one-to-one within 3.5 mm define the human inter-observer floor. To measure false alarms on anatomy that contains no catheters at all, we use all 48 pelvic T2W scans of the public Medical Segmentation Decathlon prostate task [50] (MSD Task05_Prostate; 0.60 to 0.75 mm in-plane, 3 to 4 mm slices).

2.2 *Stage 0: Analytical Dark-Core Symmetry Prior*

The physical idea is simple. On an axial slice a catheter cross-section is a dark core that is surrounded by brighter tissue in *every* direction, whereas a fascial plane or a muscle border is dark on one side only. Stage 0 turns this into a parameter-free score $S_0(x, y, z)$, computed slice by slice in multi-threaded native Rust.

For each voxel $\mathbf{p}$ on slice z we sample six diameters through $\mathbf{p}$ at angles $\theta_d = d\pi/6$, $d \in \{0, \dots, 5\}$, and four physical radii $r \in \mathcal{R} = \{1.35, 1.85, 2.5, 3.3\}$ mm, converting each radius to pixels with the in-plane spacing. For diameter d and radius r the darker of the two ring samples is kept,

$$m_{d,r} = \min\big(I(\mathbf{p} + r\mathbf{u}_d),\ I(\mathbf{p} - r\mathbf{u}_d)\big),$$

so that a one-sided edge, where one end of the diameter stays dark, produces a small $m_{d,r}$. The six values are sorted, $m_{(1),r} \le \dots \le m_{(6),r}$, and the ring brightness is a robust weighted order statistic that still tolerates one dark neighbor,

$$S_{\text{ring}}(\mathbf{p}) = \max_{r \in \mathcal{R}} \left[0.45\, m_{(2),r} + 0.55\, m_{(3),r}\right].$$

The response S_0 grows with the positive contrast between this ring brightness and the 3 × 3 smoothed center intensity and with the relative darkness of the core (the exact expression, a product of the absolute contrast, the relative contrast, and a squared dark-core ratio, is in the released source), and is normalized per volume. Candidate centers are the local maxima of S_0 after non-maximum suppression. Stage 0 takes 0.41 s per full volume on average on an 8-core CPU (0.12 s on standard 320 × 256 scans and 0.59 s on high-resolution 640 × 416 / 768 × 768 scans); the exact weights are in the released source code.

2.3 *Stage 1: Multi-Head 3D U-Net with Implant-Region Gating*

Resampling the scan to 1 mm isotropic voxels, as most pipelines do, blurs a 0.3 mm in-plane catheter hole by a factor of three and destroys its boundary. Working on the native grid instead, a standard 3×3×3 convolution mixes information over up to ten times more distance along z than in-plane. Stage 1 (NeedleUNet3D, Figure 2a) therefore operates on the native grid, starts with a 1×3×3 in-plane residual block, and continues with 3×3×3 residual blocks with squeeze-and-excitation attention (24, 48, 96, and 144 channels). Its input has six channels: the normalized intensity, the Stage 0 dark-core response and ring-symmetry contrast, the smoothed intensity, and the normalized slice position and in-plane distance from the field-of-view center.

The network predicts four outputs:




FIGURE 2 **Stage 1 and Stage 2 in detail. (a)** Stage 1 takes the normalized T2W volume and the Stage 0 maps into an anisotropic 3D residual U-Net with squeeze-and-excitation blocks, which predicts five outputs: tubular mask, centerline peak, two direction components, and obturator. **(b)** Stage 2 traces tracks from the seeds in native code, scores their nodes and edges with a 4-layer triplet message-passing network, then scores every trajectory in the context of the whole bundle with a graph transformer (two edge-biased attention layers and one global self-attention layer) and smooths it with a cubic polynomial.

1. **Tubular mask** P_{mask}: binary cross-entropy with a positive weight of 12 plus soft Dice on the tubular label around each reference trajectory.
2. **Centerline peak** P_{peak}: a sharp 2D Gaussian ($\sigma = 0.70$ mm) at the sub-pixel intersection of each catheter with the slice, trained with a mean-squared-error loss up-weighted 25-fold at the peaks. Two touching catheters whose masks merge still produce two distinct peaks.
3. **Direction field** $(V_x, V_y) = (\Delta x/\Delta z, \Delta y/\Delta z)$: the local slope of the trajectory, trained with a mean-squared-error loss inside the mask.
4. **Obturator channel** P_{obt}: a dedicated segmentation of the large central obturator and tandem, so that its lumen and rim do not spawn false needles.

As an ablation we also evaluate AxialTransUNet3D, which adds multi-head self-attention layers along the slice axis at the deepest encoder level and the bottleneck.

2.4 *Stage 2: From Slice Detections to a Coherent 3D Bundle*

Stage 2 answers two questions: which detections on neighboring slices belong to the same catheter, and which of the resulting 3D curves are real.

2.4.1 *Stage 2A: Native Seed Tracking and Triplet Message Passing.*

On each slice, local maxima of P_{peak} and of S_0 become nodes with a 16-dimensional feature vector (normalized position, P_{mask}, P_{peak}, S_0, local 3×3 intensity statistics, predicted slope, and obturator proximity). Directed edges connect nodes across 1 to $\max(4, 24\ \text{mm}/\Delta z)$ slices. An edge is kept when its chord slope is at most 0.56 and its in-plane displacement lies within $2.8 + 0.14\,\Delta z$ mm of the position predicted by the Stage 1 direction

field, or within $3.5 + 0.22\,\Delta z$ mm outright. Each edge carries 12 features (displacement, chord slope, and agreement with the predicted direction).

A pairwise graph network cannot penalize a kink, because a kink involves three nodes. NeedleSliceMPNN is a 4-layer triplet-collinearity message-passing network (hidden size 64). Each layer updates an edge from its own features, the embeddings of its two nodes, and attention-aggregated messages from the edges entering and leaving each of those nodes, so that the update sees the incoming and outgoing chords together and can discourage a kink. The network outputs refined sub-pixel offsets $(\delta x_i, \delta y_i)$ and node and edge probabilities.

Candidate 3D tracks themselves are traced in native code: from each Stage 0 and Stage 1 seed, a bidirectional Kalman tracker follows the catheter slice by slice in both directions and trims it from the inside out where the evidence ends. Tracks that run within 1.65 to 2.05 mm of each other, depending on their confidence, are merged as duplicates; the margin is chosen so that genuinely adhering twins survive. The surviving tracks, scored by the message-passing network, are passed to Stage 2B.

2.4.2 *Stage 2B: Bundle Graph Transformer, Synthetic Pre-Training, and Smoothing.* After Stage 1, recall is above 92% (96% on long shafts), but each patient still carries about ten false tracks along vessels and obturator edges, and needle-free pelvic scans produce about twenty; Stage 2A removes roughly half of them. Real catheters have a property these false tracks lack: they form a *bundle*. They enter through a common template or corridor, run roughly parallel to the obturator axis, keep at least 2 mm apart, and their tips form a coherent surface over the target.

Each candidate trajectory becomes a 26-dimensional token (length, linear-fit residual, 75th-percentile S_0 and P_{mask}, P_{peak}, P_{obt}, the three message-passing probabilities, agreement between models, the deviation of its entry position and slope from the bundle median, and the contrast gradient at its cranial end), together with 8 pairwise features between trajectories (minimum 3D clearance, mean in-plane separation, angular divergence, z-overlap, and duplicate indicators). The TipAwareBundleGraphTransformer applies two edge-biased attention layers over neighboring trajectories and one global self-attention layer over the whole bundle. It outputs three logits (a recall-oriented score, a strict one-to-one score, and a long-shaft one-to-one score), combined with weights 0.45 and 0.55 into a confidence p_m, and one cranial tip offset Δz_m that trims or extends the deep end. An implant-region gate inspired by the IRGM of Xiao *et al.* [45] is applied at this point: whereas they train a separate region branch, ours is geometric. The implant center is the median in-plane position of the candidates with confidence of at least 55%, the implant radius is the 98th percentile of their distances from that center plus 16 mm, and a candidate outside this radius is kept only if its confidence exceeds the threshold by more than 22 points. Trajectories with $p_m \geq \tau$ (default $\tau = 50\%$) are kept, their cranial tips refined with the predicted offset and the local contrast drop-off, and each is smoothed with a cubic polynomial in z (Poly3) to remove the voxel staircase across thick slices.

Annotated clinical scans are scarce, and rare configurations such as crossing catheters are rarer still. We therefore built **SynthBrachy3T**, a generator of synthetic candidate-trajectory bundles with the same 26 node and 8 edge features as real candidates. Each of the 300 bundles contains 12 to 23 true long shafts arranged around

a common axis, of which 30% receive an adhering twin 1.65 to 2.75 mm away, 2 to 8 short obturator tracks, 3 to 11 split duplicates of true shafts, and 4 to 14 vessel-like false tracks, with simulated tip-tracking noise. SynthBrachy3T does not simulate MRI voxels; it supplies the bundle-level reasoning task with far more adhesion and tip configurations than any clinical dataset contains. Stage 2B is pre-trained on these bundles, then fine-tuned on the clinical bundles of the fold's training patients together with five synthetically augmented copies of each, and its scores are blended with those of a transformer trained on clinical bundles only (weight 0.55 on the synthetic model).

3 EXPERIMENTAL SETUP

3.1 *Cross-Validation Protocol and Baselines*

Every learned component and every baseline is trained and evaluated with patient-level 5-fold cross-validation on the 54 patients. No slice, patch, or trajectory from a test patient is used for training or model selection in that fold. We re-implemented three published approaches on the same splits:

- **PMB 2019: 3D U-Net + DBSCAN** [4]. Volumes resampled to 1 mm isotropic, segmented by a 3D U-Net, then grouped by principal-axis estimation, DBSCAN clustering, and spline fitting.
- **ResUNet + clDice + skeleton** [21,40]. A native-resolution 3D residual U-Net trained with binary cross-entropy, soft Dice, and the centerline-Dice loss, followed by morphological skeletonization, connected components, and polynomial centerline fitting.
- **Xiao 2026 Sync-Track** [45]. Implant-region gating, direction-constrained 3D Hough and PCA initialization, synchronous inward tracking, and cubic smoothing. As in their paper, we run it both on our out-of-fold segmentation masks and on the reference masks (an oracle upper bound).

3.2 *Metrics and Statistics*

Predicted and reference trajectories are matched one-to-one with the Hungarian algorithm (SciPy `linear_sum_assignment`) using a 3.5 mm in-plane cutoff over the slices they share. We report:

1. **Detection**: recall, precision, and F1 under this strict one-to-one matching.
2. **Shaft error** E_{shaft}: the in-plane distance between the two curves at 40 evenly spaced positions along their shared length (pairs with a mean distance above 8 mm are treated as unmatched), summarized by mean, standard deviation, median, interquartile range, and 95th percentile.
3. **3D curve distance** D_{avg} and **95th-percentile Hausdorff distance** HD_{95}, computed symmetrically on 50 resampled points per curve.
4. **Tip error** E_{tip}: the 3D distance between the predicted and reference deepest (cranial) tip.

Differences between methods are tested over the 54 patients with paired two-sided Wilcoxon signed-rank tests, and 95% confidence intervals come from patient-level bootstrap resamples (1,000 for the headline rows; 200 to 500 for the threshold sweep, subgroup, and few-shot rows).

3.3 *Computing Environment, Software, and Code*

Training and the 5-fold benchmarks ran on one Linux workstation (Arch Linux) with a 24-core AMD Ryzen Threadripper PRO 7965WX, 256 GB of RAM, and two NVIDIA GPUs (GeForce RTX 4090, 24 GB, and RTX 4070 Ti SUPER, 16 GB), The benchmark, graph-network training, and few-shot experiments ran with Python 3.13.2, PyTorch 2.11.0 (CUDA 13.0), SciPy 1.16.3, and NumPy 2.4.3; the Stage 1 networks were trained on both GPUs in parallel and exported to ONNX in a separate environment with Python 3.12.9, PyTorch 2.4.1 (CUDA 12.1), SciPy 1.16.1, and ONNX 1.23.2, the PyTorch version being the producer recorded in the model files. The Stage 0 kernel is written in Rust (crate `needlefinder_rust`, multi-threaded with the standard library, exposed to Python through a C interface and `ctypes`, compiled in release mode with link-time optimization). Hungarian matching uses SciPy. Figures were produced with Matplotlib.

For clinical use the trained Stage 1 and Stage 2B networks are exported to ONNX and run with ONNX Runtime inside an extension for 3D Slicer 5.10 [51], the open-source platform in which the 2013 and 2019 NeedleFinder modules were also built [4,18]; the Stage 2A weights are stored as NumPy archives. No GPU is required. The runtimes reported in the Results were measured on an 8-core desktop CPU and, for the GPU figures, on the RTX 4090.

The 3D Slicer extension, the Rust kernel, and the exported models are available at https://github.com/needlefinder/NeedleFinder. The training scripts (Stage 1 U-Nets, graph networks, and all baselines), the benchmark and statistics code, the SynthBrachy3T generator, the prediction export used for Table 3 and the qualitative figures, and the figure and table generators are collected in the repository's `vault/paper/benchmarks` directory and will be made public on publication. See also the Data and Code Availability statements at the end of the paper.

4 RESULTS

4.1 *Accuracy Against Published Methods*

Table 2 is the main result. Part A compares all methods on the whole cohort and on the severe-adhesion subgroup; Part B reports ablations on the same folds. Figure 3 shows the distribution of per-catheter errors.

Four findings stand out.

Voxel segmentation alone over-fragments. The ResUNet + clDice baseline finds almost every catheter (90.4% recall) and is accurate where it matches (0.43 mm median shaft error), but it reports 4,015 fragments across the 54 volumes. Its

TABLE 2 **Catheter reconstruction accuracy on 54 BWH patients (1 063 catheters), 5-fold patient-level out-of-fold.** Part A compares end-to-end methods on the whole cohort and the severe catheter-adhesion subgroup; Part B ablates **NeedleFinder 2026** components, `SynthBrachy3T` synthetic-to-real pre-training + tip-aware refinement, and ground-truth-mask tracking on identical folds. Bold: best; underline: second best within each block. Shaded: **NeedleFinder 2026** variants (pre-specified $\tau = 50\,\%$).

	Detection, 1-to-1 (%)			E_{shaft} (mm)			E_{tip} (mm)	Time (s)
Method	Rec.	Prec.	F1	Mean	Median	P95	Median	CPU
Part A — End-to-end comparison								
Whole cohort (54 patients, 1 063 catheters)								
PMB 2019 3D U-Net[a]	86.9	76.4	81.3	**0.52**	**0.42**	**1.11**	3.35	14.80
ResUNet + clDice + Skel.[b]	90.4	23.9	37.9	0.56	0.43	1.38	3.08	3.85
Xiao 2026 Sync-Track[c]	86.9	38.9	53.7	0.66	0.48	1.83	6.02	3.42
NeedleFinder 2026 + IRGM + Poly3[d]	89.8	82.5	86.0	0.57	0.46	1.29	2.82	**2.68**
NeedleFinder 2026 + SynthBrachy3T + Tip[e]	**91.3**	**85.3**	**88.2**	0.58	0.46	1.32	**2.61**	2.74
Severe catheter-adhesion subgroup (23 patients, 571 catheters)[f]								
PMB 2019 3D U-Net	94.2	79.0	85.9	0.54	0.43	1.22	3.68	–
ResUNet + clDice + Skel.	**94.9**	32.0	47.8	**0.52**	**0.42**	**1.19**	3.37	–
Xiao 2026 Sync-Track	90.5	48.5	63.1	0.64	0.47	1.84	7.25	–
NeedleFinder 2026 + IRGM + Poly3	91.9	82.8	87.1	0.56	0.45	1.25	3.33	–
NeedleFinder 2026 + SynthBrachy3T + Tip	93.4	**86.3**	**89.7**	0.56	0.45	1.27	**2.98**	–
Part B — Controlled ablation & oracle-mask benchmark (whole cohort, identical folds)								
Physics + HybridNeedleNet	76.8	64.0	69.8	0.62	0.48	1.54	3.04	1.77
NeedleUNet3D + tubes	92.8	62.6	74.7	0.58	0.47	1.31	2.91	**1.62**
AxialTransUNet3D + tubes	92.5	60.3	73.0	0.58	0.47	1.30	2.98	1.84
Triplet MPNN only	88.3	77.7	82.7	**0.53**	**0.41**	1.29	3.31	1.95
NeedleFinder 2026 ($\tau = 50\,\%$)	88.0	81.8	84.8	0.57	0.46	1.31	2.72	2.65
NeedleFinder 2026 + IRGM + Poly3 ($\tau = 50\,\%$)	89.8	82.5	86.0	0.57	0.46	1.29	2.82	2.68
NeedleFinder 2026 + SynthBrachy3T ($\tau = 25\,\%$)	**94.7**	76.8	84.8	0.58	0.47	1.32	2.69	2.74
NeedleFinder 2026 + SynthBrachy3T ($\tau = 50\,\%$)	91.3	85.3	88.2	0.58	0.46	1.32	**2.61**	2.74
NeedleFinder 2026 + SynthBrachy3T ($\tau = 55\,\%$)	90.5	**86.4**	**88.4**	0.57	0.46	**1.28**	**2.61**	2.74
Xiao 2026 on GT masks[g]	97.4[r]	83.1[r]	89.7[r]	0.21[r]	0.11[r]	0.86[r]	1.69[r]	0.12[r]

Rec./Prec./F1: strict one-to-one Hungarian matching (≤ 3.5 mm shaft threshold); E_{shaft}: in-plane centerline distance on matched slices; E_{tip}: 3D distal-tip distance. P95: 95th percentile over matched catheters.

[a] Zaffino *et al.* (2019) ($1 \times 1 \times 1$ mm^3 isotropic 3D U-Net + 2D Hough / DBSCAN clustering), retrained on our 5-fold splits.

[b] Native-resolution 3D ResUNet trained with soft-skeleton clDice loss (2021) + 3D morphological skeletonization.

[c] Xiao *et al.* (2026b) Implant-Region Gating Module (IRGM) + direction-constrained 3D Hough + intensity-weighted PCA + synchronous cost-matrix inward tracking + 3rd-order polynomial fitting evaluated on 5-fold OOF masks.

[d] **NeedleFinder 2026** with the geometric implant-region gate, adhesion-aware duplicate merging, and 3rd-order polynomial regularization ($p = 3.97 \times 10^{-10}$ vs. Xiao 2026 on F1; the variant without gate and Poly3 differs from PMB 2019 at $p = 0.0085$).

[e] **NeedleFinder 2026** with Stage 2B pre-trained on `SynthBrachy3T` synthetic bundles, fine-tuned on clinical plus synthetically augmented bundles, and cranial tip refinement ($p = 6.42 \times 10^{-5}$ vs. **NeedleFinder 2026** + IRGM + Poly3; $p = 1.15 \times 10^{-5}$ vs. PMB 2019; $p = 2.28 \times 10^{-10}$ vs. Xiao 2026 on F1; $p = 4.96 \times 10^{-4}$ vs. **NeedleFinder 2026** + IRGM + Poly3 and $p = 1.06 \times 10^{-5}$ vs. PMB 2019 on E_{tip}).

[f] The 23/54 patients in which ≥ 4 long or ≥ 6 total catheters pass within ≤ 3.0 mm of another catheter on shared axial slices (571 catheters evaluated, 191 of them adhering).

[g] Oracle upper-bound check following Xiao *et al.* (2026b) Table 2 Part B (tracker run directly on ground-truth binary masks).




FIGURE 3 **Per-catheter error distributions on the whole cohort** (54 patients; each method's matched catheters). (a) In-plane shaft error and (b) 3D tip error. Each raincloud shows the density (top), the median and interquartile range (white box), the 5th to 95th percentile (capped bar), the mean (white diamond), and individual catheters (points). Numbers at the right give median / 95th percentile in mm. In (a) the grey band and dashed line mark the human inter-observer interquartile range (0.34 to 0.85 mm) and 95th percentile (2.05 mm); in (b) the dashed line marks the human inter-observer median tip discrepancy (3.20 mm).

precision is 23.9% and its F1 37.9%. Signal dropouts along the shaft cut one catheter into several pieces, and dark veins become extra tracks.

Structure-aware tracking needs a clean mask. On the reference masks (Table 2, Part B, last row), the Xiao 2026 tracker achieves 97.4% recall, 89.7% F1, and 0.11 mm median shaft error: given a perfect mask, synchronous tracking separates touching shafts well. On our out-of-fold masks its F1 drops to 53.7% (precision 38.9%), because vessels inside the implant region still become tracks and nothing rejects them. Tracking alone does improve on skeletonization, by 15.9 F1 points on the whole cohort.

NeedleFinder 2026 is the most accurate on every metric. At the pre-specified threshold $\tau = 50\%$ it reaches 91.3% recall, 85.3% precision, and 88.2% F1 (88.4% at $\tau = 55\%$; 94.7% recall at $\tau = 25\%$). The gains over the other methods are all significant: 2.2 F1 points over our own pipeline without synthetic pre-training ($p = 6\times10^{-5}$), 6.9 points over PMB 2019 ($p = 1\times10^{-5}$), 34.5 points over Xiao 2026, and 50.3 points over ResUNet + clDice (both $p < 10^{-9}$). Median tip error falls to 2.61 mm, from 3.35 mm (PMB 2019), 6.02 mm (Xiao 2026), and 2.82 mm (our pipeline without synthetic pre-training; $p = 5\times10^{-4}$). Median 3D curve distance is 1.03 mm and median HD_{95} is 3.44 mm. The median shaft error is 0.46 mm on all matched catheters and 0.45 mm on the 759 long shafts, on which recall is 92.5% at the default threshold and 96.4% at $\tau = 25\%$.

The advantage holds when catheters touch. On the 23 severe-adhesion patients (571 catheters), NeedleFinder 2026 reaches 89.7% F1 (93.4% recall, 86.3% precision) with 0.45 mm median shaft error and 2.98 mm median tip error, compared with 85.9% F1 for PMB 2019, 63.1% for Xiao 2026, and 47.8% for ResUNet + clDice.

4.2 *Contribution of Each Stage and Learning From Few Patients*




FIGURE 4 **(a) Precision versus recall** for each pipeline stage, each baseline, and the NeedleFinder 2026 threshold sweep on the whole cohort. Dashed curves are lines of constant F1; the star marks the default threshold $\tau = 50\%$. **(b) Accuracy versus number of annotated training patients per fold**, $k \in \{0, 1, 5, 10, 20, 43\}$, with and without SynthBrachy3T pre-training. Shaded bands are the patient-level bootstrap 95% confidence intervals of F1.

Figure 4(a) shows what each stage adds. Stage 1 alone has high recall (92.8%) but only 62.6% precision. Stage 2A raises precision by 15.1 points and tightens median shaft error to 0.41 mm. Stage 2B with implant gating and cubic smoothing reaches 86.0% F1. Pre-training Stage 2B on SynthBrachy3T adds a further 2.2 F1 points at the default threshold (2.4 points at $\tau = 55\%$), to 88.2%.

Figure 4(b) asks how many annotated patients are really needed. With no clinical training data at all ($k = 0$), the analytical Stage 0 prior and the synthetically pre-trained Stage 2B already reach 75.1% F1. With a single annotated patient ($k = 1$) the method reaches 82.0% F1, which is 3.9 points better than the same pipeline without synthetic pre-training and better than the PMB 2019 U-Net trained on all 43 patients of a fold (81.3%). Five patients give 83.7% F1, ten give 86.6%, and all 43 give 88.2%.

Figure 5 makes the stages visible on two difficult cases. Stage 0 lights up every round dark core in the pelvis, including vessels. Stage 1 suppresses those outside the implant gate and resolves adjacent catheters into separate peaks even where their masks merge. Stage 2 then produces smooth bundles with no swapped identities.




FIGURE 5 **Two out-of-fold test cases.** (a) Case 017, a dense implant with 26 reference trajectories including the obturator channels. (b) Case 031, a severe-adhesion implant with 23 catheters of which 10 lie within 3 mm of a neighbor. Left to right: the T2W slice with the expert intersections (white circles) and the out-of-fold NeedleFinder 2026 intersections (vermilion crosses); the Stage 0 map S_0; the Stage 1 centerline peaks (blue) with the mask contour (solid) and the obturator gate (dashed orange); and the out-of-fold 3D bundle (vermilion) over the expert reference (grey). Case 017: 27 predicted trajectories for 26 reference trajectories; Case 031: 23 for 23.

4.3 *Agreement With Experts Compared With Agreement Between Experts*

Because a catheter is only a few pixels wide and slices are thick, two experts clicking centerlines never agree perfectly; inter-observer discrepancies of 0.6 to 0.8 mm have been reported for breast and prostate implants [52,53], and algorithms within that range are considered clinically equivalent to a second observer [19,35]. Table 3 measures this on the 40 doubly traced scans and compares it with the algorithm's agreement with each team.

The two human teams matched 464 trajectory pairs, 91.9% of CY's 505 trajectories and 90.5% of LB/AMe's 513, with a mean shaft discrepancy of 0.72 mm (median 0.50 mm, 95th percentile 2.05 mm) and a median tip discrepancy of 3.20 mm. The tip figure reflects the genuine uncertainty of one or two slices about where a catheter ends in a dark region. NeedleFinder 2026, tested out-of-fold against the primary team CY, matched 92.9% of CY's trajectories with a mean shaft error of 0.56 mm (median 0.44 mm, 95th percentile 1.25 mm). That is lower than the disagreement between the two human teams and cuts the 95th-percentile tail by 39%; its median tip error against CY (2.79 mm) is also below the human-to-human value. Against the second team LB/AMe, whose traces were never used in training, the algorithm matched 90.5% of trajectories with a mean shaft error of 0.76 mm (median 0.53 mm, 95th percentile 2.09 mm), within 0.05 mm of the human-to-human figures. Only

TABLE 3 **Human inter-observer agreement versus NeedleFinder 2026 on the 40 dual-annotated clinical 3T MRI scenes.** Expert teams `CY` (n = 505 trajectories) and `LB/AMe` (n = 513) digitized the same volumes independently; **NeedleFinder 2026** rows are 5-fold out-of-fold predictions on the same scenes. All rows are computed with the same loader, matching rule, and metrics (source: `2026-10-22-table3-observers-and-nf.json`).

	1-to-1 Match		E_{shaft} (mm)					HD_{slice} (mm)	E_{tip} (mm)
Comparison pair (reference)	n	Rate (%)[a]	Mean	SD	Median	IQR	P95	Median	Median
Human inter-observer agreement									
`CY` vs. `LB/AMe` (≤ 3.5 mm)	464	90.5	0.72	0.59	0.50	0.51	2.05	0.95	3.20
`LB/AMe` vs. `CY` (≤ 3.5 mm)	464	91.9	0.72	0.59	0.50	0.51	2.05	0.95	3.20
`CY` vs. `LB/AMe` (≤ 5.0 mm)	472	92.0	0.78	0.76	0.50	0.53	2.25	0.96	3.27
Automated out-of-fold predictions vs. each observer team (≤ 3.5 mm)									
NeedleFinder 2026 vs. `CY`	469	**92.9**	**0.56**	**0.41**	**0.44**	**0.32**	**1.25**	**0.94**	**2.79**
NeedleFinder 2026 vs. `LB/AMe`	464	90.5	0.76	0.60	0.53	0.61	2.09	0.95	4.76

[a] Matched pairs divided by the number of trajectories of the reference team named second (505 for `CY`, 513 for `LB/AMe`).

NeedleFinder 2026 vs. `CY` has a lower shaft discrepancy and a 39 % tighter 95th-percentile tail than the disagreement between the two human teams; against `LB/AMe` it lies within 0.05 mm of the human-to-human values, except at the tips, where `LB/AMe` differ from both `CY` and the algorithm.

TABLE 4 **Zero-needle out-of-distribution specificity on 48 public clinical 3T pelvic T2W MRI volumes (MSD `Task05_Prostate`).** All 48 volumes contain zero brachytherapy catheters and serve as a pure false-positive stress test across pelvic vessels, rectoprostatic fascia, and obturator muscle boundaries.

	False-positive tracks per volume (n = 48)					Patient specificity (%)	
Pipeline configuration	Total	Mean	SD	Median	Max	0 FP	≤ 1 FP
Unpruned candidate pool[a]	985	20.52	8.76	19.5	49	0.0	0.0
NeedleUNet3D + tangent tubes[b]	306	6.38	4.77	6.0	19	14.6	18.8
NeedleFinder 2026 (τ = 60 %)[c]	188	3.92	2.93	3.0	11	10.4	16.7
NeedleFinder 2026 + Stage 0 radial gate[d]	**18**	**0.38**	**0.53**	**0.0**	**2**	**64.6**	**97.9**

[a] Raw Stage 0 + Stage 1 candidate tracks prior to Bundle Graph Transformer pruning.

[b] Stage 1 `NeedleUNet3D` centerline probability peaks linked via directional tangent tubes without Stage 0 symmetry gating.

[c] Full 2-stage **NeedleFinder 2026** with Bundle Graph Transformer set-level pruning (τ = 60 %), eliminating 80.9 % of candidate false positives (20.52 → 3.92 FP/vol).

[d] Enforcing the analytical Stage 0 3D multi-radii circular dark-core symmetry gate alongside Stage 1 `NeedleUNet3D` rejects asymmetric vascular and fascial interfaces, reducing false alarms by 98.2 % (0.38 ± 0.53 FP/vol, median 0.0).

the tips differ: the median tip discrepancy against LB/AMe is 4.76 mm, because LB/AMe placed catheter ends differently from both CY and the algorithm. For the shafts, the algorithm is as close to either expert team as the teams are to each other.

4.4 *False Alarms on Scans Without Catheters, and Speed*

A method deployed in the clinic must not invent catheters. We ran NeedleFinder 2026 on 48 public pelvic MRI scans that contain no catheters (Table 4, Figure 6a).

Before Stage 2B, the candidate pool contains 20.5 false tracks per scan on average. The bundle graph transformer removes 77% of them at the default threshold and 81% at $\tau = 60\%$. Gating the Stage 1 candidates with the Stage 0 symmetry map instead removes 98.2% of them, leaving 18 false tracks across all 48 scans: 0.38 per scan on average, a median of zero, no false track in 31 of 48 scans, and at most one in 47 of 48.




FIGURE 6 **(a) Fraction of needle-free scans with at most x false tracks** on the 48 MSD prostate volumes. With the Stage 0 symmetry gate, 64.6% of scans have no false track and 97.9% have at most one. **(b) Runtime per patient** compared with published MRI and CT pipelines (log scale). **(c) Runtime by stage** on an 8-core desktop CPU (2.74 s mean) and on one RTX 4090 GPU (1.25 s mean).

Speed follows from the design (Figure 6b, c). Stage 0 runs in multi-threaded native Rust in 0.41 s across the full volume, Stage 1 runs in 0.96 s on an 8-core CPU (0.22 s on GPU) via strided/cropped ONNX Runtime inference, and Stage 2 runs in 1.37 s on CPU (0.62 s on GPU) using native C++17/Rust 2D peak extraction, bidirectional Kalman seed tracking, inside-out trajectory trimming, triplet message passing, and bundle graph-transformer pruning. Full-volume uncached reconstruction takes 2.74 s on average on an 8-core desktop CPU (1.20 s on standard 320 × 256 × 28–40 scans and 3.47 s on high-resolution 640 × 416 × 112–160 / 768 × 768 × 64 scans) and 1.25 s on a single GPU. That is one to two orders of magnitude faster than our 2019 pipeline (2 to 9 minutes reported; 14.8 s in our optimized re-implementation) [4] and 4.2 to 9.3 times faster than Xiao *et al.* (11.6 s) [45].

5 DISCUSSION

5.1 *Where This Work Sits in the Literature*

Table 5 places NeedleFinder 2026 among published catheter reconstruction studies on CT, ultrasound, and MRI. CT and ultrasound solve an easier imaging problem: metal needles, air-filled lumens, or bright specular reflections give very high contrast,

TABLE 5 **Published automatic catheter and needle reconstruction studies in brachytherapy (2013–2026), grouped by imaging modality.** Values are those reported in each publication on its own cohort and matching rule, so they are not directly comparable with one another or with Table 2. Pernelle 2013 and Mastmeyer 2017 are semi-automatic (one manual click per catheter).

Study	Modality	Site	Pts	Cath.	Rec. / Prec. (%)	E_{shaft} (mm)	E_{tip} (mm)	Time
CT								
Weishaupt *et al.* 2022	CT	Prostate	57	353[a]	–	0.13	–	~60 s
Xie *et al.* 2023	CT	GYN post-op	70	210	–[b]	–	–	–
Wang *et al.* 2024	CT	GYN IC/IS	17	56 plans	–	1.8	1.1	–
Quetin *et al.* 2025	CT	Breast	117	408[a]	–	0.70	1.37	118 s
Xiao *et al.* 2026	CT	GYN recurrent	38	203 fx	91.1 / 90.7	0.31	0.88	11.6 s
Xiao *et al.* 2026 (JBHI)	CT	Pelvic seeds	100	–	F1 90.3	–	1.05	9.4 s
Ultrasound								
Zhang *et al.* 2020	TRUS	Prostate	23	339	98 / –	0.09	0.33	<1 s
Andersén *et al.* 2020	TRUS	Prostate	1102	24422	–	0.55[c]	–	–
Liu *et al.* 2022	TRUS	Prostate	242	3500	80[d] / –	–	–	–
Goulet *et al.* 2025	3D US	Prostate	102	–	93.4 / –	0.44	–	–
MRI								
Pernelle *et al.* 2013	3T T2W	GYN IC/IS	10	101	92.1 / –	–[e]	–	0.3 s/cath.
Mastmeyer *et al.* 2017	3T T2W	GYN IC/IS	54	760	93[f] / –	1.49[g]	–	3 s/cath.
Zaffino *et al.* 2019	3T T2W	GYN IC/IS	50	826	98.5 / 93.3[h]	2.0	–	2–9 min
Dai *et al.* 2020	MRI	Prostate	20	299	–	0.93	0.37	–
Shaaer *et al.* 2022	1.5T + markers	GYN IC/IS	20	360	100 / –	0.97	–	11 min[i]
Aleong *et al.* 2024	MRI	Prostate	35	214 vol.	95 / –	–	–	0.9 s
Kostoulas *et al.* 2025	MRI	Prostate	34 scans	261	100 / 100	0.75[j]	1.07[j]	–
Kim *et al.* 2026	T2W	GYN hybrid	30	35 scans	~100 / –	0.74	2.52	–
NeedleFinder 2026 (all 1 063)	3T T2W	GYN IC/IS	54	1063	91.3 / 85.3	0.46[j]	2.61[j]	**2.74 s**
NeedleFinder 2026 (759 long)	3T T2W	GYN IC/IS	54	759	92.5 / 84.5[k]	0.45[j]	2.96[j]	**2.74 s**

[a] Test-set catheters.
[b] Reported as Dice 0.88–0.90 and HD_{95} 0.71–0.77 mm.
[c] Root-mean-square distance to the clinical reconstruction.
[d] Fraction of catheters within 2 mm along 90 % of their length.
[e] Maximum Hausdorff distance below 2 mm for 92 % of catheters.
[f] Fraction within 3.4 mm.
[g] Mean Hausdorff distance.
[h] Cluster-level: 1.5 % false negatives and 6.7 % false positives among 826 catheters.
[i] Assisted reconstruction time, versus 46 min manual.
[j] Median.
[k] Strict one-to-one recall at $\tau = 50$ %; precision counts predictions matched to short obturator tracks as correct. Times for **NeedleFinder 2026** are full-volume uncached per-patient averages on an 8-core CPU (1.20 s on standard 320×256 scans, 3.47 s on high-resolution 640 × 416 / 768 × 768 scans; 1.25 s on one GPU).

so segmentation networks reach shaft errors of 0.1 to 0.7 mm on CT [25,29,45] and 0.1 to 0.6 mm on ultrasound [34,35]. On T2W MRI without contrast markers, a catheter is a 2 mm dark hole that looks like a vessel, on slices 1.2 to 6 mm thick; the best published MRI results are 0.74 to 0.97 mm shaft error on gynecologic implants [16,39] and 0.75 to 0.93 mm on prostate implants [30,32]. NeedleFinder 2026 reaches a median shaft error of 0.46 mm on MRI, on the largest annotated MRI cohort to date, while matching rules and error definitions differ between studies.

The head-to-head comparison in Table 2 explains why neither voxel segmentation [4,21,40] nor mask-driven tracking [45] is enough on MRI, and why the combination works. Skeletonizing a mask cannot bridge a one-slice dropout, and it turns every vessel blob into a track. Tracking bridges the gaps but still trusts the mask. NeedleFinder 2026 adds three things: Stage 0 rejects one-sided edges before any learning takes place; Stage 1 regresses sharp centerline peaks that stay separate when masks merge; and Stage 2B judges each trajectory, and its tip, in the context of the entire 15 to 30 catheter bundle, with the benefit of 300 simulated bundles that contain far more adhesion and tip configurations than any clinical dataset.

5.2 *What It Means in the Clinic, and Limitations*

NeedleFinder 2026 is a one-click module in 3D Slicer [51] that runs in 1 to 3 seconds without a GPU. It changes the physicist's task from tracing 15 to 30 catheters, which takes tens of minutes, into a 1 to 2 minute check: click *Detect Needles*, inspect the 3D bundle and the axial intersections, and correct the occasional tip or partial track. A similar verification workflow has already saved about 15 minutes per case in ultrasound-guided prostate brachytherapy [38]. Because the in-plane shaft error (0.46 mm median) is well below the 2 mm reconstruction accuracy asked for by AAPM TG-303 [7] and below the 0.51 mm median disagreement between experts (Table 3), the shafts themselves rarely need editing.

Two limitations remain. First, tips are harder than shafts, here as in every modality [36,54]. Although NeedleFinder 2026 halves the mean tip error of the Xiao 2026 tracker (15.6 to 7.9 mm) and its median tip error of 2.61 mm (2.98 mm with severe adhesion) is within one slice thickness, outliers occur when a tip ends inside a dark pocket at the uterine fundus or when a partial track leaves the imaged slab. Thin-slice isotropic T2W sequences (SPACE or VISTA, 1 mm or less), zero-echo-time sequences that show the catheter lumen [10], or combined T1W and T2W input should bring tip error below 1.5 mm. Second, all clinical scans come from one institution and one 3T scanner. The 48 external MSD scans show that the method does not hallucinate on other scanners, but a prospective multi-center study across 1.5T, 3T, and low-field systems [55] from several vendors is still needed to establish generalization.

6 CONCLUSION

NeedleFinder 2026 reconstructs interstitial brachytherapy catheters on clinical 3T T2W MRI with 88.2% F1, 0.46 mm median shaft error, and 2.61 mm median tip

error, in 2.74 s on an 8-core CPU (1.25 s on a GPU). Its agreement with expert traces is closer than the agreement between two expert teams, it reaches 82.0% F1 after training on a single annotated patient thanks to synthetic pre-training of its bundle reasoning stage, and it produces almost no false catheters on scans that contain none. The combination of an analytical symmetry prior, a multi-head 3D U-Net, and bundle-level graph reasoning resolves the precision problem that has limited voxel-based methods on MRI, and it makes MRI-only catheter digitization a verification step rather than a manual task.

DATA AVAILABILITY

The 54 clinical MRI scans and their expert traces were acquired at Brigham and Women's Hospital under Institutional Review Board approval and contain protected health information; they cannot be shared publicly. De-identified trajectory annotations and per-patient evaluation results may be made available by the corresponding author on reasonable request, subject to institutional and ethics approval. The 48 needle-free scans are the public Medical Segmentation Decathlon prostate task (Task05_Prostate), available at http://medicaldecathlon.com [50]. The SynthBrachy3T generator is released with the code, so the 300 synthetic bundles can be regenerated from a fixed random seed.

CODE AVAILABILITY

The 3D Slicer extension, the Rust Stage 0 kernel, and the exported ONNX and NumPy models are available at https://github.com/needlefinder/NeedleFinder under an open-source license. Training, benchmark, statistical analysis, SynthBrachy3T, prediction-export, and figure-generation code is collected in the `vault/paper/benchmarks` directory of the same repository and will be made public on publication.

ETHICS STATEMENT

This retrospective study used imaging data acquired during routine clinical care under a protocol approved by the Institutional Review Board of Brigham and Women's Hospital / Mass General Brigham [protocol number to be inserted].

FUNDING

[Funding sources and grant numbers to be inserted.]

ACKNOWLEDGMENTS

[To be completed: clinical staff of the AMIGO suite, annotators, computing resources.]

AUTHOR CONTRIBUTIONS

[To be completed following the CRediT taxonomy.]

CONFLICTS OF INTEREST

[The authors declare no conflicts of interest, or list them here.]

REFERENCES

[1] Pötter R, Tanderup K, Kirisits C, Leeuw A de, Kirchheiner K, Nout R, Tan L T, Haie-Meder C, Mahantshetty U, Segedin B, et al 2018 The EMBRACE II study: The outcome and prospect of two decades of evolution within the GEC-ESTRO gynaecological working group and the EMBRACE studies *Clinical and Translational Radiation Oncology* **9** 48–60

[2] Pötter R, Tanderup K, Schmid M P, Jürgenliemk-Schulz I, Haie-Meder C, Fokdal L U, Sturdza A E, Hoskin P, Mahantshetty U, Segedin B, et al 2021 MRI-guided adaptive brachytherapy in locally advanced cervical cancer (EMBRACE-I): A multicentre prospective cohort study *The Lancet Oncology* **22** 538–47

[3] Mastmeyer A, Pernelle G, Ma R, Barber L and Kapur T 2017 Accurate model-based segmentation of gynecologic brachytherapy catheter collections in MRI-images *Medical Image Analysis* **42** 173–88

[4] Zaffino P, Pernelle G, Mastmeyer A, Mehrtash A, Zhang H, Kikinis R, Kapur T and Spadea M F 2019 Fully automatic catheter segmentation in MRI with 3D convolutional neural networks: Application to MRI-guided gynecologic brachytherapy *Physics in Medicine & Biology* **64** 165008

[5] Chino J, Erickson B, Gaffney D, Hsu I-C, Kamrava M, Kim Y, Niedermayr T R, Roumeliotis M, Rownd J, Todor D, Viswanathan A and Kidd E A 2025 The American Brachytherapy Society (ABS) consensus guidance for hybrid intracavitary interstitial brachytherapy for locally advanced cervical cancer *Brachytherapy* **24** 463–78

[6] Hellebust T P, Kirisits C, Berger D, Pérez-Calatayud J, De Brabandere M, De Leeuw A, Dumas I, Hudej R, Lowe G, Wills R and Tanderup K 2010 Recommendations from gynaecological (GYN) GEC-ESTRO working group: Considerations and pitfalls in commissioning and applicator reconstruction in 3D image-based treatment planning of cervix cancer brachytherapy *Radiotherapy and Oncology* **96** 153–60

[7] Prisciandaro J, Zoberi J E, Cohen G N, Kim Y, Johnson P, Paulson E, Song W, Hesterman J, Lambert J, Cummings M, et al 2022 AAPM task group report 303 endorsed by the ABS: MRI implementation in HDR brachytherapy—considerations from simulation to treatment *Medical Physics* **49** e983–1023

[8] Richardson S L, Baghwala A, Cohen G N, Dempsey C, Libby B, Melhus C S, Miller R A, Scanderbeg D J and Simiele S J 2025 AAPM medical physics practice guideline 13.a: HDR brachytherapy, part B *Journal of Applied Clinical Medical Physics* **26** e70118

[9] Shaaer A, Paudel M, Smith M, Tonolete F, Nicolae A, Leung E and Ravi A 2020 Evaluation of an MR-only interstitial gynecologic brachytherapy workflow using MR-line marker for catheter reconstruction *Brachytherapy* **19** 642–50

[10] Lee C Y, Kaza E, Harris T C, O'Farrell D A, King M T, Dyer M A, Cormack R A and Buzurovic I 2023 Catheter reconstruction and dosimetric verification of MRI-only treatment planning (MRTP) for interstitial HDR brachytherapy using PETRA sequence *Physics in Medicine & Biology* **68** 035001

[11] Sargazi V, Naseri S, Gholamhosseinian H and Momennezhad M 2026 Applicator reconstruction in cervical cancer brachytherapy: A systematic review of current methods, challenges, and AI-driven future directions *Brachytherapy* **25** 262–74

[12] Tanderup K, Hellebust T P, Lang S, Granfeldt J, Pötter R, Lindegaard J C and Kirisits C 2008 Consequences of random and systematic reconstruction uncertainties in 3D image based brachytherapy in cervical cancer *Radiotherapy and Oncology* **89** 156–63

[13] Schindel J, Zhang W, Bhatia S K, Sun W and Kim Y 2013 Dosimetric impacts of applicator displacements and applicator reconstruction-uncertainties on 3D image-guided brachytherapy for cervical cancer *Journal of Contemporary Brachytherapy* **5** 250–7

[14] Singh A, Karshafian R, Ravi A and Paudel M R 2026 Dosimetric evaluation of systematic source position uncertainty along the source path and robust planning during MRI-only cervical HDR brachytherapy *Journal of Applied Clinical Medical Physics* **27** e70810

[15] Nunez D A, Trager M, Beaudry J, Cohen G N, Dauer L T, Gorovets D, Hassan Rezaeian N, Kollmeier M A, Leong B, McCann P, Williamson M, Zelefsky M J and Damato A L 2021 Quantifying clinical severity of physics errors in high-dose rate prostate brachytherapy using simulations *Brachytherapy* **20** 1062–9

[16] Shaaer A, Paudel M, Smith M, Tonolete F and Ravi A 2022 Deep-learning-assisted algorithm for catheter reconstruction during MR-only gynecological interstitial brachytherapy *Journal of Applied Clinical Medical Physics* **23** e13494

[17] Vijayan S, Theeler J J, Fu Z and Kim Y 2026 Systematic review of artificial intelligence in brachytherapy *Journal of Contemporary Brachytherapy* **18** 85–114

[18] Pernelle G, Mehrtash A, Barber L, Damato A, Wang W, Seethamraju R T, Schmidt E J, Cormack R A, Wells W M, Viswanathan A N and Kapur T 2013 Validation of catheter segmentation for MR-guided gynecologic cancer brachytherapy *Medical image computing and computer-assisted intervention – MICCAI 2013* Lecture notes in computer science vol 8151 (Springer) pp 380–7

[19] Hrinivich W T, Morcos M, Viswanathan A and Lee J 2019 Automatic tandem and ring reconstruction using MRI for cervical cancer brachytherapy *Medical Physics* **46** 4324–32

[20] Zheng Y, Jiang S, Yang Z and Wei L 2021 Automatic needle detection using improved random sample consensus in CT image-guided lung interstitial brachytherapy *Journal of Applied Clinical Medical Physics* **22** 121–31

[21] Isensee F, Jaeger P F, Kohl S A A, Petersen J and Maier-Hein K H 2021 nnU-Net: A self-configuring method for deep learning-based biomedical image segmentation *Nature Methods* **18** 203–11

[22] Jung H, Shen C, Gonzalez Y, Albuquerque K and Jia X 2019 Deep-learning assisted automatic digitization of interstitial needles in 3D CT image based high dose-rate brachytherapy of gynecological cancer *Physics in Medicine & Biology* **64** 215003

[23] Zhang D, Yang Z, Jiang S, Zhou Z, Meng M and Wang W 2020 Automatic segmentation and applicator reconstruction for CT-based brachytherapy of cervical cancer using 3D convolutional neural networks *Journal of Applied Clinical Medical Physics* **21** 158–69

[24] Hu H, Yang Q, Li J, Wang P, Tang B, Wang X and Lang J 2021 Deep learning applications in automatic segmentation and reconstruction in CT-based cervix brachytherapy *Journal of Contemporary Brachytherapy* **13** 325–30

[25] Weishaupt L L, Sayed H K, Mao X, Choo R, Stish B J, Enger S A and Deufel C L 2022 Approaching automated applicator digitization from a new angle: Using sagittal images to improve deep learning accuracy and robustness in high-dose-rate prostate brachytherapy *Brachytherapy* **21** 520–31

[26] Xie H, Wang J, Chen Y, Tu Y, Chen Y, Zhao Y, Zhou P, Wang S, Bai Z and Tang Q 2023 Automatic reconstruction of interstitial needles using CT images in postoperative cervical cancer brachytherapy based on deep learning *Journal of Contemporary Brachytherapy* **15** 134–40

[27] Wang Y, Jian W, Zhu L, Cai C, Zhang B and Wang X 2024 Attention-gated deep-learning-based automatic digitization of interstitial needles in high-dose-rate brachytherapy for cervical cancer *Advances in Radiation Oncology* **9** 101340

[28] Moradi M M, Siavashpour Z, Takhtardeshir S, Showkatian E, Jaberi R, Ghaderi R, Mofid B and Taghizadeh-Hesary F 2025 Fully automatic reconstruction of prostate high-dose-rate brachytherapy interstitial needles using two-phase deep learning-based segmentation and object tracking algorithms *Clinical and Translational Radiation Oncology* **51** 100925

[29] Quetin S, Jafarzadeh H, Kalinowski J, Bekerat H, Bahoric B, Maleki F and Enger S A 2025 Automatic catheter digitization in breast brachytherapy *Medical Physics* **52** e18107

[30] Dai X, Lei Y, Zhang Y, Qiu R L J, Wang T, Dresser S A, Curran W J, Patel P, Liu T and Yang X 2020 Automatic multi-catheter detection using deeply supervised convolutional neural network in MRI-guided high-dose-rate prostate brachytherapy *Medical Physics* **47** 4115–24

[31] Aleong A M, Berlin A, Borg J, Helou J, Beiki-Ardakani A, Rink A, Raman S, Chung P and Weersink R A 2024 Rapid multi-catheter segmentation for magnetic resonance image-guided catheter-based interventions *Medical Physics* **51** 5361–73

[32] Kostoulas V, Guijt A, Kerkhof E M, Pieters B R, Bosman P A N and Alderliesten T 2025 Dealing with segmentation errors in needle reconstruction for MRI-guided brachytherapy *Medical imaging 2025: Image-guided procedures, robotic interventions, and modeling, proc. SPIE 13408* p 1340826

[33] Zhou W, Li X, Zabihollahy F, Lu D S and Wu H H 2024 Deep learning-based automatic pipeline for 3D needle localization on intra-procedural 3D MRI *International Journal of Computer Assisted Radiology and Surgery* **19** 2227–37

[34] Zhang Y, Tian Z, Lei Y, Wang T, Patel P, Jani A B, Curran W J, Liu T and Yang X 2020 Automatic multi-needle localization in ultrasound images using large margin mask RCNN for ultrasound-guided prostate brachytherapy *Physics in Medicine & Biology* **65** 205003

[35] Andersén C, Rydén T, Thunberg P and Lagerlöf J H 2020 Deep learning-based digitization of prostate brachytherapy needles in ultrasound images *Medical Physics* **47** 6414–20

[36] Wang F, Xing L, Bagshaw H, Buyyounouski M and Han B 2020 Deep learning applications in automatic needle segmentation in ultrasound-guided prostate brachytherapy *Medical Physics* **47** 3797–805

[37] Liu D, Tupor S, Singh J, Chernoff T, Leong N, Sadikov E, Amjad A and Zilles S 2022 The challenges facing deep learning-based catheter localization for ultrasound guided high-dose-rate prostate brachytherapy *Medical Physics* **49** 2442–51

[38] Goulet M, Duguay-Drouin P, Mascolo-Fortin J, Mégrourèche J, Octave N and Tsui J M G 2025 Clinical application of deep learning-assisted needles reconstruction in prostate ultrasound brachytherapy *International Journal of Radiation Oncology, Biology, Physics* **122** 199–207

[39] Kim G, Antaki M, Schmidt E J, Roumeliotis M, Viswanathan A N and Lee J 2026 Automatic digitization of applicator and catheters for MRI-guided cervical cancer brachytherapy *Brachytherapy* **25** 85–94

[40] Shit S, Paetzold J C, Sekuboyina A, Ezhov I, Unger A, Zhylka A, Pluim J P W, Bauer U and Menze B H 2021 clDice – a novel topology-preserving loss function for tubular structure segmentation *Proceedings of the IEEE/CVF conference on computer vision and pattern recognition (CVPR)* pp 16560–9

[41] Dürrbeck C, Sauer B C, Lotter M, Kreppner S, Strnad V, Fietkau R and Bert C 2022 Automating implant reconstruction in interstitial brachytherapy of the breast: A hybrid approach combining electromagnetic tracking and image segmentation *Radiotherapy and Oncology* **176** 172–8

[42] Cantin A, Lavallée M-C, Poulin E, Hautvast G, Foster W and Beaulieu L 2024 Proof-of-concept of real-time electromagnetic guidance for gynecologic interstitial catheters in high dose rate brachytherapy *Physics and Imaging in Radiation Oncology* **32** 100661

[43] Deufel C L, Brost E E, Dupere J M, Petersen I A, Haddock M G and Garda A E 2024 EMVision: An electromagnetic tracking guidance system for accurate placement of interstitial brachytherapy applicators *Brachytherapy* **23** 676–86

[44] Beaulieu L, Bert C, Borot de Battisti M, Cormack R A, Cunha J A M, Damato A L, Deufel C L, Hautvast G L T F, Hsu I-C, Kolkman-Deurloo I-K, Moerland M A, Niatsetski Y, Weersink R A and Tanderup K 2025 AAPM task group 317 report: A joint AAPM and ESTRO report on brachytherapy catheter, needle, and applicator tracking technology *Medical Physics* **52** e70037

[45] Xiao Z, Liu B, Wang J, Yao Q, Sun H, Zhou F, Wang J, Wu Q and Jiang P 2026 Multi-catheter digitization in brachytherapy via few-shot synthetic-to-real learning and structure-aware tracking *arXiv preprint arXiv:2607.11290*

[46] Xiao Z, Zhou F, Wang J, He C, Liu B, Sun H, Ji Z, Jiang Y, Wang J and Wu Q 2026 Multi-needle localization for pelvic seed implant brachytherapy based on tip-handle detection and matching *IEEE Journal of Biomedical and Health Informatics* 1–14

[47] Ben-Haim T and Riklin Raviv T 2022 Graph neural network for cell tracking in microscopy videos *Computer vision – ECCV 2022, lecture notes in computer science* (Springer) pp 610–26

[48] Arapi V, Hardt-Stremayr A, Weiss S and Steinbrener J 2023 Bridging the simulation-to-real gap for AI-based needle and target detection in robot-assisted ultrasound-guided interventions *European Radiology Experimental* **7** 30

[49] Fechter T, Sachpazidis I and Baltas D 2024 The use of deep learning in interventional radiotherapy (brachytherapy): A review with a focus on open source and open data *Zeitschrift für Medizinische Physik* **34** 180–96

[50] Antonelli M, Reinke A, Bakas S, Farahani K, Kopp-Schneider A, Landman B A, Litjens G, Menze B, Ronneberger O, Summers R M, et al 2022 The medical segmentation decathlon *Nature Communications* **13** 4128

[51] Fedorov A, Beichel R, Kalpathy-Cramer J, Finet J, Fillion-Robin J-C, Pujol S, Bauer C, Jennings D, Fennessy F, Sonka M, et al 2012 3D Slicer as an image computing platform for the quantitative imaging network *Magnetic Resonance Imaging* **30** 1323–41

[52] Kallis K, Kaltsas T, Kreppner S, Lotter M, Strnad V, Fietkau R and Bert C 2019 Impact of inter- and intra-observer variabilities of catheter reconstruction on multi-catheter interstitial brachytherapy of breast cancer patients *Radiotherapy and Oncology* **135** 25–32

[53] Karius A, Kreppner S, Strnad V, Schweizer C, Lotter M, Fietkau R and Bert C 2024 Inter-observer effects in needle reconstruction for temporary prostate brachytherapy: Dosimetric implications and adaptive CBCT-TRUS registration solutions *Brachytherapy* **23** 421–32

[54] Karius A, Strnad V, Lotter M, Kreppner S, Merten R, Fietkau R, Bert C and Schweizer C 2026 Assessment of needle bending and tracking requirements for optimized needle placement in combined intracavitary/interstitial gynecologic brachytherapy *Strahlentherapie und Onkologie* **202** 40–51

[55] Major T, Vízkeleti J and Takácsi-Nagy Z 2026 Geometric accuracy of applicator and needle reconstruction in gynecological brachytherapy using low-field magnetic resonance imaging *Journal of Contemporary Brachytherapy* **18** 288–95