

INTENT TO AWARD SOLE SOURCE CONTRACT

THIS IS NOT A SOLICITATION OR REQUEST FOR OFFERS. NO SOLICITATION PACKAGE IS AVAILABLE AND TELEPHONE REQUESTS WILL NOT BE HONORED AND NO SOLICITATION WILL BE POSTED TO CONTRACT OPPORTUNITIES.

INTENT TO AWARD: The Department of Veterans Affairs, VAMC Manchester, NH intends to award one (1) Purchase Order for a one-year service contract, for the MGC Diagnostics Ultima PFX pulmonary function/stress testing system, and for the MGC Diagnostics Platinum Elite DX body plethysmography device. All services and maintenance must be performed as described in the enclosed Statement of Work (SOW).

INTERESTED PARTIES: However, any Contractor who believes that they have experience and capability to provide equipment and perform installation on equipment described in the enclosed Statement of Work (SOW) may express interest by submitting a written capability statement or letter of interest to pearl.wray@va.gov by **4:30 PM EST, by 3/3/2025**. Any questions related to this notice can be sent to pearl.wray@va.gov

PROJECT TITLE: MGC DIAGNOSTICS EQUIPMENT SERVICE CONTRACT

BACKGROUND: Manchester VA Medical Center has one MGC Diagnostics Ultima PFX pulmonary function/stress testing system and one MGC Diagnostics Platinum Elite DX body plethysmography device. The devices will need to be maintained, by a qualified and reliable company, for ongoing quality assurance, software package modifications including upgrades & patches, telephone technical support, and future onsite repairs, calibrations, and preventive maintenance.

WORK DESCRIPTION: Replace parts coverage, service support, telephone support, software and hardware support, continuing education for VA staff and support service, full maintenance coverage by providing all replacement parts, labor, travel expenses, tools, test equipment and other incidentals necessary to maintain or restore the listed equipment to operating specifications according to the manufacturer's recommendations.

NAICS: The intended North American Classification System Code (NAICS) is 622110 General Medical and Surgical Hospitals, \$47.M Size Standard.

PSC: The intended Product or Service Code (PSC) is J065 Maint/Repair/Rebuild of Equipment—Medical, Dental, and Veterinary Equipment and Supplies

SET-ASIDE: It is anticipated the procurement will be a Total SDVOSB set-aside.

AUTHORITY: The authority for applying the Simplified Acquisition Procedures for Commercial Items of FAR 13.5 is 41 U.S.C. 1901. This statute is implemented for restricted competition on this procurement via FAR 8.405-6(a)(1)(i)(B). A determination by the Government not to compete this proposed award upon response to this notice is solely within the discretion of the Contracting Officer.

ENCLOSED: Statement of Work

STATEMENT OF WORK FOR SERVICE AND SUPPORT OF MGC Diagnostics Equipment Service Contract

A. BACKGROUND

Manchester VA Medical Center has one MGC Diagnostics Ultima PFX pulmonary function/stress testing system and one MGC Diagnostics Platinum Elite DX body plethysmography device. The devices will need to be maintained, by a qualified and reliable company, for ongoing quality assurance, software package modifications including upgrades & patches, telephone technical support, and future onsite repairs, calibrations, and preventive maintenance.

B. OBJECTIVES

The contract effort goals are to maintain, upgrade, service and support all components of the equipment listed in Section F throughout the total contract period at the Manchester VA Medical Center. Objectives to meet the service and support goal include, but are not limited to, replacement parts coverage, service support, telephone support, software and hardware support, continuing education for VA staff and support service, full maintenance coverage by providing all replacement parts, labor, travel expenses, tools, test equipment and other incidentals necessary to maintain or restore the listed equipment to operating specifications according to the manufacturer's recommendations throughout the term of the contract.

C. PERIOD OF PERFORMANCE

After the contract has been awarded, the period of performance will be defined within the following time parameters.

- 1 Year from award date

D. PLACE OF PERFORMANCE

Delivery of services in this case will be conducted virtual and/or at the following address:

- Manchester VA Medical Center
 - 718 Smyth Road, Manchester, NH 03104

E. SCOPE OF WORK (TASKS FOR SERVICES)

The work requirements for this contract performance period are introduced as stated in the preceding sections. However, it is here within the scope of work they will be captured explicitly. The contractor will meet the following requirements as bulleted below, THE CONTRACTOR WILL:

- **REPAIRS**

- All Field Engineer on site visits shall be performed during the normal VA business hours
 - 8:00am-5:00pm, Monday through Friday, excluding Federal holidays
 - On-site response time shall be within 24 hours of request
- All Field Engineers or company representative must check in with Healthcare Technology Management (HTM) (aka: Biomedical Engineering or Clinical Engineering) prior to starting their work at the facility.
- All telephone technical support shall be provided at no additional cost
- Repair parts shall be provided and shipped, next day 10:30 am at no extra cost to the Government

- **SCHEDULE MAINTENANCE (PREVENTIVE MAINTENANCE)**

- One Planned Maintenance (PM) visit with complete inspection of listed instrumentation. Service (work) includes labor, travel, a PM kit, complete inspection of the instrument with cleaning. Includes one additional PM or cleaning if necessary.
- Preventive maintenance shall be scheduled during normal VA business hours as identified above.
- Preventive maintenance shall be coordinated with the Service Line and HTM so that there is minimal negative impact to patient care.
- All parts and equipment necessary to complete the preventive maintenance shall be provided and shipped, next day 10:30 am at no extra cost to the Government
- **SOFTWARE**
 - All software updates, inclusive of installation of any minor patches or major revisions, shall be included at no extra cost to the Government
 - The Contractor shall assist in any/all software and/or network vulnerabilities identified by the Government that need remediation as expressed by the Government.
 - All software will be approved and/or scanned by HTM (aka: Biomedical Engineering or Clinical Engineering) prior to installation.
- **SERVICE DOCUMENTATION**
 - At the conclusion of each repair and scheduled maintenance call, the Contractor shall provide a written service report to HTM (aka: Biomedical Engineering or Clinical Engineering) within 3 business days indicating:
 - Date(s) of service
 - Make, model, serial number, and barcode/ID number of the equipment serviced
 - Name of FSE who performed the work
 - Hours worked
 - Description of the reported problem, the services performed, and parts replaced
 - Make, model, serial number, barcode/ID number, and proof of up-to-date calibration of the test equipment used in the repair or preventive maintenance service.
 - Quantifiable measurement results, inclusive of range and tolerance, for all tests conducted

F. EQUIPMENT LIST

Item #	Catalog Number	Serial Number	Product
1	MGC Diagnostics	251000224	MGC Diagnostics Ultima PFX
1	MGC Diagnostics	243000745	MGC Diagnostics Platinum Elite DX