

Purchase Description

Procurement of Dew Point Mirror

The U.S. Army Test, Measurement and Diagnostic Equipment Activity (USATA) has a requirement to procure (1) Model 373H Dew Point Mirror along with an ISO calibration or equivalent for the Army's Primary Standards Laboratory (APSL). The equipment shall be delivered to USATA's Army Physical Standards Labs located at Building 5435, Fowler Road Redstone Arsenal. The Dew Point Mirror supports USATA's mission as follow:

1.0 Requirements:

<u>Model</u>	<u>Description</u>	<u>Quantity</u>
373H	Dew Point Mirror Range: -30°C FP... +70°C DP Ext temp probe RH indication, Data Acquisition software, Heated hose Internal components. Case included	1
373H/HX	ISO Calibration of 373H -20, 0, 20, 40, 60, 69°C FP/DP Ext.Temp: 0, 25, 50, 70°C Head & Ext. Pressure: 80, 101, 120k Pa add:1.00 10, 20, 45, 70, 90 %RH @ 23°C.	1

2.0 Delivery Requirements

Shipping Address: U.S. Army TMDE Activity, Attn: Matt Watts, Building 5435, Redstone Arsenal, AL 35898