

[REDACTED]

CLASS JUSTIFICATION FOR AN EXCEPTION TO FAIR OPPORTUNITY (JEFO)
Zoll Medical Propaq MD Vital Signs Monitor/Defibrillator
CORPORATE EXIGENCY CONTRACT SPE2D1-24-D-0017

(1) The Defense Logistics Agency (DLA) Troop Support, Medical Supply Chain, Capital Equipment Division, is the contracting activity.

(2) The action submitted for approval is a class JEFO for the placement of sole-source delivery orders for Zoll Medical Propaq MD Vital Signs Monitor/Defibrillator and accessories and parts under Zoll's Corporate Exigency Contract (CEC) SPE2D1-24-D-0017. This class JEFO will apply to all orders placed under the subject contract commencing with the date of approval of the JEFO through the CEC's expiration on [REDACTED].

(3) The United States (US) Army, US Navy, and US Air Force (DoD Tri-Services), and other US Federal Agencies/Services have ongoing requirements for Zoll Medical Propaq MD Vital Signs Monitor/Defibrillator, US Air Force (USAF) Configuration part number 306-2451312-01 without the case (NSN: 6515-01-644-9586), and for the option of US Army Configuration part number 303-2451312-01 with the case (NSN: 6515-01-659-4040), and all the accessories and working components.

The total maximum dollar value of the Zoll contract is [REDACTED]. This contract was awarded on [REDACTED] for a one (1) year base period with four (4) one-year option periods. The forecasted dollar value of the sole source delivery orders to be placed under the referenced contract for Zoll Medical Propaq MD Vital Signs Monitor/Defibrillator and accessories is [REDACTED].

(4) The authority waiving the competition requirements is FAR 16.505(b)(2)(i)(B), only one awardee is capable of providing the supplies or services required at the level of quality required because the supplies or services ordered are unique or highly specialized.

This acquisition is to allow for the necessary and urgent purchases of Zoll Medical Propaq MD Vital Signs Monitor/Defibrillator, USAF Configuration without the case, and for the option of US Army Configuration with the case, and all the accessories and working components for the DoD Tri-Services and US Federal Agencies for critical care duties through [REDACTED].

[REDACTED]

The required Vital Signs Monitor/Defibrillator must satisfy the following minimum requirements (MRs) that were established by DoD Tri-Services and other US Agencies/ Services for a monitor/defibrillator in a field environment:

- a) Shall have a Joint Airworthiness Certification (JAC) consisting of a “Safe-to-Fly” letter from the USAF for fixed and rotary wing aircraft and a Fleet Airworthiness Release (AWR) and an Aeromedical Certification Letter (AWC) from the U.S. Army.
- b) Shall provide a minimum of two (2) ports each for temperature monitoring and invasive blood pressure (IBP) monitoring.
- c) Shall provide Pleth Variability Index (PVI).
- d) Shall be capable of operating on line voltages of 110-240V AC, at frequencies of 50/60 Hz, and on aircraft power at a line voltage of 110V AC at a frequency of 400 Hz.
- e) Shall be capable of mounting to the existing Vital Signs Monitor bracket on the Special Medical Emergency Evacuation Device (SMEED).
- f) Shall offer a night vision goggle (NVG) display mode for military and air medical nighttime operation.
- g) Shall be a portable vital signs monitor/defibrillator capability that has FDA 510k/ Pre-Market Approval (PMA) and that is Automated External Defibrillation (AED) capable.
- h) Shall be operable solely on battery power for a minimum of 2 hours while providing monitoring of all patient parameters and a minimum of ten (10) defibrillations or cardioversions at maximum energy levels.
- i) Shall enable achievement of a 25% reduction in weight over the current defibrillator and patient monitor used in USAF Aeromedical Operations; the current weight of the two systems is 26 pounds; therefore, the threshold for a 25% reduction is 19.65 pounds

The Zoll Propaq MD Vital Signs Monitor/Defibrillator identified above is the only product available in the U.S. market that can meet all the MRs listed above.

(5) The Zoll Propaq MD Vital Signs Monitor/Defibrillators that are required are available on Zoll's CEC SPE2D1-24-D-0017. The contract pricing has previously been determined fair and reasonable [REDACTED]

[REDACTED] In view of this earlier determination, the Contracting Officer has determined that the anticipated cost will be fair and reasonable, and that the subject procurement represents the best value to the government.

[REDACTED]

(6) Market research was performed by DLA Troop Support's Engineer in January-February 2025. Following is a summary of the results:

- [REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
- [REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

- [REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

The market research revealed that none of the other vendors' similar monitors/ defibrillators can meet all the MRs. Propaq MD Vital Sign Monitor/Defibrillator supplied by Zoll Medical is the only monitor/defibrillator in the U.S. market that has JAC and can meet all the MRs.

(7) There are no other facts supporting the use of this class JEFO.

(8) In an effort to remove barriers to fair opportunity and competition on subsequent acquisitions, DLA Troop Support will continue to monitor other medical vital signs monitor/defibrillator vendors as they may have the opportunity to achieve JAC and obtain a CEC. Additional vendors obtaining CECs and certification will enhance competition for future acquisitions.

(9) As the Contracting Officer, I hereby certify that the information contained in this justification is accurate and complete to the best of my knowledge and belief.

[REDACTED]

Date

ANTHONY GIUNTA
Contracting Officer, FSD

[REDACTED]

[Redacted]

I hereby certify that the requirements that form the basis of this justification are accurate, that these supplies meet the Government's minimum needs, and that this justification is accurate and complete to the best of my knowledge and belief.

[Redacted]

Date

Biomedical Engineer, FSDC

I have reviewed and hereby approve this justification.

KENNY.W
ILLIAM.J.
12289688
26

Digitally signed
by
KENNY.WILLIAM
.J.1228968826
Date: 2025.04.07
12:43:29 -04'00'

Date

WILLIAM J. KENNY
Executive Director
DLA Troop Support Contracting
and Acquisition Management

[Redacted]