

Purchase Description

Procurement of Fluke Units

The U.S. Army Test, Measurement and Diagnostic Equipment Activity (USATA) has a requirement to procure two (2) Model 9114-256 Metrology Furnace, one (1) Model 9117-256 Annealing Furnace, one (1) Model 7381-26 Deep-Well Compact Baths, and one (1) 2019-DCB Liquid-in-Glass Thermometer Calibration Kit or equivalent to support USATA's mission. The equipment shall be delivered to USATA's Army Physical Standards Labs located at Building 5435, Fowler Road Redstone Arsenal. The Metrology Furnace to support USATA's mission as follow:

1.0 Requirements:

<u>Model</u>	<u>Description</u>	<u>Quantity</u>
9114-256	Metrology Furnace Range of 100 °C to 680 °C Three-zone furnace with the best in digital controller technology Power Range 230 VAC (±10 %), 50/60 Hz, 1 Phase, 22 A maximum	2
9117-256	Annealing Furnace Range of 200 °C and 450 °C Thermal Wells Five: 8 mm diameter x 430 mm long (0.31 x 16.9 in) Over-Temp Protection: Separate circuit protects furnace from exceeding rated temperature limit	1
7381-26	Deep-Well Compact Baths Range of –80 °C to 110 °C Heating Time: 60 minutes, from 25 °C to 100 °C (oil 5012) Size: 1067 x 356 x 788 mm (940 mm from floor to tank access opening) [42 x 14 x 31 in (37 in from floor to tank access opening)]	1
2019-DCB	Liquid-in-Glass Thermometer Bath adapter tube Thermometer Carousel	1

2.0 Delivery Requirements

Shipping Address: U.S. Army TMDE Activity, Attn: Matt Watts, Building 5435, Redstone Arsenal, AL 35898