

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT		BPA NO.	1. CONTRACT ID CODE	PAGE 1 OF 2 PAGES 2
2. AMENDMENT/MODIFICATION NUMBER 0002		3. EFFECTIVE DATE 05-09-2025		4. REQUISITION/PURCHASE REQ. NUMBER
5. PROJECT NUMBER (if applicable)		6. ISSUED BY CODE 36C257 Department of Veterans Affairs NCO 17 North Texas Healthcare System Regional Procurement Office West (RPO-W) 124 E. Hwy 67, Suite 100 Duncanville TX 75137		7. ADMINISTERED BY (If other than Item 6) CODE 36C257 NCO 17 Department of Veterans Affairs North Texas Veterans Health Care Center 124 E. Hwy 67, Suite 100 Duncanville TX 75137
8. NAME AND ADDRESS OF CONTRACTOR (Number, street, county, State and ZIP Code)		(X)	9A. AMENDMENT OF SOLICITATION NUMBER 36C25725Q0505	
		X	9B. DATED (SEE ITEM 11)	
			10A. MODIFICATION OF CONTRACT/ORDER NUMBER	
			10B. DATED (SEE ITEM 13)	
CODE		FACILITY CODE		
11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS				
X The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers is extended, X is not extended. Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods: (a) By completing Items 8 and 15, and returning _____ copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or electronic communication which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by letter or electronic communication, provided each letter or electronic communication makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.				
12. ACCOUNTING AND APPROPRIATION DATA (If required) See CONTINUATION Page				
13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS, IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.				
CHECK ONE	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.			
	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).			
	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:			
	D. OTHER (Specify type of modification and authority)			
E. IMPORTANT: Contractor is not, is required to sign this document and return _____ copies to the issuing office.				
14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.) The purpose of this amendment is to post questions and answers. See continuation page				
Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.				
15A. NAME AND TITLE OF SIGNER (Type or print)		16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)		
15B. CONTRACTOR/OFFEROR (Signature of person authorized to sign)		15C. DATE SIGNED		16B. UNITED STATES OF AMERICA (Signature of Contracting Officer)
		16C. DATE SIGNED		

PREVIOUS EDITION NOT USABLE

STANDARD FORM 30 (REV. 11/2016)
Prescribed by GSA - FAR (48 CFR) 53.243

CONTINUATION PAGE

1. A2LA is recognized by NELAC. Could a laboratory accredited by A2LA be accepted in lieu of the Texas accreditation?

Answer: The A2LA accreditation is acceptable in lieu of the NELAP.

2. within the tables on the SOW, it doesn't mention daily conductivity testing which is required in the standard 108.

Answer: Daily conductivity monitoring of the critical water system will be performed by the VA SPS Staff.

3. It also mentions daily visual inspections which are in the ST108 but intended to be completed by the on-site technicians (VA's technicians). In the solicitation it could be interpreted as the contractor is required to do the visual inspections but the ST108's intention is for the facility to do them

Answer: The VA SPS Staff will be doing the visual inspection inside the equipment of the spray arm and chambers

End of Document