

**NAVAL MEDICAL RESEARCH UNIT SAN ANTONIO
PERFORMANCE WORK STATEMENT**

Life-Technologies (under the Thermo Fisher Scientific business umbrella) QuantStudio 6 Flex Real Time PCR System Service Contract

1. Scope of Work

The Naval Medical Research Unit San Antonio (NAMRU-SA) requires, via this PWS, a vendor to provide annual maintenance and an on-site service agreement for its QuantStudio 6 Flex PCR System. NAMRU-SA conducts multiple research studies investigating gene expression markers, bioburden, and host responses related to infections and wound healing that require use of the QuantStudio 6 Flex real-time PCR System and the related software necessary to acquire and analyze the data. Any downtime due to inoperable components, parameters out of line with factory specifications, or other reductions in performance due to equipment not maintained optimally and within factory determined specifications leads to significant losses of time, man-hours, critical data, and research productivity. This loss of time, data, and productivity directly harms the mission of NAMRU-SA to provide gap driven research to improve the survival, operational readiness, and safety of DoD personnel.

2. Definitions and Terms

The Applied Biosystems® QuantStudio™ 6 Flex Real-Time PCR System facilitates real-time PCR based applications with an upgradable block selection, allowing scientists to utilize technology available in the present and as well as the future. With a simplified workflow, intuitive software, and touch-screen interface, the QuantStudio™ 6 Flex system offers exceptional reproducibility. It is compatible with over 8 million TaqMan® assays for gene expression, genetic variation, gene regulation, or protein expression experiments.

3. Support Requirements

The contracted laboratory/company shall provide a preventative service contract as required by NAMRU-SA on an as-needed basis. The vendor shall provide one on-site preventive maintenance per year, including replacement parts, for the following component: QuantStudio 6 Flex PCR System (SN: 278862709, ECN: 105993). The vendor must also provide factory certified parts, labor, and travel for remedial repairs. Provide on-site service calls for remedial repairs with a guaranteed priority response time of two business days after receipt of a service call (Service Zone 2). Service calls may be preceded by telephone troubleshooting during regular business hours Monday through Friday. Online troubleshooting and service via remote access to any system component is not permitted. All troubleshooting and service must be provided on-site. The vendor must also provide priority telephone and email access to instrument and application technical support.

4. Additional Requirements/Qualifications

The vendor selected for the performance of this statement of work must provide a trained and certified service engineer/technician. This training and certification is required in order to ensure the equipment is operating within factory specifications. The equipment must operate within factory specifications to maintain consistency in research data. Furthermore, it is preferred that the vendor have previous experience working with the Navy and Department of Defense laboratories.

This contract shall include a Non-disclosure Agreement (NDA) signed by the personnel performing on the contract and counter-signed by their management.

5. Delivery Address

Naval Medical Research Unit San Antonio
3650 Chambers Pass, Bldg 3610
JBSA Fort Sam Houston, TX 78234
ATTN: John W. Simecek
Phone: (210) 539-6991

6. Contract Period of Performance

The anticipated period of performance is 12 months.

7. Points of Contact

Technical POC

Dr. John W. Simecek, DDS, MPH
Phone: (210) 539-6991

Email: john.w.simecek2.civ@health.mil

Contract POC

Ms. Abigail Johnston
Phone: (210) 539-6252

Email: Abigail.z.johnston.civ@health.mil