

The measurement International 6020A, 6011D and 6000C equipment in question matches up with the 4020A and 4610 currently in the primary electrical lab. The 3 models will seamlessly work into our current calibration software and procedures without any added disruptions. NASA currently has the older versions of the 6020A and 6000C which are not compatible with the Windows 11 push. Our current 4020 and 4610 will wire right up and work with the newer versions as they are slave assets to the resistance bridges. NASA require these assets to obtain minimal added uncertainties of .03ppm. The minimum added accuracy to our primary resistance standards makes it possible to calibrate and have a TUR >1.2:1 36 GFE highly accurate 7.5 and 8.5 digit, 3458A and 8508/8588 multimeters that have accuracies as tight of 7.7 ppm. The MI resistance bridges also gives us the ability to calibrate our 12 primary Voltage and Resistance calibrators Fluke 5520/22, 5700's and 5720's that have accuracies as tight as 9 ppm. These calibrators and digital multimeters give us the ability to calibrate all the voltage and resistance assets that come through the calibration lab. Maintaining the ability to keep the inhouse calibration of the calibration labs standards cuts down on turn time and direct cost to the center. This system also gives us the accuracies to calibrate our GFE super thermometers Fluke 1594A in house that have accuracies of 4ppm at the tightest. These assets have a massive trickle down to the calibration laboratory affecting both temperature and the voltage labs. Having these assets makes it a plug and play situation and won't cause a delay to future calibration in rewriting procedures and training.

ppm (parts per million)

TUR (Total uncertainty ratio)