

PERFORMANCE WORK STATEMENT
FOR REPAIR AND CALIBRATION of
ROHDE & SCHWARZ MODEL

1.0 GENERAL. The U.S. Army Test, Measurement and Diagnostic Equipment Activity (USATA) has a requirement to acquire commercial calibration and repair services (C&RS) for one (1) model FSV40-N Spectrum Analyzer, manufactured by Rohde & Schwarz USA, Inc.

2.0 GOVERNMENT TASKS.

2.1 The government shall be responsible for shipment of the unit from USATA facility to the Contractor's facility at government expense. The unit shall be shipped to the Contractor's facility not later than 48 hours (excluding weekends and holidays) after receipt of notice of contract award.

2.2 The U.S. Army TMDE Activity, Redstone Arsenal, AL (Acceptance DODAAC: W80RA6) shall be responsible for receiving, inspection and acceptance in DFAS Wide Area Workflow (WAWF).

3.0 CONTRACTOR TASKS.

3.1 Model FSV40-N Spectrum Analyzer, S/N 102514, under RMA number 967350 needs repair. The contractor shall repair the unit to its original state and provide accredited calibration compliant to ISO/IEC 17025. As a minimum, a calibration report with the following shall be included: as found date, as left data, calibration standards used, and the measurement uncertainty. The Contractor shall furnish all labor, proprietary information, material and equipment required to provide diagnosis, repair and calibration of the item in accordance with the manufacturer's specifications.

4.0 **PLACE OF PERFORMANCE.** Performance shall take place at the Contractor's facilities.

5.0 **PERIOD OF PERFORMANCE.** In accordance with Section B of the contract.

6.0 **GOVERNMENT-FURNISHED EQUIPMENT.** None.

7.0 **QUALITY REQUIREMENTS.** Not Applicable.