

**Statement of Work
NWIHCS-Omaha Imaging Service
Nuclear Medicine Software**

- I. **Background:** The current radiopharmaceutical dose tracking equipment (Syntrac) in the Nuclear Medicine department has been end-of-life and is no longer serviceable. A replacement system is required to maintain storage of patient and radiopharmaceutical dose tracking for regulatory compliance.
- II. **Purpose of Project:** The current dose tracking system is end-of-life and is no longer serviceable. The purchase of a Nuclear Medicine software solution to replace the current system and maintain regulatory compliance by archiving radiopharmaceutical doses and monitoring and recording health physics procedures including surveys, instruments, and sealed sources.
- III. **Statement of Work**
 - a. **Description of Work:** Contractor to provide NMIS 3 user network software system that includes HL7 and Dose Calibrator interfaces. Contractor will also ensure data conversion from Syntrac (current end-of-life system) to the NMIS software network for archival. Contractor will also provide Web Training and Access for Nuclear Medicine technologists utilizing the software.
 - b. **Schedule of Deliverables:**
 - 1. Software components installed following delivery.
 - 2. Software components to be delivered for installation on government furnished computer. Setup/technologist training to follow software installation.
 - 3. Reports: N/A
 - c. **Contractor Qualifications:** The Contractor's employees must possess knowledge required for installation of dose tracking software with HL7 and Dose Calibrator interfaces and troubleshooting methods. They must also possess knowledge of the required record keeping for NRC and NHPP Nuclear Medicine regulatory compliance.
- IV. **Evaluation and Award Factors:** The award shall be made to the offer representing the best value to the Government considering technical and price factors as stated below. The technical aspects of HL7 and Dose Calibrator interfaces and the archival of past data from the existing system into the new software system for maintaining regulatory compliance will take precedence over price when selection is made.

- a. The software should be created and sold by those with an intimate knowledge of Nuclear Medicine and associated regulatory compliance measures. Offerors shall include in their proposals specific information relative to the evaluation factors listed below to ensure proper evaluation of their proposal.
 1. The product should have a proven track record with history of upgrade over 10 years or more
 2. Staff providing training/setup must have thorough knowledge of Nuclear Medicine isotopes, procedures, QA/QC procedures, and regulatory compliance measure.
 3. < \$30,000
 4. Lowest price technically acceptable

V. Program Office Point of Contact and/or Contracting Officer

Technical Representative (COTR): Indicate name, telephone number, email address (Include a copy of the COTR training certificate and continuing education certificates if appropriate). COTR will not be required to be assigned for non-complex procurements.

VI. Government Furnished Property: Server: Windows Server 2016 SP2, 2019 or Windows 10 64 bit, Windows, 8GB RAM, 1 TB Hard Drive, Windows 10 64 bit, Intel Core i5
System: 8 GB RAM, 500 GB Hard Drive, 2GB Flash Drive, 32GB Flash Drive, 20: Flat Panel LCD, DVD Drive, Mouse and Keyboard, Printer, Internet access

VII. Contractor Personnel Security Requirement: Software will connect to the Network. Contractor will provide training demonstration to staff utilizing the HIS/RIS.

VIII. Market Research: Research consisted of online searches for Nuclear Medicine tracking software vendors who provide HL7 and Dose Calibrator Interfaces and a method of transferring existing Syntrac data to an updated database. A TEAMS meeting was held with a vendor to demonstrate the product, available options for the software, and to answer any questions the Nuclear Medicine staff and Radiation Safety Officer had. Online brochures and PDFs were provided by a vendor for the product and interfaces.

IX. Delivery and Packing, Inspection/Acceptance requirements, Warranty: Limited warranty, as described in the attached Information Guide.