

STATEMENT OF WORK (SOW) InnoWave PFC Sonic Washer

Purpose: The Michael E. DeBakey VA Medical Center (MEDVAMC) is requesting to replace sonic washers EE#1165706, EE1204539, and EE 1204545 the sonic washers has reached the end of its life usefulness and has been determine by the Biomedical Section as uneconomical to repair, thus the need for a replacement Ultrasonic Cleaner.

Item	Equipment #	Description	Quantity
1.0	IWPCF GTIN: 00724995152710	InnoWave PCF Sonic Irrigator Thermal Disinfection & Robotic Program	2 EA
1.1	IW1038 GTIN: 00724995152956	IW PCF Xi Multi Bskt Hose Set for 4 Stndrd 2 Stplr	2 EA
1.2	MED3088	ELBOW 6MM X 1/8" FLOW	8 EA
1.3	SE130152	Install InnoWave PCF	2 EA
1.4	SE1301521	Deinstall InnoWave PCF	2 EA
1.41	SE6025100071	Disposal of Washer	2 EA
1.5	SE130151	1st Year Preventive Care InnoWave PCF	2 EA
2.0	IW523546NPR GTIN: 00724995193737	InnoWave Unity 20 Sonic Irrigator 208V 60Hz	1 EA
2.1	MP22683	IW Unity Xi Tubing Only (8 Pack)	1 EA
2.2	SE130202	Install InnoWave Unity	1 EA
2.3	SE1302021	Deinstall InnoWave Unity	1 EA
2.31	SE6025100071	Disposal of Washer	1 EA
2.4	SE1302010045	1st Year Preventive Care InnoWave Unity	1 EA

Scope:

Discover effortless operation, driven by increased efficiency with the high capacity InnoWave PFC Sonic Irrigator (208V 3Ph/60Hz) which will ultrasonically clean up to 60 lumened devices & up to 66 lbs of instruments. The InnoWave Pro offers Sonic Irrigation technology which exposes both the instruments external surfaces and internal channels (if connected) to ultrasonic activity. The InnoWave PFC offers multiple cycle phases like pre-wash, wash, disinfection, and rinse phases.

Performance Monitoring:

N/A

Security Requirements:

There are no security requirements the purchase of the needed surgical instruments.

Place of Performance

MEDVAMC Operating Suites